

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** Q1523 **SAS No.:** Q1523 **SDG NO.:** Q1523

Continuing Calib Date: 03/12/2025 **Initial Calibration Date(s):** 02/21/2025 02/21/2025

Continuing Calib Time: 13:26 **Initial Calibration Time(s):** 19:56 21:32

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.16	7.16	7.06	7.26	0.00
2,4-D	8.27	8.27	8.17	8.37	0.00
2,4,5-TP(Silvex)	9.13	9.14	9.04	9.24	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** Q1523 **SAS No.:** Q1523 **SDG NO.:** Q1523

Continuing Calib Date: 03/12/2025 **Initial Calibration Date(s):** 02/21/2025 02/21/2025

Continuing Calib Time: 13:26 **Initial Calibration Time(s):** 19:56 21:32

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.62	7.63	7.53	7.73	0.01
2,4-D	8.85	8.86	8.76	8.96	0.01
2,4,5-TP(Silvex)	9.74	9.76	9.66	9.86	0.02