

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: HSTDCCC750 Data File : PS029396.D Time Analyzed: 09:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.132	9.038	9.238	754.570	712.500	5.9
2,4-D	8.267	8.171	8.371	724.490	705.000	2.8
2,4-DCAA	7.156	7.060	7.260	777.430	750.000	3.7

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: HSTDCCC750 Data File : PS029396.D Time Analyzed: 09:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.741	9.656	9.856	742.300	712.500	4.2
2,4-D	8.848	8.761	8.961	701.440	705.000	-0.5
2,4-DCAA	7.620	7.532	7.732	730.650	750.000	-2.6