

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: HSTDCCC750 Data File : PS029405.D Time Analyzed: 13:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.132	9.038	9.238	769.220	712.500	8.0
2,4-D	8.266	8.171	8.371	737.740	705.000	4.6
2,4-DCAA	7.155	7.060	7.260	795.560	750.000	6.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: HSTDCCC750 Data File : PS029405.D Time Analyzed: 13:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.741	9.656	9.856	758.770	712.500	6.5
2,4-D	8.847	8.761	8.961	719.310	705.000	2.0
2,4-DCAA	7.619	7.532	7.732	746.890	750.000	-0.4