

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL03 Date Analyzed: 03/12/2025

Lab Sample No.: HSTDCCC750 Data File : PS029417.D Time Analyzed: 18:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.131	9.038	9.238	768.960	712.500	7.9
2,4-D	8.267	8.171	8.371	735.450	705.000	4.3
2,4-DCAA	7.154	7.060	7.260	797.340	750.000	6.3

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL03 Date Analyzed: 03/12/2025

Lab Sample No.: HSTDCCC750 Data File : PS029417.D Time Analyzed: 18:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.739	9.656	9.856	761.330	712.500	6.9
2,4-D	8.846	8.761	8.961	720.000	705.000	2.1
2,4-DCAA	7.619	7.532	7.732	749.630	750.000	0.0