

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

Instrument ID: ECD_L

Calibration Date(s): 03/11/2025 03/11/2025

Calibration Times: 10:35 11:29

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:							CF 100 =	<u>PL094569.D</u>	CF 075 =	<u>PL094570.D</u>	
CF 050 =							<u>PL094571.D</u>	CF 025 =	<u>PL094572.D</u>	CF 005 =	<u>PL094573.D</u>
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD				
Decachlorobiphenyl	1938960000	1961100000	2020410000	2196220000	2420590000	2107460000	10				
Endrin	2585940000	2584670000	2628790000	2810860000	3250100000	2772070000	10				
gamma-BHC (Lindane)	3886090000	3824550000	3841420000	4060770000	4338530000	3990270000	5				
Heptachlor	3645730000	3661000000	3717820000	3965990000	4417600000	3881630000	8				
Heptachlor epoxide	3127680000	3124100000	3197280000	3386170000	3890740000	3345200000	10				
Methoxychlor	1109170000	1148450000	1148150000	1275940000	1303650000	1197070000	7				
Tetrachloro-m-xylene	2677630000	2681900000	2722740000	2922670000	3148400000	2830670000	7				

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** Q1523 **SAS No.:** Q1523 **SDG NO.:** Q1523
Instrument ID: ECD_L **Calibration Date(s):** 03/11/2025 03/11/2025
Calibration Times: 10:35 11:29

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL094569.D</u>	CF 075 = <u>PL094570.D</u>				
CF 050 = <u>PL094571.D</u>		CF 025 = <u>PL094572.D</u>	CF 005 = <u>PL094573.D</u>				
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	3950320000	3868520000	3903950000	4095240000	4378770000	4039360000	5
Endrin	4434550000	4325290000	4326060000	4330520000	4401830000	4363650000	1
gamma-BHC (Lindane)	5372230000	5230100000	5132530000	5070750000	4891640000	5139450000	3
Heptachlor	5363060000	5262730000	5228770000	5273220000	5215450000	5268650000	1
Heptachlor epoxide	4619190000	4519300000	4529520000	4593130000	4631780000	4578580000	1
Methoxychlor	2089160000	2085630000	2099350000	2155560000	2175640000	2121070000	2
Tetrachloro-m-xylene	3572520000	3517330000	3502400000	3623970000	3630140000	3569270000	2