

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 19:38 Initial Calibration Time(s): 10:35 11:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.05	9.06	8.96	9.16	0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.01
Heptachlor	4.91	4.92	4.82	5.02	0.01
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 19:38 Initial Calibration Time(s): 10:35 11:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.90	7.91	7.81	8.01	0.01
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
gamma-BHC (Lindane)	3.60	3.60	3.50	3.70	0.00
Heptachlor	3.94	3.94	3.84	4.04	0.00
Heptachlor epoxide	4.72	4.73	4.63	4.83	0.01
Endrin	5.63	5.63	5.53	5.73	0.00
Methoxychlor	6.60	6.61	6.51	6.71	0.01