

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094640.D Time Analyzed: 15:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.053	8.956	9.156	51.250	50.000	2.5
Endrin	6.571	6.474	6.674	45.720	50.000	-8.6
gamma-BHC (Lindane)	4.326	4.227	4.427	48.680	50.000	-2.6
Heptachlor	4.913	4.815	5.015	47.570	50.000	-4.9
Heptachlor epoxide	5.682	5.583	5.783	48.580	50.000	-2.8
Methoxychlor	7.499	7.400	7.600	49.660	50.000	-0.7
Tetrachloro-m-xylene	3.537	3.438	3.638	49.080	50.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094640.D Time Analyzed: 15:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.904	7.807	8.007	53.140	50.000	6.3
Endrin	5.631	5.534	5.734	50.500	50.000	1.0
gamma-BHC (Lindane)	3.603	3.504	3.704	52.540	50.000	5.1
Heptachlor	3.940	3.842	4.042	50.950	50.000	1.9
Heptachlor epoxide	4.722	4.625	4.825	51.880	50.000	3.8
Methoxychlor	6.604	6.507	6.707	50.260	50.000	0.5
Tetrachloro-m-xylene	2.771	2.672	2.872	51.710	50.000	3.4