

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094653.D Time Analyzed: 19:38

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Decachlorobiphenyl   | 9.050 | 8.956                | 9.156 | 57.460             | 50.000            | 14.9 |
| Endrin               | 6.570 | 6.474                | 6.674 | 46.520             | 50.000            | -7.0 |
| gamma-BHC (Lindane)  | 4.325 | 4.227                | 4.427 | 51.810             | 50.000            | 3.6  |
| Heptachlor           | 4.912 | 4.815                | 5.015 | 48.640             | 50.000            | -2.7 |
| Heptachlor epoxide   | 5.681 | 5.583                | 5.783 | 50.640             | 50.000            | 1.3  |
| Methoxychlor         | 7.497 | 7.400                | 7.600 | 50.870             | 50.000            | 1.7  |
| Tetrachloro-m-xylene | 3.537 | 3.438                | 3.638 | 52.660             | 50.000            | 5.3  |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094653.D Time Analyzed: 19:38

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Decachlorobiphenyl   | 7.903 | 7.807                | 8.007 | 56.290             | 50.000            | 12.6 |
| Endrin               | 5.631 | 5.534                | 5.734 | 51.880             | 50.000            | 3.8  |
| gamma-BHC (Lindane)  | 3.602 | 3.504                | 3.704 | 55.610             | 50.000            | 11.2 |
| Heptachlor           | 3.940 | 3.842                | 4.042 | 52.170             | 50.000            | 4.3  |
| Heptachlor epoxide   | 4.721 | 4.625                | 4.825 | 54.580             | 50.000            | 9.2  |
| Methoxychlor         | 6.604 | 6.507                | 6.707 | 50.690             | 50.000            | 1.4  |
| Tetrachloro-m-xylene | 2.771 | 2.672                | 2.872 | 55.740             | 50.000            | 11.5 |