

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/13/2025

Lab Sample No.: PSTDCCC050 Data File : PL094677.D Time Analyzed: 15:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.051	8.956	9.156	58.110	50.000	16.2
Endrin	6.571	6.474	6.674	52.220	50.000	4.4
gamma-BHC (Lindane)	4.325	4.227	4.427	50.970	50.000	1.9
Heptachlor	4.912	4.815	5.015	49.860	50.000	-0.3
Heptachlor epoxide	5.680	5.583	5.783	50.170	50.000	0.3
Methoxychlor	7.497	7.400	7.600	59.780	50.000	19.6
Tetrachloro-m-xylene	3.536	3.438	3.638	50.700	50.000	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/13/2025

Lab Sample No.: PSTDCCC050 Data File : PL094677.D Time Analyzed: 15:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.903	7.807	8.007	56.820	50.000	13.6
Endrin	5.631	5.534	5.734	53.420	50.000	6.8
gamma-BHC (Lindane)	3.602	3.504	3.704	53.970	50.000	7.9
Heptachlor	3.940	3.842	4.042	53.200	50.000	6.4
Heptachlor epoxide	4.721	4.625	4.825	53.810	50.000	7.6
Methoxychlor	6.604	6.507	6.707	59.420	50.000	18.8
Tetrachloro-m-xylene	2.770	2.672	2.872	53.610	50.000	7.2