

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
 Data File : BF141898.D
 Acq On : 10 Mar 2025 11:30
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Quant Time: Mar 10 15:32:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 15:01:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	230406	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	926118	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	546702	20.000	ng	0.00
64) Phenanthrene-d10	11.398	188	1002901	20.000	ng	0.00
76) Chrysene-d12	14.033	240	745624	20.000	ng	0.00
86) Perylene-d12	15.504	264	527129	20.000	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	145959	10.570	ng	0.00
7) Phenol-d6	6.486	99	186951	10.630	ng	-0.02
23) Nitrobenzene-d5	7.433	82	159777	9.670	ng	-0.01
42) 2,4,6-Tribromophenol	10.698	330	63996	9.238	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	390913	10.888	ng	0.00
79) Terphenyl-d14	12.980	244	500607	9.924	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	2.675	88	28210	4.874	ng	99
3) Pyridine	3.451	79	70090	4.928	ng	99
4) n-Nitrosodimethylamine	3.369	42	31893	4.717	ng	# 99
6) Aniline	6.534	93	92116	5.305	ng	100
8) 2-Chlorophenol	6.657	128	81845	5.347	ng	98
10) Phenol	6.504	94	98369	5.301	ng	96
11) bis(2-Chloroethyl)ether	6.604	93	74158	5.334	ng	98
12) 1,3-Dichlorobenzene	6.816	146	84952	5.136	ng	100
13) 1,4-Dichlorobenzene	6.898	146	87237	5.213	ng	98
14) 1,2-Dichlorobenzene	7.045	146	84183	5.342	ng	99
15) Benzyl Alcohol	7.010	79	71182	5.000	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.151	45	91659	5.470	ng	97
17) 2-Methylphenol	7.116	107	62211	5.114	ng	99
18) Hexachloroethane	7.392	117	31373	5.000	ng	96
19) n-Nitroso-di-n-propyla...	7.281	70	58981	5.218	ng	98
20) 3+4-Methylphenols	7.269	107	83461	5.349	ng	97
22) Acetophenone	7.281	105	119040	5.358	ng	99
24) Nitrobenzene	7.451	77	80338	4.904	ng	99
25) Isophorone	7.692	82	149807	5.135	ng	99
26) 2-Nitrophenol	7.775	139	34757	4.329	ng	98
27) 2,4-Dimethylphenol	7.804	122	57106	5.164	ng	96
28) bis(2-Chloroethoxy)met...	7.904	93	95214	5.210	ng	97
29) 2,4-Dichlorophenol	8.010	162	67710	4.934	ng	99
30) 1,2,4-Trichlorobenzene	8.098	180	78127	5.160	ng	98
31) Naphthalene	8.180	128	255632	5.388	ng	98
33) 4-Chloroaniline	8.228	127	88635	5.294	ng	99
34) Hexachlorobutadiene	8.298	225	48885	4.966	ng	99
35) Caprolactam	8.557	113	21358	5.118	ng	94
36) 4-Chloro-3-methylphenol	8.692	107	77370	5.070	ng	99
37) 2-Methylnaphthalene	8.869	142	172734	5.427	ng	99
38) 1-Methylnaphthalene	8.969	142	166050	5.387	ng	99
40) 1,2,4,5-Tetrachloroben...	9.033	216	84034	5.041	ng	99
41) Hexachlorocyclopentadiene	9.022	237	27931	4.246	ng	96
43) 2,4,6-Trichlorophenol	9.145	196	52998	4.877	ng	99
44) 2,4,5-Trichlorophenol	9.180	196	54111	4.901	ng	97
46) 1,1'-Biphenyl	9.333	154	223637	5.395	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031025\
 Data File : BF141898.D
 Acq On : 10 Mar 2025 11:30
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Quant Time: Mar 10 15:32:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 15:01:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
47) 2-Chloronaphthalene	9.357	162	163642	5.298	ng	97
48) 2-Nitroaniline	9.451	65	40475	4.561	ng	98
49) Acenaphthylene	9.774	152	241253	5.268	ng	99
50) Dimethylphthalate	9.627	163	202373	5.328	ng	100
51) 2,6-Dinitrotoluene	9.686	165	37796	4.785	ng	99
52) Acenaphthene	9.945	154	168975	5.225	ng	98
53) 3-Nitroaniline	9.857	138	39761	4.902	ng	98
55) Dibenzofuran	10.116	168	252715	5.481	ng	97
56) 4-Nitrophenol	10.004	139	26252	4.309	ng	99
57) 2,4-Dinitrotoluene	10.092	165	49913	4.800	ng	98
58) Fluorene	10.457	166	197266	5.542	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.233	232	48472	4.839	ng	98
60) Diethylphthalate	10.327	149	201554	5.286	ng	99
61) 4-Chlorophenyl-phenyle...	10.451	204	99747	5.380	ng	98
62) 4-Nitroaniline	10.463	138	38863	4.945	ng	94
63) Azobenzene	10.610	77	190233	5.370	ng	98
66) n-Nitrosodiphenylamine	10.569	169	169225	5.223	ng	98
67) 4-Bromophenyl-phenylether	10.945	248	60238	4.823	ng	98
68) Hexachlorobenzene	11.004	284	65620	4.753	ng	98
69) Atrazine	11.092	200	54251	6.867	ng	98
70) Pentachlorophenol	11.198	266	32716	3.843	ng	100
71) Phenanthrene	11.421	178	289580	5.340	ng	98
72) Anthracene	11.474	178	294198	5.407	ng	98
73) Carbazole	11.627	167	249634	5.329	ng	99
74) Di-n-butylphthalate	11.957	149	330999	5.352	ng	99
75) Fluoranthene	12.610	202	309273	5.432	ng	99
77) Benzidine	12.733	184	64673	6.138	ng	99
78) Pyrene	12.839	202	315969	4.941	ng	99
80) Butylbenzylphthalate	13.457	149	117256	4.653	ng	99
81) Benzo(a)anthracene	14.021	228	255308	5.205	ng	100
82) 3,3'-Dichlorobenzidine	13.986	252	68969	4.865	ng	100
83) Chrysene	14.062	228	213144	4.837	ng	99
84) Bis(2-ethylhexyl)phtha...	14.015	149	167165	4.836	ng	99
85) Di-n-octyl phthalate	14.627	149	213395	4.435	ng	99
87) Indeno(1,2,3-cd)pyrene	16.980	276	144536	4.228	ng	99
88) Benzo(b)fluoranthene	15.068	252	187154	5.234	ng	99
89) Benzo(k)fluoranthene	15.098	252	153142	4.903	ng	99
90) Benzo(a)pyrene	15.439	252	137043	4.860	ng	99
91) Dibenzo(a,h)anthracene	16.998	278	119937	4.252	ng	99
92) Benzo(g,h,i)perylene	17.427	276	120142	4.294	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 10 15:32:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:01:52 2025
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.881 min Scan# 814
 Delta R.T. -0.000 min
 Lab File: BF141898.D
 Acq: 10 Mar 2025 11:30




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Tgt Ion:152 Resp: 230406
 Ion Ratio Lower Upper
 152 100
 150 155.9 127.4 191.2
 115 64.6 51.9 77.9



#2
 1,4-Dioxane
 Concen: 4.874 ng
 RT: 2.675 min Scan# 99
 Delta R.T. -0.000 min
 Lab File: BF141898.D
 Acq: 10 Mar 2025 11:30

Tgt Ion: 88 Resp: 28210
 Ion Ratio Lower Upper
 88 100
 58 71.4 57.4 86.2
 43 27.5 21.5 32.3









#3
Pyridine
Concen: 4.928 ng
RT: 3.451 min Scan# 217
Delta R.T. 0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 79 Resp: 70090
Ion Ratio Lower Upper
79 100
52 67.3 53.3 79.9
51 35.2 27.6 41.4



#4
n-Nitrosodimethylamine
Concen: 4.717 ng
RT: 3.369 min Scan# 217
Delta R.T. -0.024 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 42 Resp: 31893
Ion Ratio Lower Upper
42 100
74 129.9 103.4 155.0
44 9.5 5.0 7.4#









#5
2-Fluorophenol
Concen: 10.570 ng
RT: 5.487 min Scan# 51
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 112 Resp: 145959
Ion Ratio Lower Upper
112 100
64 57.8 46.5 69.7
63 29.9 25.4 38.2



#6
Aniline
Concen: 5.305 ng
RT: 6.534 min Scan# 755
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 93 Resp: 92116
Ion Ratio Lower Upper
93 100
66 39.6 31.9 47.9
65 19.6 15.8 23.6









#7
Phenol-d6
Concen: 10.630 ng
RT: 6.486 min Scan# 747
Delta R.T. -0.018 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 99 Resp: 186951
Ion Ratio Lower Upper
99 100
42 17.5 14.6 21.8
71 36.4 30.8 46.2



#8
2-Chlorophenol
Concen: 5.347 ng
RT: 6.657 min Scan# 776
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 128 Resp: 81845
Ion Ratio Lower Upper
128 100
130 31.5 12.8 52.8
64 45.5 24.4 64.4









#10
Phenol
Concen: 5.301 ng
RT: 6.504 min Scan# 753
Delta R.T. -0.018 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 94 Resp: 98369
Ion Ratio Lower Upper
94 100
65 28.4 10.5 50.5
66 43.3 25.8 65.8



#11
bis(2-Chloroethyl)ether
Concen: 5.334 ng
RT: 6.604 min Scan# 767
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 93 Resp: 74158
Ion Ratio Lower Upper
93 100
63 75.3 53.0 93.0
95 32.9 12.9 52.9





#12

1,3-Dichlorobenzene

Concen: 5.136 ng

RT: 6.816 min Scan# 804

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDIC005





Tgt Ion:146 Resp: 84952

Ion Ratio Lower Upper

146 100

148 63.8 51.3 76.9

75 33.3 26.6 40.0



#13

1,4-Dichlorobenzene

Concen: 5.213 ng

RT: 6.898 min Scan# 817

Delta R.T. -0.000 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30




Tgt Ion:146 Resp: 87237


Ion Ratio Lower Upper

146 100

148 64.3 51.0 76.6

111 42.2 35.7 53.5









#14
1,2-Dichlorobenzene
Concen: 5.342 ng
RT: 7.045 min Scan# 841
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDIC005


Tgt Ion:146 Resp: 84183
Ion Ratio Lower Upper
146 100
148 64.7 50.8 76.2
111 47.4 38.0 57.0



#15
Benzyl Alcohol
Concen: 5.000 ng
RT: 7.010 min Scan# 836
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 79 Resp: 71182
Ion Ratio Lower Upper
79 100
108 76.1 59.7 89.5
77 64.8 51.0 76.6









#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.470 ng
RT: 7.151 min Scan# 80
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 45 Resp: 91659
Ion Ratio Lower Upper
45 100
77 16.5 0.0 37.8
79 13.0 0.0 34.2



#17
2-Methylphenol
Concen: 5.114 ng
RT: 7.116 min Scan# 854
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 107 Resp: 62211
Ion Ratio Lower Upper
107 100
108 112.2 90.2 135.2
77 47.6 39.6 59.4
79 48.5 39.1 58.7









#18
Hexachloroethane
Concen: 5.000 ng
RT: 7.392 min Scan# 901
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:117 Resp: 31373
Ion Ratio Lower Upper
117 100
119 99.3 76.9 115.3
201 104.7 79.9 119.9



#19
n-Nitroso-di-n-propylamine
Concen: 5.218 ng
RT: 7.281 min Scan# 882
Delta R.T. -0.018 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 70 Resp: 58981
Ion Ratio Lower Upper
70 100
42 53.7 41.4 62.0
101 9.9 8.2 12.2
130 22.5 18.2 27.4









#20
3+4-Methylphenols
Concen: 5.349 ng
RT: 7.269 min Scan# 81
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:107	Resp:	83461
Ion	Ratio	Lower Upper
107	100	
108	88.4	68.4 108.4
77	83.1	69.7 109.7
79	30.9	11.4 51.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:136	Resp:	926118
Ion	Ratio	Lower Upper
136	100	
137	11.1	8.8 13.2
54	7.7	5.8 8.8
68	5.3	4.1 6.1









#22
Acetophenone
Concen: 5.358 ng
RT: 7.281 min Scan# 81
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:	105	Resp:	119040
Ion	Ratio	Lower	Upper
105	100		
71	8.0	5.8	8.8
51	29.0	23.4	35.0
120	22.0	17.3	25.9



#23
Nitrobenzene-d5
Concen: 9.670 ng
RT: 7.433 min Scan# 908
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:	82	Resp:	159777
Ion	Ratio	Lower	Upper
82	100		
128	42.0	36.0	54.0
54	49.2	39.6	59.4









#24
Nitrobenzene
Concen: 4.904 ng
RT: 7.451 min Scan# 911
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 77 Resp: 80338
Ion Ratio Lower Upper
77 100
123 45.2 37.0 55.4
65 13.6 11.3 16.9



#25
Isophorone
Concen: 5.135 ng
RT: 7.692 min Scan# 952
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 82 Resp: 149807
Ion Ratio Lower Upper
82 100
95 7.8 6.4 9.6
138 18.3 14.9 22.3









#26
2-Nitrophenol
Concen: 4.329 ng
RT: 7.775 min Scan# 90
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:139 Resp: 34757
Ion Ratio Lower Upper
139 100
109 30.8 26.5 39.7
65 49.4 40.3 60.5



#27
2,4-Dimethylphenol
Concen: 5.164 ng
RT: 7.804 min Scan# 971
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:122 Resp: 57106
Ion Ratio Lower Upper
122 100
107 113.8 94.8 142.2
121 56.3 47.8 71.6









#28
bis(2-Chloroethoxy)methane
Concen: 5.210 ng
RT: 7.904 min Scan# 989
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 93	Resp:	95214
Ion Ratio	Lower	Upper
93	100	
95	32.1	26.6 40.0
123	11.3	10.2 15.4



#29
2,4-Dichlorophenol
Concen: 4.934 ng
RT: 8.010 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 162	Resp:	67710
Ion Ratio	Lower	Upper
162	100	
164	64.4	43.8 83.8
98	38.2	18.1 58.1









#30
1,2,4-Trichlorobenzene
Concen: 5.160 ng
RT: 8.098 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:180 Resp: 78127
Ion Ratio Lower Upper
180 100
182 94.3 77.5 116.3
145 31.6 25.8 38.6



#31
Naphthalene
Concen: 5.388 ng
RT: 8.180 min Scan# 1035
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:128 Resp: 255632
Ion Ratio Lower Upper
128 100
129 10.5 9.0 13.6
127 12.9 10.8 16.2









#33
4-Chloroaniline
Concen: 5.294 ng
RT: 8.228 min Scan# 1043
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:	127	Resp:	88635
Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.2	39.4
65	30.5	25.4	38.2
92	19.4	16.2	24.4



#34
Hexachlorobutadiene
Concen: 4.966 ng
RT: 8.298 min Scan# 1055
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:	225	Resp:	48885
Ion	Ratio	Lower	Upper
225	100		
223	63.6	50.6	75.8
227	64.4	52.2	78.2





#35

Caprolactam

Concen: 5.118 ng

RT: 8.557 min Scan# 1107

Delta R.T. -0.047 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:113 Resp: 21358

Ion Ratio Lower Upper

113 100

55 145.9 129.6 169.6

56 113.1 104.0 144.0



#36

4-Chloro-3-methylphenol

Concen: 5.070 ng

RT: 8.692 min Scan# 1122

Delta R.T. -0.012 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30




Tgt Ion:107 Resp: 77370


Ion Ratio Lower Upper

107 100

144 22.8 19.0 28.6

142 74.2 59.2 88.8









#37
2-Methylnaphthalene
Concen: 5.427 ng
RT: 8.869 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:142 Resp: 172734
Ion Ratio Lower Upper
142 100
141 88.1 71.2 106.8



#38
1-Methylnaphthalene
Concen: 5.387 ng
RT: 8.969 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:142 Resp: 166050
Ion Ratio Lower Upper
142 100
141 93.9 74.2 111.4









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1180
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:164 Resp: 546702
Ion Ratio Lower Upper
164 100
162 101.8 81.8 122.6
160 45.2 36.7 55.1



#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.041 ng
RT: 9.033 min Scan# 1180
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:216 Resp: 84034
Ion Ratio Lower Upper
216 100
214 78.6 61.9 92.9
179 22.8 18.1 27.1
108 18.7 14.7 22.1









#41
Hexachlorocyclopentadiene
Concen: 4.246 ng
RT: 9.022 min Scan# 1179
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 237 Resp: 27931
Ion Ratio Lower Upper
237 100
235 66.6 43.1 83.1
272 13.2 0.0 32.0



#42
2,4,6-Tribromophenol
Concen: 9.238 ng
RT: 10.698 min Scan# 1463
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 330 Resp: 63996
Ion Ratio Lower Upper
330 100
332 96.3 77.6 116.4
141 33.4 24.7 37.1





#43

2,4,6-Trichlorophenol

Concen: 4.877 ng

RT: 9.145 min Scan# 1199

Delta R.T. -0.000 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:196 Resp: 52998

Ion Ratio Lower Upper

196 100

198 99.7 80.6 121.0

200 31.6 25.8 38.6



#44

2,4,5-Trichlorophenol

Concen: 4.901 ng

RT: 9.180 min Scan# 1205

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Tgt Ion:196 Resp: 54111




Ion Ratio Lower Upper


196 100

198 98.4 80.7 121.1

97 49.4 42.1 63.1

132 27.2 23.4 35.2









#45
2-Fluorobiphenyl
Concen: 10.888 ng
RT: 9.233 min Scan# 1214
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:172 Resp: 390913
Ion Ratio Lower Upper
172 100
171 34.9 29.3 43.9
170 23.0 19.4 29.0



#46
1,1'-Biphenyl
Concen: 5.395 ng
RT: 9.333 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:154 Resp: 223637
Ion Ratio Lower Upper
154 100
153 39.8 21.1 61.1
76 11.9 0.0 31.3





#47

2-Chloronaphthalene

Concen: 5.298 ng

RT: 9.357 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:162 Resp: 163642

Ion Ratio Lower Upper

162 100

127 41.7 35.5 53.3

164 32.8 26.6 40.0



#48

2-Nitroaniline

Concen: 4.561 ng

RT: 9.451 min Scan# 1251

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30




Tgt Ion: 65 Resp: 40475


Ion Ratio Lower Upper

65 100

92 71.9 56.6 84.8

138 112.5 88.6 133.0





#49

Acenaphthylene

Concen: 5.268 ng

RT: 9.774 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:152 Resp: 241253

Ion Ratio Lower Upper

152 100

151 20.1 16.3 24.5

153 13.6 11.1 16.7



#50

Dimethylphthalate

Concen: 5.328 ng

RT: 9.627 min Scan# 1281

Delta R.T. -0.012 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30




Tgt Ion:163 Resp: 202373


Ion Ratio Lower Upper

163 100

194 3.8 3.1 4.7

164 10.1 8.2 12.4









#51
2,6-Dinitrotoluene
Concen: 4.785 ng
RT: 9.686 min Scan# 1291
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDIC005


Tgt Ion:165 Resp: 37796
Ion Ratio Lower Upper
165 100
63 59.1 47.0 70.4
89 61.0 47.4 71.2



#52
Acenaphthene
Concen: 5.225 ng
RT: 9.945 min Scan# 1335
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:154 Resp: 168975
Ion Ratio Lower Upper
154 100
153 111.0 90.7 136.1
152 50.2 41.1 61.7





#53

3-Nitroaniline

Concen: 4.902 ng

RT: 9.857 min Scan# 11

Delta R.T. -0.012 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:138 Resp: 39761

Ion Ratio Lower Upper

138 100

108 12.6 9.3 13.9

92 118.6 93.5 140.3



#55

Dibenzofuran

Concen: 5.481 ng

RT: 10.116 min Scan# 1364

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30




Tgt Ion:168 Resp: 252715


Ion Ratio Lower Upper

168 100

139 39.8 33.9 50.9

169 13.6 11.0 16.4









#56
4-Nitrophenol
Concen: 4.309 ng
RT: 10.004 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:139	Resp:	26252	
Ion	Ratio	Lower	Upper
139	100		
109	81.1	61.6	101.6
65	100.7	79.4	119.4



#57
2,4-Dinitrotoluene
Concen: 4.800 ng
RT: 10.092 min Scan# 1360
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt	Ion:165	Resp:	49913
Ion	Ratio	Lower	Upper
165	100		
63	41.2	35.4	53.2
89	70.9	56.6	84.8
182	2.8	2.2	3.4





#58

Fluorene

Concen: 5.542 ng

RT: 10.457 min Scan# 1422

Delta R.T. -0.012 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:166 Resp: 197266

Ion Ratio Lower Upper

166 100

165 98.0 79.0 118.4

167 13.2 11.0 16.4



#59

2,3,4,6-Tetrachlorophenol

Concen: 4.839 ng

RT: 10.233 min Scan# 1384

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Tgt Ion:232 Resp: 48472




Ion Ratio Lower Upper


232 100

131 44.9 37.3 55.9

130 2.1 1.9 2.9

166 30.3 24.8 37.2









#60
Diethylphthalate
Concen: 5.286 ng
RT: 10.327 min Scan# 1400
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 201554
Ion Ratio Lower Upper
149 100
177 22.1 18.0 27.0
150 12.1 10.2 15.4



#61
4-Chlorophenyl-phenylether
Concen: 5.380 ng
RT: 10.451 min Scan# 1421
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:204 Resp: 99747
Ion Ratio Lower Upper
204 100
206 31.1 26.8 40.2
141 64.5 52.4 78.6









#62
4-Nitroaniline
Concen: 4.945 ng
RT: 10.463 min Scan# 1426
Delta R.T. -0.018 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:138 Resp: 38863
Ion Ratio Lower Upper
138 100
92 50.5 32.6 72.6
108 77.1 65.0 105.0



#63
Azobenzene
Concen: 5.370 ng
RT: 10.610 min Scan# 1448
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion: 77 Resp: 190233
Ion Ratio Lower Upper
77 100
182 25.8 6.6 46.6
105 21.0 1.3 41.3
51 29.1 10.4 50.4





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.398 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:188 Resp: 1002901

Ion Ratio Lower Upper

188 100

94 9.0 6.8 10.2

80 10.2 7.6 11.4



#66

n-Nitrosodiphenylamine

Concen: 5.223 ng

RT: 10.569 min Scan# 1441

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30




Tgt Ion:169 Resp: 169225


Ion Ratio Lower Upper

169 100

168 65.7 53.7 80.5

167 35.5 29.2 43.8









#67
4-Bromophenyl-phenylether
Concen: 4.823 ng
RT: 10.945 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:248 Resp: 60238
Ion Ratio Lower Upper
248 100
250 97.1 77.8 116.8
141 74.7 62.6 93.8



#68
Hexachlorobenzene
Concen: 4.753 ng
RT: 11.004 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:284 Resp: 65620
Ion Ratio Lower Upper
284 100
142 35.6 27.0 40.6
249 31.9 25.2 37.8





#69

Atrazine

Concen: 6.867 ng

RT: 11.092 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:200 Resp: 54251

Ion Ratio Lower Upper

200 100

173 25.1 4.7 44.7

215 47.1 25.4 65.4



#70

Pentachlorophenol

Concen: 3.843 ng

RT: 11.198 min Scan# 1548

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30




Tgt Ion:266 Resp: 32716


Ion Ratio Lower Upper

266 100

268 64.2 51.0 76.6

264 62.8 50.2 75.2









#71
Phenanthrene
Concen: 5.340 ng
RT: 11.421 min Scan# 1586
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:178 Resp: 289580
Ion Ratio Lower Upper
178 100
176 19.0 15.8 23.8
179 15.3 12.6 19.0



#72
Anthracene
Concen: 5.407 ng
RT: 11.474 min Scan# 1595
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:178 Resp: 294198
Ion Ratio Lower Upper
178 100
176 18.1 15.1 22.7
179 15.1 12.6 19.0









#73
 Carbazole
 Concen: 5.329 ng
 RT: 11.627 min Scan# 1621
 Delta R.T. -0.006 min
 Lab File: BF141898.D
 Acq: 10 Mar 2025 11:30




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Tgt Ion:167 Resp: 249634
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.4 26.2
 139 13.3 11.0 16.6



#74
 Di-n-butylphthalate
 Concen: 5.352 ng
 RT: 11.957 min Scan# 1677
 Delta R.T. -0.006 min
 Lab File: BF141898.D
 Acq: 10 Mar 2025 11:30

Tgt Ion:149 Resp: 330999
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.6 11.4
 104 5.0 4.4 6.6





#75

Fluoranthene

Concen: 5.432 ng

RT: 12.610 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:202 Resp: 309273

Ion Ratio Lower Upper

202 100

101 9.8 0.0 29.7

203 17.2 0.0 37.8



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.033 min Scan# 2030

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30




Tgt Ion:240 Resp: 745624


Ion Ratio Lower Upper

240 100

120 10.4 8.4 12.6

236 25.3 20.5 30.7









#77
Benzidine
Concen: 6.138 ng
RT: 12.733 min Scan# 1809
Delta R.T. -0.000 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:184 Resp: 64673
Ion Ratio Lower Upper
184 100
185 14.0 11.5 17.3
183 12.3 9.7 14.5



#78
Pyrene
Concen: 4.941 ng
RT: 12.839 min Scan# 1827
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:202 Resp: 315969
Ion Ratio Lower Upper
202 100
200 20.9 17.3 25.9
203 17.5 14.4 21.6









#79
 Terphenyl-d14
 Concen: 9.924 ng
 RT: 12.980 min Scan# 1851
 Delta R.T. -0.006 min
 Lab File: BF141898.D
 Acq: 10 Mar 2025 11:30




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Tgt Ion:244 Resp: 500607
 Ion Ratio Lower Upper
 244 100
 212 6.6 6.0 9.0
 122 9.7 7.7 11.5



#80
 Butylbenzylphthalate
 Concen: 4.653 ng
 RT: 13.457 min Scan# 1932
 Delta R.T. -0.006 min
 Lab File: BF141898.D
 Acq: 10 Mar 2025 11:30

Tgt Ion:149 Resp: 117256
 Ion Ratio Lower Upper
 149 100
 91 71.4 56.5 84.7
 206 21.4 17.3 25.9









#81
Benzo(a)anthracene
Concen: 5.205 ng
RT: 14.021 min Scan# 2028
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:228 Resp: 255308
Ion Ratio Lower Upper
228 100
226 27.3 21.8 32.8
229 19.6 16.0 24.0



#82
3,3'-Dichlorobenzidine
Concen: 4.865 ng
RT: 13.986 min Scan# 2022
Delta R.T. -0.006 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:252 Resp: 68969
Ion Ratio Lower Upper
252 100
254 64.0 51.0 76.6
126 10.7 8.4 12.6





#83

Chrysene

Concen: 4.837 ng

RT: 14.062 min Scan# 2036

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:228 Resp: 213144

Ion Ratio Lower Upper

228 100

226 29.8 24.1 36.1

229 19.9 15.7 23.5



#84

Bis(2-ethylhexyl)phthalate

Concen: 4.836 ng

RT: 14.015 min Scan# 2027

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30




Tgt Ion:149 Resp: 167165


Ion Ratio Lower Upper

149 100

167 26.0 21.4 32.0

279 4.4 3.8 5.6





#85

Di-n-octyl phthalate

Concen: 4.435 ng

RT: 14.627 min Scan# 2132

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:149 Resp: 213395

Ion Ratio Lower Upper

149 100

167 1.5 1.3 1.9

43 8.4 6.9 10.3



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.504 min Scan# 2280

Delta R.T. -0.006 min

Lab File: BF141898.D

Acq: 10 Mar 2025 11:30




Tgt Ion:264 Resp: 527129


Ion Ratio Lower Upper

264 100

260 23.9 19.1 28.7

265 21.7 17.5 26.3









#87
Indeno(1,2,3-cd)pyrene
Concen: 4.228 ng
RT: 16.980 min Scan# 21
Delta R.T. -0.018 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:276 Resp: 144536
Ion Ratio Lower Upper
276 100
138 23.7 18.4 27.6
277 25.1 20.3 30.5



#88
Benzo(b)fluoranthene
Concen: 5.234 ng
RT: 15.068 min Scan# 2206
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:252 Resp: 187154
Ion Ratio Lower Upper
252 100
253 21.5 17.5 26.3
125 9.1 7.0 10.4









#89
Benzo(k)fluoranthene
Concen: 4.903 ng
RT: 15.098 min Scan# 211
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:252 Resp: 153142
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 9.4 7.2 10.8



#90
Benzo(a)pyrene
Concen: 4.860 ng
RT: 15.439 min Scan# 2269
Delta R.T. -0.012 min
Lab File: BF141898.D
Acq: 10 Mar 2025 11:30

Tgt Ion:252 Resp: 137043
Ion Ratio Lower Upper
252 100
253 21.1 17.4 26.0
125 10.4 7.8 11.8









#91
 Dibenzo(a,h)anthracene
 Concen: 4.252 ng
 RT: 16.998 min Scan# 21
 Delta R.T. -0.024 min
 Lab File: BF141898.D
 Acq: 10 Mar 2025 11:30

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Tgt Ion:278 Resp: 119937
 Ion Ratio Lower Upper
 278 100
 139 14.8 11.4 17.2
 279 23.9 18.8 28.2



#92
 Benzo(g,h,i)perylene
 Concen: 4.294 ng
 RT: 17.427 min Scan# 2607
 Delta R.T. -0.024 min
 Lab File: BF141898.D
 Acq: 10 Mar 2025 11:30

Tgt Ion:276 Resp: 120142
 Ion Ratio Lower Upper
 276 100
 277 23.1 18.7 28.1
 138 19.6 15.8 23.8

