

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1523

Client: ENTACT

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q1523-06MS		Client Sample ID: WC-A1-02-CMS		DataFile: BG064101.D							
Pyridine	500	0	280	ug/Kg	56				20 (10)	160 (109)	
1,4-Dichlorobenzene	500	0	400	ug/Kg	80				70 (55)	130 (125)	
2-Methylphenol	500	0	450	ug/Kg	90				70 (60)	130 (131)	
3+4-Methylphenols	500	0	440	ug/Kg	88				20 (54)	160 (136)	
Hexachloroethane	500	0	440	ug/Kg	88				20 (19)	160 (146)	
Nitrobenzene	500	0	460	ug/Kg	92				70 (62)	130 (112)	
Hexachlorobutadiene	500	0	440	ug/Kg	88				70 (52)	130 (125)	
2,4,6-Trichlorophenol	500	0	520	ug/Kg	104				70 (78)	130 (112)	
2,4,5-Trichlorophenol	500	0	520	ug/Kg	104				70 (71)	130 (111)	
2,4-Dinitrotoluene	500	0	480	ug/Kg	96				70 (74)	130 (137)	
Hexachlorobenzene	500	0	470	ug/Kg	94				70 (72)	130 (115)	
Pentachlorophenol	1000	0	1000	ug/Kg	100				20 (52)	160 (162)	