

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG NO.: Q1523

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/11/2025

Lab File ID: BF141910.D Time Analyzed: 11:51

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	212930	6.881	840081	8.16	479107	9.92
UPPER LIMIT	425860	7.381	1680160	8.663	958214	10.416
LOWER LIMIT	106465	6.381	420041	7.663	239554	9.416
EPA SAMPLE NO.						
01 WC-A1-01-C	132841	6.88	498756	8.16	254340	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.