

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q1523 SAS No.: Q1523 SDG No.: Q1523

Instrument ID: MSVOA_N Calibration Date/Time: 03/11/2025 13:22

Lab File ID: VN085928.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.723	0.746		3.18	20
1,1-Dichloroethene	0.548	0.594		8.39	20
2-Butanone	0.302	0.338		11.92	20
Carbon Tetrachloride	0.550	0.625		13.64	20
Chloroform	1.156	1.221		5.62	20
Benzene	1.484	1.669		12.47	20
1,2-Dichloroethane	0.483	0.541		12.01	20
Trichloroethene	0.364	0.369		1.37	20
Tetrachloroethene	0.387	0.405		4.65	20
Chlorobenzene	1.145	1.212	0.3	5.85	20
1,2-Dichloroethane-d4	0.648	0.719		10.96	20
Dibromofluoromethane	0.327	0.367		12.23	20
Toluene-d8	1.190	1.369		15.04	20
4-Bromofluorobenzene	0.392	0.506		29.08	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.