

Surrogate Summary

SDG No.: Q1523

Client: ENTACT

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1523-01	WC-A1-01-G	1,2-Dichloroethane-d4	50	58.8	118	70 (74)	130 (125)
		Dibromofluoromethane	50	41.1	82	70 (75)	130 (124)
		Toluene-d8	50	48.5	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.9	110	70 (77)	130 (121)
Q1523-04	WC-A1-02-G	1,2-Dichloroethane-d4	50	55.7	111	70 (74)	130 (125)
		Dibromofluoromethane	50	38.3	77	70 (75)	130 (124)
		Toluene-d8	50	47.6	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.9	106	70 (77)	130 (121)
VN0311WBL01	VN0311WBL01	1,2-Dichloroethane-d4	50	56.1	112	70 (74)	130 (125)
		Dibromofluoromethane	50	49.0	98	70 (75)	130 (124)
		Toluene-d8	50	44.5	89	70 (86)	130 (113)
		4-Bromofluorobenzene	50	42.2	84	70 (77)	130 (121)
VN0311WBS01	VN0311WBS01	1,2-Dichloroethane-d4	50	46.9	94	70 (74)	130 (125)
		Dibromofluoromethane	50	46.9	94	70 (75)	130 (124)
		Toluene-d8	50	46.3	93	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.7	97	70 (77)	130 (121)
VN0311WBSD0	VN0311WBSD01	1,2-Dichloroethane-d4	50	47.6	95	70 (74)	130 (125)
		Dibromofluoromethane	50	47.6	95	70 (75)	130 (124)
		Toluene-d8	50	46.6	93	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.2	100	70 (77)	130 (121)

