

Instrument ID: IC-2

Daily Analysis Runlog For Sequence/QC Batch ID # LB134959

Review By	Niha	Review On	3/11/2025 9:15:10 AM
Supervise By	Iwona	Supervise On	3/11/2025 4:16:22 PM
SubDirectory	LB134959	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP112016,WP112017,WP112018,WP112019,WP112020,WP112021,WP112022		
ICV Standard	WP112023		
CCV Standard	WP112231		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP112232		
Chk Standard	WP112024,WP112025		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	02/21/25 11:05	All standards, samples, and	NF/IZ	OK
2	STD2	STD2	CAL2	02/21/25 11:26	QC are filtered through	NF/IZ	OK
3	STD3	STD3	CAL3	02/21/25 11:48	0.45um, filter lot W3160	NF/IZ	OK
4	STD4	STD4	CAL4	02/21/25 12:09		NF/IZ	OK
5	STD5	STD5	CAL5	02/21/25 12:31		NF/IZ	OK
6	STD6	STD6	CAL6	02/21/25 12:52		NF/IZ	OK
7	STD7	STD7	CAL7	02/21/25 13:13		NF/IZ	OK
8	ICV1	ICV1	ICV	02/21/25 13:35		NF/IZ	OK
9	ICB1	ICB1	ICB	02/21/25 13:56		NF/IZ	OK
10	CCV1	CCV1	CCV	03/07/25 10:50		NF/IZ	OK
11	CCB1	CCB1	CCB	03/07/25 11:11		NF/IZ	OK
12	LB134959BLW	LB134959BLW	MB	03/07/25 11:33		NF/IZ	OK
13	LB134959BSW	LB134959BSW	LCS	03/07/25 11:54		NF/IZ	OK
14	Q1525-01	MW10	SAM	03/07/25 12:16	Cl high	NF/IZ	Dilution
15	Q1522-02	TW-WTS-04	SAM	03/07/25 12:37		NF/IZ	OK
16	Q1522-02MS	TW-WTS-04MS	MS	03/07/25 12:59	9.5ml of sample, 0.5mL W3092	NF/IZ	OK
17	Q1522-02MSD	TW-WTS-04MSD	MSD	03/07/25 13:21	9.5ml of sample, 0.5mL W3092	NF/IZ	OK
18	Q1522-02DL	TW-WTS-04DL	SAM	03/07/25 13:42	5X for Cl, but still high	NF/IZ	Dilution

Instrument ID: IC-2

Daily Analysis Runlog For Sequence/QCBatch ID # LB134959

Review By	Niha	Review On	3/11/2025 9:15:10 AM
Supervise By	Iwona	Supervise On	3/11/2025 4:16:22 PM
SubDirectory	LB134959	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP112016,WP112017,WP112018,WP112019,WP112020,WP112021,WP112022		
ICV Standard	WP112023		
CCV Standard	WP112231		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP112232		
Chk Standard	WP112024,WP112025		

19	Q1522-02DL2	TW-WTS-04DL2	SAM	03/07/25 14:04	20X for Cl	NF/IZ	Confirms
20	CCV2	CCV2	CCV	03/07/25 14:25		NF/IZ	OK
21	CCB2	CCB2	CCB	03/07/25 14:47		NF/IZ	OK

Instrument ID: TITRATOR

Daily Analysis Runlog For Sequence/QC Batch ID # LB135065

Review By	Niha	Review On	3/19/2025 9:01:06 AM
Supervise By	Iwona	Supervise On	3/19/2025 9:31:24 AM
SubDirectory	LB135065	Test	Alkalinity
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP112337		
Chk Standard	W3150,W3071,W3107		

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	LB135065BLW	LB135065BLW	MB	03/18/25 09:30	pH=3.87	Niha	OK
2	LB135065BSW	LB135065BSW	LCS	03/18/25 09:35	pH=4.40	Niha	OK
3	Q1505-05	PT-MIN1-WP	SAM	03/18/25 09:40	pH=4.46	Niha	OK
4	Q1525-01	MW10	SAM	03/18/25 09:45	pH=4.46	Niha	OK
5	Q1525-01DUP	MW10DUP	DUP	03/18/25 09:50	pH=4.45	Niha	OK