

TitriSoft Report

Reviewed By: Iwona
On: 3/19/2025 9:31:24 AM
Inst Id: Titroline Alpha
Titrator

Ident	Amount	DateTime	Alkalinity	Final pH (sample)	Initial pH (sample)	slope	offset
LB135065BLW	100	18/03/2025 09:30	0.00	3.87	3.91	102.5	6.86
LB135065BSW	100	18/03/2025 09:35	44.96	4.40	8.70	102.5	6.86
Q1505-05	50	18/03/2025 09:40	344.44	4.46	8.32	102.5	6.86
Q1525-01	50	18/03/2025 09:45	153.62	4.46	7.68	102.5	6.86
Q1525-01DUP	50	18/03/2025 09:50	153.68	4.45	7.68	102.5	6.86

03/18/2025

03.18.2025

NF