

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q1525

Project: DPW

RunNo.: LB134959

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.2	10	102	90-110	02/21/2025
Chloride	mg/L	3	3	100	90-110	02/21/2025
Fluoride	mg/L	2	2	100	90-110	02/21/2025
Nitrite	mg/L	3	3	100	90-110	02/21/2025
Nitrate	mg/L	2.5	2.5	100	90-110	02/21/2025
Sulfate	mg/L	15.1	15	101	90-110	02/21/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	02/21/2025
Sample ID: CCV1						
Bromide	mg/L	10.3	10	103	90-110	03/07/2025
Chloride	mg/L	3	3	100	90-110	03/07/2025
Fluoride	mg/L	2	2	100	90-110	03/07/2025
Nitrite	mg/L	3	3	100	90-110	03/07/2025
Nitrate	mg/L	2.6	2.5	104	90-110	03/07/2025
Sulfate	mg/L	15.2	15	101	90-110	03/07/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	03/07/2025
Sample ID: CCV2						
Bromide	mg/L	10.3	10	103	90-110	03/07/2025
Chloride	mg/L	3.1	3	103	90-110	03/07/2025
Fluoride	mg/L	2	2	100	90-110	03/07/2025
Nitrite	mg/L	3.1	3	103	90-110	03/07/2025
Nitrate	mg/L	2.6	2.5	104	90-110	03/07/2025
Sulfate	mg/L	15.1	15	101	90-110	03/07/2025
Orthophosphate as P	mg/L	5	5	100	90-110	03/07/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q1525

Project: DPW

RunNo.: LB134959

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------