

Duplicate Sample Summary

Client: G Environmental	SDG No.: Q1525
Project: DPW	Sample ID: Q1522-02
Client ID: TW-WTS-04MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Orthophosphate as P	mg/L	+/-20	5.20		5.20		1	0		03/07/2025
Chloride	mg/L	+/-20	82.7	OR	82.7	OR	1	0		03/07/2025
Sulfate	mg/L	+/-20	375	OR	375	OR	1	0		03/07/2025
Nitrate	mg/L	+/-20	14.6	OR	14.7	OR	1	1		03/07/2025
Nitrite	mg/L	+/-20	11.2	OR	11.4	OR	1	2		03/07/2025
Fluoride	mg/L	+/-20	3.20		3.30		1	3		03/07/2025
Bromide	mg/L	+/-20	10.2		10.8		1	6		03/07/2025

Duplicate Sample Summary

Client:	G Environmental	SDG No.:	Q1525
Project:	DPW	Sample ID:	Q1525-01
Client ID:	MW10DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Alkalinity	mg/L	+/-20	154		154		1	0		03/18/2025