

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** G Environmental

**SDG No.:** Q1525

**Contract:** GENV01

**Lab Code:** CHEM

**Case No.:** Q1525

**SAS No.:** Q1525

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interement Correction Factors For: |           |           |           |           |
|-----------|-------------------------|----------------------------------------|-----------|-----------|-----------|-----------|
|           |                         | Al                                     | Ca        | Fe        | Mg        | Ag        |
| Iron      | 240.488                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Manganese | 257.610                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Sodium    | 589.592                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** G Environmental

**SDG No.:** Q1525

**Contract:** GENV01

**Lab Code:** CHEM

**Case No.:** Q1525

**SAS No.:** Q1525

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |           |           |           |            |
|-----------|-------------------------|------------------------------------------|-----------|-----------|-----------|------------|
|           |                         | As                                       | Ba        | Be        | Cd        | Co         |
| Iron      | 240.488                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | -0.0039600 |
| Manganese | 257.610                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Sodium    | 589.592                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** G Environmental

**SDG No.:** Q1525

**Contract:** GENV01

**Lab Code:** CHEM

**Case No.:** Q1525

**SAS No.:** Q1525

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |           |           |           |            |
|-----------|-------------------------|------------------------------------------|-----------|-----------|-----------|------------|
|           |                         | Cr                                       | Cu        | K         | Mn        | Mo         |
| Iron      | 240.488                 | 0.0000000                                | 0.0000000 | 0.0000730 | 0.0000000 | -0.0015250 |
| Manganese | 257.610                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Sodium    | 589.592                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** G Environmental

**SDG No.:** Q1525

**Contract:** GENV01

**Lab Code:** CHEM

**Case No.:** Q1525

**SAS No.:** Q1525

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |            |           |           |           |
|-----------|-------------------------|------------------------------------------|------------|-----------|-----------|-----------|
|           |                         | Na                                       | Ni         | Pb        | Sb        | Se        |
| Iron      | 240.488                 | 0.0000000                                | -0.0017000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Manganese | 257.610                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Sodium    | 589.592                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** G Environmental

**SDG No.:** Q1525

**Contract:** GENV01

**Lab Code:** CHEM

**Case No.:** Q1525

**SAS No.:** Q1525

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte   | Wave-<br>Length<br>(nm) | ICP Interement Correction Factors For: |           |           |           |           |
|-----------|-------------------------|----------------------------------------|-----------|-----------|-----------|-----------|
|           |                         | Sn                                     | Ti        | Tl        | V         | Zn        |
| Iron      | 240.488                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Manganese | 257.610                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Sodium    | 589.592                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000 |