



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q1525

Contract: GENV01

Lab Code: CHEM

Case No.: Q1525

SAS No.: Q1525

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	98.7	100	99	40 - 160	P	03/12/2025	11:05	LB135011
	Manganese	19.2	20.0	96	40 - 160	P	03/12/2025	11:05	LB135011
	Sodium	2100	2000	105	40 - 160	P	03/12/2025	11:05	LB135011
CRI01	Iron	110	100	110	40 - 160	P	03/13/2025	12:21	LB135035
	Manganese	20.6	20.0	103	40 - 160	P	03/13/2025	12:21	LB135035
	Sodium	1830	2000	91	40 - 160	P	03/13/2025	12:21	LB135035