

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: G Environmental **SDG No.:** Q1525
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1525 **SAS No.:** Q1525
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	104000	101000	103	85600	116500	03/12/2025	11:09	LB135011
	Manganese	2.80	7.0	40	-13	27	03/12/2025	11:09	LB135011
	Sodium	41.4			0	0	03/12/2025	11:09	LB135011
ICSAB01	Iron	102000	99300	103	84400	114500	03/12/2025	11:14	LB135011
	Manganese	451	507	89	430	584	03/12/2025	11:14	LB135011
	Sodium	43.6			0	0	03/12/2025	11:14	LB135011
ICSA01	Iron	103000	101000	102	85600	116500	03/13/2025	12:25	LB135035
	Manganese	3.77	7.0	54	-13	27	03/13/2025	12:25	LB135035
	Sodium	81.3			0	0	03/13/2025	12:25	LB135035
ICSAB01	Iron	100000	99300	101	84400	114500	03/13/2025	12:30	LB135035
	Manganese	476	507	94	430	584	03/13/2025	12:30	LB135035
	Sodium	93.6			0	0	03/13/2025	12:30	LB135035