

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031125\
 Data File : VX045211.D
 Acq On : 11 Mar 2025 09:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 12 01:41:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.629	0.659	-4.8	103	0.00
3 P	Chloromethane	0.770	0.731	5.1	95	0.00
4 C	Vinyl Chloride	0.764	0.723	5.4#	93	0.00
5 T	Bromomethane	0.300	0.303	-1.0	102	0.00
6 T	Chloroethane	0.352	0.386	-9.7	102	0.00
7 T	Trichlorofluoromethane	1.011	1.035	-2.4	97	0.00
8 T	Diethyl Ether	0.396	0.369	6.8	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.581	0.644	-10.8	106	0.00
10 T	Methyl Iodide	0.726	0.769	-5.9	98	0.00
11 T	Tert butyl alcohol	0.151	0.117	22.5	86	0.00
12 CM	1,1-Dichloroethene	0.614	0.615	-0.2#	97	0.00
13 T	Acrolein	0.174	0.153	12.1	87	0.00
14 T	Allyl chloride	1.094	1.165	-6.5	103	0.00
15 T	Acrylonitrile	0.404	0.413	-2.2	99	0.00
16 T	Acetone	0.369	0.381	-3.3	106	0.00
17 T	Carbon Disulfide	1.636	1.547	5.4	92	0.00
18 T	Methyl Acetate	0.888	1.002	-12.8	114	0.00
19 T	Methyl tert-butyl Ether	2.061	2.061	0.0	97	0.00
20 T	Methylene Chloride	0.731	0.714	2.3	100	0.00
21 T	trans-1,2-Dichloroethene	0.607	0.616	-1.5	96	0.00
22 T	Diisopropyl ether	2.224	2.299	-3.4	99	0.00
23 T	Vinyl Acetate	1.856	1.965	-5.9	98	0.00
24 P	1,1-Dichloroethane	1.247	1.245	0.2	98	0.00
25 T	2-Butanone	0.555	0.581	-4.7	99	-0.01
26 T	2,2-Dichloropropane	0.585	0.905	-54.7#	149	0.00
27 T	cis-1,2-Dichloroethene	0.749	0.755	-0.8	97	0.00
28 T	Bromochloromethane	0.594	0.579	2.5	94	0.00
29 T	Tetrahydrofuran	0.373	0.367	1.6	97	0.00
30 C	Chloroform	1.239	1.234	0.4#	97	0.00
31 T	Cyclohexane	1.082	1.111	-2.7	99	0.00
32 T	1,1,1-Trichloroethane	1.000	1.017	-1.7	98	0.00
33 S	1,2-Dichloroethane-d4	0.795	0.792	0.4	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.334	0.355	-6.3	102	-0.01
36 T	1,1-Dichloropropene	0.459	0.476	-3.7	96	0.00
37 T	Ethyl Acetate	0.591	0.613	-3.7	96	0.00
38 T	Carbon Tetrachloride	0.465	0.493	-6.0	100	0.00
39 T	Methylcyclohexane	0.549	0.622	-13.3	101	0.00
40 TM	Benzene	1.448	1.486	-2.6	94	0.00
41 T	Methacrylonitrile	0.319	0.341	-6.9	95	0.00
42 TM	1,2-Dichloroethane	0.521	0.546	-4.8	98	0.00
43 T	Isopropyl Acetate	0.874	0.953	-9.0	97	0.00
44 TM	Trichloroethene	0.340	0.354	-4.1	98	0.00
45 C	1,2-Dichloropropane	0.372	0.380	-2.2#	97	0.00
46 T	Dibromomethane	0.263	0.278	-5.7	98	0.00
47 T	Bromodichloromethane	0.516	0.557	-7.9	101	0.00
48 T	Methyl methacrylate	0.438	0.470	-7.3	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031125\
 Data File : VX045211.D
 Acq On : 11 Mar 2025 09:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 12 01:41:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	84	0.00
50 S	Toluene-d8	1.212	1.251	-3.2	98	0.00
51 T	4-Methyl-2-Pentanone	0.587	0.636	-8.3	99	0.00
52 CM	Toluene	0.849	0.900	-6.0#	95	0.00
53 T	t-1,3-Dichloropropene	0.431	0.514	-19.3	104	0.00
54 T	cis-1,3-Dichloropropene	0.503	0.580	-15.3	103	0.00
55 T	1,1,2-Trichloroethane	0.350	0.361	-3.1	96	0.00
56 T	Ethyl methacrylate	0.532	0.594	-11.7	99	0.00
57 T	1,3-Dichloropropane	0.605	0.631	-4.3	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.302	-16.6	101	0.00
59 T	2-Hexanone	0.425	0.458	-7.8	97	0.00
60 T	Dibromochloromethane	0.366	0.404	-10.4	100	0.00
61 T	1,2-Dibromoethane	0.349	0.373	-6.9	99	0.00
62 S	4-Bromofluorobenzene	0.402	0.446	-10.9	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.320	0.336	-5.0	98	0.00
65 PM	Chlorobenzene	1.058	1.120	-5.9	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.352	0.372	-5.7	99	0.00
67 C	Ethyl Benzene	1.843	1.996	-8.3#	97	0.00
68 T	m/p-Xylenes	0.675	0.738	-9.3	97	0.00
69 T	o-Xylene	0.681	0.724	-6.3	97	0.00
70 T	Styrene	1.101	1.222	-11.0	98	0.00
71 P	Bromoform	0.266	0.306	-15.0	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.903	4.178	-7.0	99	0.00
74 T	N-amyl acetate	1.831	1.958	-6.9	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.432	1.392	2.8	97	0.00
76 T	1,2,3-Trichloropropane	1.164	1.182	-1.5	101	0.00
77 T	Bromobenzene	0.921	0.953	-3.5	100	0.00
78 T	n-propylbenzene	4.409	4.921	-11.6	101	0.00
79 T	2-Chlorotoluene	2.808	2.904	-3.4	100	0.00
80 T	1,3,5-Trimethylbenzene	3.156	3.410	-8.0	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.340	0.405	-19.1	118	0.00
82 T	4-Chlorotoluene	3.060	3.281	-7.2	100	0.00
83 T	tert-Butylbenzene	3.242	3.424	-5.6	100	0.00
84 T	1,2,4-Trimethylbenzene	3.175	3.431	-8.1	100	0.00
85 T	sec-Butylbenzene	3.884	4.364	-12.4	102	0.00
86 T	p-Isopropyltoluene	3.139	3.522	-12.2	101	0.00
87 T	1,3-Dichlorobenzene	1.643	1.694	-3.1	100	0.00
88 T	1,4-Dichlorobenzene	1.662	1.696	-2.0	99	0.00
89 T	n-Butylbenzene	2.703	3.210	-18.8	105	0.00
90 T	Hexachloroethane	0.568	0.625	-10.0	102	0.00
91 T	1,2-Dichlorobenzene	1.648	1.671	-1.4	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.293	-5.0	100	0.00
93 T	1,2,4-Trichlorobenzene	0.938	1.037	-10.6	101	0.00
94 T	Hexachlorobutadiene	0.385	0.438	-13.8	110	0.00
95 T	Naphthalene	3.605	3.723	-3.3	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031125\
Data File : VX045211.D
Acq On : 11 Mar 2025 09:59
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 12 01:41:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.990	1.037	-4.7	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6