

Report of Analysis

Client:	G Environmental	Date Collected:	
Project:	DPW	Date Received:	
Client Sample ID:	VX0311WBSD01	SDG No.:	Q1525
Lab Sample ID:	VX0311WBSD01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045215.D	1		03/11/25 11:40	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
108-90-7	Chlorobenzene	19.6		0.13	1.00	ug/L
100-41-4	Ethyl Benzene	19.7		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	39.9		0.31	2.00	ug/L
95-47-6	o-Xylene	20.2		0.14	1.00	ug/L
100-42-5	Styrene	20.5		0.16	1.00	ug/L
75-25-2	Bromoform	20.3		0.21	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.7		0.27	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.1		70 (74) - 130 (125)	102%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0		70 (75) - 130 (124)	102%	SPK: 50
2037-26-5	Toluene-d8	50.7		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.4		70 (77) - 130 (121)	109%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	92100	5.544			
540-36-3	1,4-Difluorobenzene	166000	6.757			
3114-55-4	Chlorobenzene-d5	149000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	68900	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products