

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1525 SAS No.: Q1525 SDG NO.: Q1525
 Lab File ID: VX045211.D Date Analyzed: 03/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:59
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	101357	5.54	176548	6.75	153652	10.05
UPPER LIMIT	202714	6.044	353096	7.251	307304	10.549
LOWER LIMIT	50678.5	5.044	88274	6.251	76826	9.549
EPA SAMPLE NO.						
MW10	69449	5.55	133868	6.76	123702	10.05
VX0311WBL01	69956	5.54	139206	6.76	126312	10.06
VX0311WBS01	94786	5.54	172566	6.76	151770	10.06
VX0311WBSD01	92082	5.54	166231	6.76	148942	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.