

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1525 SAS No.: Q1525 SDG NO.: Q1525
 Lab File ID: VX045211.D Date Analyzed: 03/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:59
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	70136	12.018				
UPPER LIMIT	140272	12.518				
LOWER LIMIT	35068	11.518				
EPA SAMPLE NO.						
MW10	55461	12.02				
VX0311WBL01	50068	12.02				
VX0311WBS01	71836	12.02				
VX0311WBSD01	68861	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.