

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/26/2025 02/26/2025

Client Sample No.: CCAL01 Date Analyzed: 03/11/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070025.D Time Analyzed: 13:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.487	4.385	4.585	265191200.000	213707000.000	-19.4
Aroclor-1016-2	4.506	4.403	4.603	455488600.000	395497000.000	-13.2
Aroclor-1016-3	4.564	4.462	4.662	299813200.000	262504000.000	-12.4
Aroclor-1016-4	4.653	4.551	4.751	237375700.000	206137000.000	-13.2
Aroclor-1016-5	4.938	4.835	5.035	232879600.000	209292000.000	-10.1
Aroclor-1260-1	6.028	5.926	6.126	397928000.000	361596000.000	-9.1
Aroclor-1260-2	6.284	6.182	6.382	429095700.000	367866000.000	-14.3
Aroclor-1260-3	6.635	6.532	6.732	347609000.000	316371000.000	-9.0
Aroclor-1260-4	6.849	6.746	6.946	376519100.000	335108000.000	-11.0
Aroclor-1260-5	7.155	7.052	7.252	662059400.000	552079000.000	-16.6
DCB	8.508	8.404	8.604	4467595000.000	3881080000.000	-13.1
TCX	3.406	3.304	3.504	10442500000.000	9785720000.000	-6.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/26/2025 02/26/2025

Client Sample No.: CCAL01 Date Analyzed: 03/11/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070025.D Time Analyzed: 13:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.788	3.687	3.887	217195800.000	199348000.000	-8.2
Aroclor-1016-2	3.803	3.702	3.902	408040000.000	396684000.000	-2.8
Aroclor-1016-3	3.963	3.863	4.063	204902400.000	201701000.000	-1.6
Aroclor-1016-4	4.011	3.911	4.111	189779300.000	196125000.000	3.3
Aroclor-1016-5	4.207	4.106	4.306	220863400.000	223928000.000	1.4
Aroclor-1260-1	5.201	5.100	5.300	398528300.000	400167000.000	0.4
Aroclor-1260-2	5.390	5.290	5.490	461970500.000	445512000.000	-3.6
Aroclor-1260-3	5.532	5.431	5.631	439916700.000	430945000.000	-2.0
Aroclor-1260-4	5.993	5.893	6.093	390238500.000	379919000.000	-2.6
Aroclor-1260-5	6.237	6.137	6.337	814273700.000	746575000.000	-8.3
DCB	7.551	7.450	7.650	5951967000.000	5618440000.000	-5.6
TCX	2.784	2.684	2.884	8098077000.000	8267280000.000	2.1