

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/26/2025 02/26/2025

Client Sample No.: CCAL02 Date Analyzed: 03/11/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070029.D Time Analyzed: 14:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.485	4.385	4.585	265191200.000	220273000.000	-16.9
Aroclor-1016-2	4.504	4.403	4.603	455488600.000	387949000.000	-14.8
Aroclor-1016-3	4.562	4.462	4.662	299813200.000	251400000.000	-16.1
Aroclor-1016-4	4.651	4.551	4.751	237375700.000	200364000.000	-15.6
Aroclor-1016-5	4.935	4.835	5.035	232879600.000	202352000.000	-13.1
Aroclor-1260-1	6.026	5.926	6.126	397928000.000	349151000.000	-12.3
Aroclor-1260-2	6.282	6.182	6.382	429095700.000	373167000.000	-13.0
Aroclor-1260-3	6.633	6.532	6.732	347609000.000	305208000.000	-12.2
Aroclor-1260-4	6.847	6.746	6.946	376519100.000	327373000.000	-13.1
Aroclor-1260-5	7.153	7.052	7.252	662059400.000	566745000.000	-14.4
DCB	8.505	8.404	8.604	4467595000.000	3797680000.000	-15.0
TCX	3.405	3.304	3.504	10442500000.000	9299630000.000	-11.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/26/2025 02/26/2025

Client Sample No.: CCAL02 Date Analyzed: 03/11/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070029.D Time Analyzed: 14:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.787	3.687	3.887	217195800.000	207889000.000	-4.3
Aroclor-1016-2	3.802	3.702	3.902	408040000.000	386521000.000	-5.3
Aroclor-1016-3	3.963	3.863	4.063	204902400.000	196683000.000	-4.0
Aroclor-1016-4	4.011	3.911	4.111	189779300.000	181362000.000	-4.4
Aroclor-1016-5	4.207	4.106	4.306	220863400.000	213045000.000	-3.5
Aroclor-1260-1	5.200	5.100	5.300	398528300.000	385796000.000	-3.2
Aroclor-1260-2	5.390	5.290	5.490	461970500.000	443279000.000	-4.0
Aroclor-1260-3	5.531	5.431	5.631	439916700.000	422368000.000	-4.0
Aroclor-1260-4	5.992	5.893	6.093	390238500.000	373552000.000	-4.3
Aroclor-1260-5	6.237	6.137	6.337	814273700.000	773151000.000	-5.1
DCB	7.550	7.450	7.650	5951967000.000	5524310000.000	-7.2
TCX	2.784	2.684	2.884	8098077000.000	7795610000.000	-3.8