

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q1526	OrderDate:	3/7/2025 11:53:00 AM
Client:	Chemtech Consulting Group	Project:	Weekly Storage Blanks
Contact:	QA Officer	Location:	F11,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1526-11	PEST-GPC2-BLANK	SOIL	Pesticide-TCL	SFAM_PEST	03/07/25	03/10/25	03/12/25	03/07/25
Q1526-12	PEST-GPC2-BLANK-SP IKE	SOIL	Pesticide-TCL	SFAM_PEST	03/07/25	03/10/25	03/12/25	03/07/25
Q1526-13	PCB-GPC2-BLANK	SOIL	PCB	SFAM_PCB	03/07/25	03/10/25	03/11/25	03/07/25
Q1526-14	PCB-GPC2-BLANK-SPI KE	SOIL	PCB	SFAM_PCB	03/07/25	03/10/25	03/11/25	03/07/25

Hit Summary Sheet
SW-846

SDG No.: Q1526

Order ID: Q1526

Client: Chemtech Consulting Group

Project ID: Weekly Storage Blanks

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : PEST-GPC2-BLANK-SPIKE								
Q1526-12	PEST-GPC2-BLANK· SOIL		gamma-BHC (Lindane)	16.0		0.48	5.10	ug/kg
Q1526-12	PEST-GPC2-BLANK· SOIL		Heptachlor	16.0		0.57	5.10	ug/kg
Q1526-12	PEST-GPC2-BLANK· SOIL		Aldrin	16.0		0.39	5.10	ug/kg
Q1526-12	PEST-GPC2-BLANK· SOIL		Dieldrin	34.0		0.45	9.90	ug/kg
Q1526-12	PEST-GPC2-BLANK· SOIL		Endrin	34.0		0.63	9.90	ug/kg
Q1526-12	PEST-GPC2-BLANK· SOIL		4,4-DDT	35.0		0.63	9.90	ug/kg
Total Concentration:				151.000				



QC SUMMARY

Surrogate Summary

SDG No.: Q1526

Client: Chemtech Consulting Group

Analytical Method: SFAM_PEST

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD087906.D	PIBLK-PD087906.D	TCX	1	20	0.23	113		30	150
		DCB	1	40	0.35	87		30	150
		TCX	2	20	0.23	113		30	150
		DCB	2	40	0.34	84		30	150
I.BLK-PD087950.D	PIBLK-PD087950.D	TCX	1	20	0.22	108		30	150
		DCB	1	40	0.33	81		30	150
		TCX	2	20	0.26	131		30	150
		DCB	2	40	0.38	95		30	150
Q1526-11	PEST-GPC2-BLANK	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
Q1526-12	PEST-GPC2-BLANK-SPIKE	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
I.BLK-PD087954.D	PIBLK-PD087954.D	TCX	1	20	0.21	106		30	150
		DCB	1	40	0.32	81		30	150
		TCX	2	20	0.25	127		30	150
		DCB	2	40	0.37	93		30	150



SAMPLE DATA

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	03/07/25
Project:	Weekly Storage Blanks	Date Received:	03/07/25
Client Sample ID:	PEST-GPC2-BLANK	SDG No.:	Q1526
Lab Sample ID:	Q1526-11	Matrix:	SOIL
Analytical Method:	SFAM_PEST	% Solid:	100
Sample Wt/Vol:	10 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	Pesticide-TCL
Extraction Type:		Injection Volume :	
GPC Factor :	2.0	PH :	
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD087952.D	1	03/10/25 10:34	03/12/25 09:16	PB167055

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.39	U	0.39	5.10	ug/kg
319-85-7	beta-BHC	0.39	U	0.39	5.10	ug/kg
319-86-8	delta-BHC	1.40	U	1.40	5.10	ug/kg
58-89-9	gamma-BHC (Lindane)	0.48	U	0.48	5.10	ug/kg
76-44-8	Heptachlor	0.57	U	0.57	5.10	ug/kg
309-00-2	Aldrin	0.39	U	0.39	5.10	ug/kg
1024-57-3	Heptachlor epoxide	0.42	U	0.42	5.10	ug/kg
959-98-8	Endosulfan I	0.39	U	0.39	5.10	ug/kg
60-57-1	Dieldrin	0.45	U	0.45	9.90	ug/kg
72-55-9	4,4-DDE	0.48	U	0.48	9.90	ug/kg
72-20-8	Endrin	0.63	U	0.63	9.90	ug/kg
33213-65-9	Endosulfan II	0.57	U	0.57	9.90	ug/kg
72-54-8	4,4-DDD	0.57	U	0.57	9.90	ug/kg
1031-07-8	Endosulfan Sulfate	0.60	U	0.60	9.90	ug/kg
50-29-3	4,4-DDT	0.63	U	0.63	9.90	ug/kg
72-43-5	Methoxychlor	0.75	U	0.75	51.0	ug/kg
53494-70-5	Endrin ketone	0.45	U	0.45	9.90	ug/kg
7421-93-4	Endrin Aldehyde	0.54	U	0.54	9.90	ug/kg
5103-71-9	cis-Chlordane	0.39	U	0.39	5.10	ug/kg
5103-74-2	trans-Chlordane	0.39	U	0.39	5.10	ug/kg
8001-35-2	Toxaphene	24.0	U	24.0	510	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	30 - 150	%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	30 - 150	%	SPK: 40

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	03/07/25
Project:	Weekly Storage Blanks	Date Received:	03/07/25
Client Sample ID:	PEST-GPC2-BLANK	SDG No.:	Q1526
Lab Sample ID:	Q1526-11	Matrix:	SOIL
Analytical Method:	SFAM_PEST	% Solid:	100
Sample Wt/Vol:	10 Units: g	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	5000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	2.0	Injection Volume :	
PH :			
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD087952.D	1	03/10/25 10:34	03/12/25 09:16	PB167055

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
Data File : PD087952.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 09:16
Operator : AR\AJ
Sample : Q1526-11
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 10:15:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 15:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
Data File : PD087952.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 09:16
Operator : AR\AJ
Sample : Q1526-11
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampled :
PEST-GPC2-BLANK


Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 10:15:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 15:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_

Signal: PD087952.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T.: 3.553 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK

Time

Response_

Signal: PD087952.D\ECD2B.ch




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 2.884 min
Response: 0
Conc: N.D.

Time

Response_

Signal: PD087952.D\ECD1A.ch




#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T.: 9.079 min
Response: 0
Conc: N.D.

Time

Response_

Signal: PD087952.D\ECD2B.ch



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T.: 8.080 min
Response: 0
Conc: N.D.

Time

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	03/07/25
Project:	Weekly Storage Blanks	Date Received:	03/07/25
Client Sample ID:	PEST-GPC2-BLANK-SPIKE	SDG No.:	Q1526
Lab Sample ID:	Q1526-12	Matrix:	SOIL
Analytical Method:	SFAM_PEST	% Solid:	100
Sample Wt/Vol:	10 Units: g	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	5000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	2.0	Injection Volume :	
PH :			
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD087953.D	1	03/10/25 10:34	03/12/25 10:37	PB167055

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
319-84-6	alpha-BHC	0.39	U	0.39	5.10	ug/kg
319-85-7	beta-BHC	0.39	U	0.39	5.10	ug/kg
319-86-8	delta-BHC	1.40	U	1.40	5.10	ug/kg
58-89-9	gamma-BHC (Lindane)	16.0		0.48	5.10	ug/kg
76-44-8	Heptachlor	16.0		0.57	5.10	ug/kg
309-00-2	Aldrin	16.0		0.39	5.10	ug/kg
1024-57-3	Heptachlor epoxide	0.42	U	0.42	5.10	ug/kg
959-98-8	Endosulfan I	0.39	U	0.39	5.10	ug/kg
60-57-1	Dieldrin	34.0		0.45	9.90	ug/kg
72-55-9	4,4-DDE	0.48	U	0.48	9.90	ug/kg
72-20-8	Endrin	34.0		0.63	9.90	ug/kg
33213-65-9	Endosulfan II	0.57	U	0.57	9.90	ug/kg
72-54-8	4,4-DDD	0.57	U	0.57	9.90	ug/kg
1031-07-8	Endosulfan Sulfate	0.60	U	0.60	9.90	ug/kg
50-29-3	4,4-DDT	35.0		0.63	9.90	ug/kg
72-43-5	Methoxychlor	0.75	U	0.75	51.0	ug/kg
53494-70-5	Endrin ketone	0.45	U	0.45	9.90	ug/kg
7421-93-4	Endrin Aldehyde	0.54	U	0.54	9.90	ug/kg
5103-71-9	cis-Chlordane	0.39	U	0.39	5.10	ug/kg
5103-74-2	trans-Chlordane	0.39	U	0.39	5.10	ug/kg
8001-35-2	Toxaphene	24.0	U	24.0	510	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	30 - 150	%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	30 - 150	%	SPK: 40

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	03/07/25
Project:	Weekly Storage Blanks	Date Received:	03/07/25
Client Sample ID:	PEST-GPC2-BLANK-SPIKE	SDG No.:	Q1526
Lab Sample ID:	Q1526-12	Matrix:	SOIL
Analytical Method:	SFAM_PEST	% Solid:	100
Sample Wt/Vol:	10 Units: g	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	5000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	2.0	Injection Volume :	
PH :			
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD087953.D	1	03/10/25 10:34	03/12/25 10:37	PB167055

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
 Data File : PD087953.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2025 10:37
 Operator : AR\AJ
 Sample : Q1526-12
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEST-GPC2-BLANK-SPIKE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 10:03:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 24 15:08:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

3) MA	gamma-BHC...	4.338	3.733	66872573	381.6E6	16.305	19.819
4) MA	Heptachlor	4.938	4.088	66128414	377.2E6	16.429	18.981
5) MB	Aldrin	5.280	4.374	65120331	388.9E6	16.338	19.830
13) MA	Dieldrin	6.356	5.519	122.3E6	681.3E6	33.571	38.046
14) MA	Endrin	6.583	5.796	101.8E6	597.2E6	33.601	37.361
17) MA	4,4'-DDT	7.031	6.191	92472250	589.2E6	34.690	40.239


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
Data File : PD087953.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 10:37
Operator : AR\AJ
Sample : Q1526-12
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampled :
PEST-GPC2-BLANK-SPIKE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 10:03:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 15:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm




Response_

Signal: PD087953.D\IECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 3.553 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK-SPIKE




Time

Response_

Signal: PD087953.D\IECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 2.884 min
Response: 0
Conc: N.D.




Time

Response_

Signal: PD087953.D\IECD1A.ch

#3 gamma-BHC (Lindane)

R.T.: 4.338 min
Delta R.T.: 0.005 min
Response: 66872573
Conc: 16.30 ng/ml




Time

Response_

Signal: PD087953.D\IECD2B.ch

#3 gamma-BHC (Lindane)


R.T.: 3.733 min
Delta R.T.: 0.000 min
Response: 381638785
Conc: 19.82 ng/ml



Time

Response_

Signal: PD087953.D\IECD1A.ch



#4 Heptachlor


R.T.: 4.938 min
Delta R.T.: 0.005 min
Response: 66128414
Conc: 16.43 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK-SPIKE

Time

Response_

Signal: PD087953.D\IECD2B.ch




#4 Heptachlor

R.T.: 4.088 min
Delta R.T.: 0.000 min
Response: 377154011
Conc: 18.98 ng/ml

Time

Response_

Signal: PD087953.D\IECD1A.ch




#5 Aldrin

R.T.: 5.280 min
Delta R.T.: 0.006 min
Response: 65120331
Conc: 16.34 ng/ml

Time

Response_

Signal: PD087953.D\IECD2B.ch



#5 Aldrin

R.T.: 4.374 min
Delta R.T.: 0.000 min
Response: 388899073
Conc: 19.83 ng/ml

Time

Response_

Signal: PD087953.D\IECD1A.ch

#13 Dieldrin

R.T.: 6.356 min

Delta R.T.: 0.005 min

Response: 122339828


Conc: 33.57 ng/ml

Instrument :

ECD_D

ClientSampleId :

PEST-GPC2-BLANK-SPIKE



Time

Response_

Signal: PD087953.D\IECD2B.ch


#13 Dieldrin

R.T.: 5.519 min

Delta R.T.: 0.001 min

Response: 681319540

Conc: 38.05 ng/ml



Time

Response_

Signal: PD087953.D\IECD1A.ch


#14 Endrin

R.T.: 6.583 min

Delta R.T.: 0.005 min

Response: 101757881

Conc: 33.60 ng/ml



Time

Response_

Signal: PD087953.D\IECD2B.ch


#14 Endrin

R.T.: 5.796 min

Delta R.T.: 0.001 min

Response: 597201759

Conc: 37.36 ng/ml



Time


Response_

Signal: PD087953.D\IECD1A.ch

#17 4,4'-DDT

R.T.: 7.031 min
Delta R.T.: 0.006 min
Response: 92472250
Conc: 34.69 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK-SPIKE




Time

Response_

Signal: PD087953.D\IECD2B.ch

#17 4,4'-DDT

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 589230175
Conc: 40.24 ng/ml




Time

Response_

Signal: PD087953.D\IECD1A.ch

#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T.: 9.079 min
Response: 0
Conc: N.D.




Time

Response_

Signal: PD087953.D\IECD2B.ch

#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T.: 8.080 min
Response: 0
Conc: N.D.



Time



CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: INDA320 Data File : PD087955.D Time Analyzed: 12:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.712	6.642	6.782	2568240.000	2354180.000	-8.3
4,4'-DDT	7.029	6.959	7.099	2665660.000	2444830.000	-8.3
alpha-BHC	4.005	3.955	4.055	4131350.000	3755830.000	-9.1
DCB	9.084	8.984	9.184	3294960.000	2981900.000	-9.5
Dieldrin	6.355	6.285	6.425	3644250.000	3400600.000	-6.7
Endosulfan I	6.082	6.012	6.152	3360700.000	3090670.000	-8.0
Endrin	6.582	6.512	6.652	3028440.000	2750360.000	-9.2
gamma-BHC (Lindane)	4.336	4.286	4.386	4101440.000	3679530.000	-10.3
Heptachlor	4.937	4.887	4.987	4025040.000	3738640.000	-7.1
Methoxychlor	7.501	7.431	7.571	1464130.000	1345270.000	-8.1
TCX	3.555	3.505	3.605	2086660.000	1844510.000	-11.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: INDA320 Data File : PD087955.D Time Analyzed: 12:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	5.936	5.866	6.006	14429100.000	14982500.000	3.8
4,4'-DDT	6.191	6.121	6.261	14643200.000	15639000.000	6.8
alpha-BHC	3.397	3.347	3.447	21057100.000	22537000.000	7.0
DCB	8.081	7.981	8.181	15809900.000	16259100.000	2.8
Dieldrin	5.519	5.449	5.589	17907700.000	18866600.000	5.4
Endosulfan I	5.253	5.183	5.323	16918600.000	18099500.000	7.0
Endrin	5.795	5.725	5.865	15984600.000	16338600.000	2.2
gamma-BHC (Lindane)	3.733	3.683	3.783	19255900.000	20745100.000	7.7
Heptachlor	4.088	4.038	4.138	19870200.000	20996800.000	5.7
Methoxychlor	6.762	6.692	6.832	6618460.000	6549350.000	-1.0
TCX	2.883	2.833	2.933	13450900.000	14405500.000	7.1

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
 Data File : PD087955.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2025 12:00
 Operator : AR\AJ
 Sample : INDA320
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3234

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 10:04:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 24 15:08:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.555	2.883	36890144	288.1E6	17.679	21.419
27)	SA Decachlor...	9.084	8.081	119.3E6	650.4E6	36.200	41.137
Target Compounds							
2)	A alpha-BHC	4.005	3.397	75116660	450.7E6	18.182	21.406
3)	MA gamma-BHC...	4.336	3.733	73590622	414.9E6	17.943	21.547
4)	MA Heptachlor	4.937	4.088	74772748	419.9E6	18.577	21.134
9)	A Endosulfan I	6.082	5.253	61813479	362.0E6	18.393	21.396
13)	MA Dieldrin	6.355	5.519	136.0E6	754.7E6	37.326	42.142
14)	MA Endrin	6.582	5.795	110.0E6	653.5E6	36.327	40.886
16)	A 4,4'-DDD	6.712	5.936	94167318	599.3E6	36.666	41.534
17)	MA 4,4'-DDT	7.029	6.191	97793365	625.6E6	36.686	42.720
20)	A Methoxychlor	7.501	6.762	269.1E6	1309.9E6	183.764	197.912


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
Data File : PD087955.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 12:00
Operator : AR\AJ
Sample : INDA320
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
Client Sampled :
INDA3234

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 10:04:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 15:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: INDB320 Data File : PD087956.D Time Analyzed: 12:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDE	6.203	6.133	6.273	3406110.000	3046720.000	-10.6
Aldrin	5.279	5.229	5.329	3985850.000	3527280.000	-11.5
beta-BHC	4.521	4.471	4.571	1752110.000	1572710.000	-10.2
cis-Chlordane	6.034	5.964	6.104	3644450.000	3237750.000	-11.2
DCB	9.084	8.984	9.184	3294960.000	3010760.000	-8.6
delta-BHC	4.768	4.718	4.818	4084440.000	3731350.000	-8.6
Endosulfan II	6.793	6.723	6.863	3152050.000	2852430.000	-9.5
Endosulfan sulfate	7.157	7.087	7.227	2912810.000	2658880.000	-8.7
Endrin aldehyde	6.922	6.852	6.992	2372520.000	2180380.000	-8.1
Endrin ketone	7.638	7.568	7.708	3198110.000	2939700.000	-8.1
Heptachlor epoxide	5.698	5.628	5.768	3686560.000	3328110.000	-9.7
TCX	3.556	3.506	3.606	2086660.000	1877080.000	-10.0
trans-Chlordane	5.953	5.883	6.023	3633010.000	3226880.000	-11.2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: INDB320 Data File : PD087956.D Time Analyzed: 12:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDE	5.382	5.312	5.452	17748400.000	18246600.000	2.8
Aldrin	4.374	4.324	4.424	19611500.000	20504400.000	4.6
beta-BHC	4.028	3.978	4.078	8910290.000	9514260.000	6.8
cis-Chlordane	5.197	5.127	5.267	18492300.000	19140800.000	3.5
DCB	8.080	7.980	8.180	15809900.000	16332800.000	3.3
delta-BHC	4.266	4.216	4.316	20135400.000	20916100.000	3.9
Endosulfan II	6.087	6.017	6.157	15828700.000	16526800.000	4.4
Endosulfan sulfate	6.489	6.419	6.559	15611500.000	16107700.000	3.2
Endrin aldehyde	6.265	6.195	6.335	12445700.000	12861700.000	3.3
Endrin ketone	6.998	6.928	7.068	16825800.000	17553700.000	4.3
Heptachlor epoxide	4.879	4.809	4.949	18226400.000	18917900.000	3.8
TCX	2.884	2.834	2.934	13450900.000	14717000.000	9.4
trans-Chlordane	5.132	5.062	5.202	19167800.000	19869500.000	3.7

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
 Data File : PD087956.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2025 12:15
 Operator : AR\AJ
 Sample : INDB320
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB3235

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/13/2025
 Supervised By : Ankita Jodhani 03/13/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 10:04:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 24 15:08:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.556	2.884	37541539	294.3E6	17.991	21.883
27)	SA Decachlor...	9.084	8.080	120.4E6	653.3E6	36.550	41.323
Target Compounds							
5)	MB Aldrin	5.279	4.374	70545669	410.1E6	17.699	20.911
6)	B beta-BHC	4.521	4.028	31454193	190.3E6	17.952	21.356m
7)	B delta-BHC	4.768	4.266	74626913	418.3E6	18.271m	20.775
8)	B Heptachlo...	5.698	4.879	66562136	378.4E6	18.055	20.759
10)	B trans-Chl...	5.953	5.132	64537631	397.4E6	17.764	20.732
11)	B cis-Chlor...	6.034	5.197	64754993	382.8E6	17.768	20.701
12)	B 4,4'-DDE	6.203	5.382	121.9E6	729.9E6	35.779	41.123
15)	B Endosulfa...	6.793	6.087	114.1E6	661.1E6	36.198	41.764
18)	B Endrin al...	6.922	6.265	87215348	514.5E6	36.761	41.337
19)	B Endosulfa...	7.157	6.489	106.4E6	644.3E6	36.513	41.271
21)	B Endrin ke...	7.638	6.998	117.6E6	702.1E6	36.768	41.730

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
Data File : PD087956.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 12:15
Operator : AR\AJ
Sample : INDB320
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_D
Client Sampled :
INDB3235

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 10:04:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 15:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/13/2025
Supervised By : Ankita Jodhani 03/13/2025

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No. (PEM): PEM - PD087890.D Date Analyzed: 02/18/2025

Lab Sample No.(PEM): PEM067 Time Analyzed: 09:42

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
alpha-BHC	4.001	3.950	4.050	9.930	10.000	-0.7
beta-BHC	4.518	4.470	4.570	10.040	10.000	0.4
gamma-BHC (Lindane)	4.334	4.280	4.380	10.050	10.000	0.5
Endrin	6.579	6.510	6.650	47.780	50.000	-4.4
4,4'-DDT	7.025	6.950	7.100	106.420	100.000	6.4
Methoxychlor	7.497	7.430	7.570	252.270	250.000	0.9
TCX	3.553	3.500	3.600	20.720	20.000	3.6
DCB	9.079	8.980	9.180	21.400	20.000	7.0

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No. (PEM): PEM - PD087890.D Date Analyzed: 02/18/2025

Lab Sample No.(PEM): PEM067 Time Analyzed: 09:42

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
alpha-BHC	3.398	3.350	3.450	10.790	10.000	7.9
beta-BHC	4.030	3.980	4.080	10.600	10.000	6.0
gamma-BHC (Lindane)	3.734	3.680	3.780	10.690	10.000	6.9
Endrin	5.795	5.720	5.870	46.650	50.000	-6.7
4,4'-DDT	6.190	6.120	6.260	97.300	100.000	-2.7
Methoxychlor	6.761	6.690	6.830	237.570	250.000	-5.0
TCX	2.884	2.830	2.930	20.880	20.000	4.4
DCB	8.081	7.980	8.180	21.870	20.000	9.4

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No. (PEM): PEM - PD087907.D Date Analyzed: 02/18/2025

Lab Sample No.(PEM): PEM068 Time Analyzed: 13:34

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
alpha-BHC	4.001	3.950	4.050	10.460	10.000	4.6
beta-BHC	4.518	4.470	4.570	10.670	10.000	6.7
gamma-BHC (Lindane)	4.334	4.280	4.380	10.630	10.000	6.3
Endrin	6.579	6.510	6.650	50.870	50.000	1.7
4,4'-DDT	7.025	6.950	7.100	112.420	100.000	12.4
Methoxychlor	7.497	7.430	7.570	265.780	250.000	6.3
TCX	3.553	3.500	3.600	21.760	20.000	8.8
DCB	9.079	8.980	9.180	22.820	20.000	14.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No. (PEM): PEM - PD087907.D Date Analyzed: 02/18/2025

Lab Sample No.(PEM): PEM068 Time Analyzed: 13:34

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
alpha-BHC	3.397	3.350	3.450	11.320	10.000	13.2
beta-BHC	4.029	3.980	4.080	11.070	10.000	10.7
gamma-BHC (Lindane)	3.734	3.680	3.780	11.200	10.000	12.0
Endrin	5.795	5.720	5.870	49.770	50.000	-0.5
4,4'-DDT	6.190	6.120	6.260	103.160	100.000	3.2
Methoxychlor	6.761	6.690	6.830	262.130	250.000	4.9
TCX	2.884	2.830	2.930	21.720	20.000	8.6
DCB	8.080	7.980	8.180	22.860	20.000	14.3

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No. (PEM): PEM - PD087951.D Date Analyzed: 03/12/2025

Lab Sample No.(PEM): PEM073 Time Analyzed: 09:03

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
alpha-BHC	4.002	3.950	4.050	9.170	10.000	-8.3
beta-BHC	4.518	4.470	4.570	9.360	10.000	-6.4
gamma-BHC (Lindane)	4.333	4.280	4.380	9.270	10.000	-7.3
Endrin	6.580	6.510	6.650	48.230	50.000	-3.5
4,4'-DDT	7.026	6.960	7.100	99.260	100.000	-0.7
Methoxychlor	7.498	7.430	7.570	234.350	250.000	-6.3
TCX	3.553	3.500	3.600	18.860	20.000	-5.7
DCB	9.082	8.980	9.180	19.270	20.000	-3.7

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No. (PEM): PEM - PD087951.D Date Analyzed: 03/12/2025

Lab Sample No.(PEM): PEM073 Time Analyzed: 09:03

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
alpha-BHC	3.397	3.350	3.450	11.820	10.000	18.2
beta-BHC	4.029	3.980	4.080	11.010	10.000	10.1
gamma-BHC (Lindane)	3.733	3.680	3.780	11.780	10.000	17.8
Endrin	5.795	5.720	5.870	49.230	50.000	-1.5
4,4'-DDT	6.190	6.120	6.260	102.290	100.000	2.3
Methoxychlor	6.761	6.690	6.830	234.510	250.000	-6.2
TCX	2.884	2.830	2.930	22.660	20.000	13.3
DCB	8.081	7.980	8.180	23.100	20.000	15.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
 Data File : PD087951.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2025 09:03
 Operator : AR\AJ
 Sample : PEM073
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM056

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/13/2025
 Supervised By : Ankita Jodhani 03/13/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 10:16:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 24 15:08:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.553	2.884	39359489	304.8E6	18.862	22.661
27)	SA Decachlor...	9.082	8.081	63498534	365.2E6	19.271	23.100
Target Compounds							
2)	A alpha-BHC	4.002	3.397	37902455	248.8E6	9.174	11.815 #
3)	MA gamma-BHC...	4.333	3.733	38027326	226.8E6	9.272	11.776 #
6)	B beta-BHC	4.518	4.029	16391786	98094920	9.355	11.009
12)	B 4,4'-DDE	6.203	5.381	368518	1891730	0.108	0.107m
14)	MA Endrin	6.580	5.795	146.1E6	786.9E6	48.233	49.226
16)	A 4,4'-DDD	6.710	5.937	15699213	16553502	6.113	1.147 #
17)	MA 4,4'-DDT	7.026	6.190	264.6E6	1497.8E6	99.257	102.288
18)	B Endrin al...	6.920	6.264	3782491	29434151	1.594	2.365 #
20)	A Methoxychlor	7.498	6.761	343.1E6	1552.1E6	234.354	234.509
21)	B Endrin ke...	7.636	6.997	6493605	60150198	2.030	3.575 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
Data File : PD087951.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 09:03
Operator : AR\AJ
Sample : PEM073
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_D

Client SampleId :

PEM056

Manual Integrations


APPROVED

Reviewed By : Abdul Mirza 03/13/2025

Supervised By : Ankita Jodhani 03/13/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 10:16:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 15:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: Chemtech Consulting Group	SDG No.: Q1526
Project: Weekly Storage Blanks	Instrument ID: ECD_D
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 02/18/2025 02/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
RESC013	RESC058	02/18/2025	09:29	PD087889.D	9.08	3.55
PEM052	PEM067	02/18/2025	09:42	PD087890.D	9.08	3.55
PIBLK486	PIBLK134	02/18/2025	13:20	PD087906.D	9.08	3.55
PEM053	PEM068	02/18/2025	13:34	PD087907.D	9.08	3.55
PIBLK491	PIBLK140	03/12/2025	08:49	PD087950.D	9.08	3.55
PEM056	PEM073	03/12/2025	09:03	PD087951.D	9.08	3.55
PEST-GPC2-BLANK	Q1526-11	03/12/2025	09:16	PD087952.D	0.00	0.00
PEST-GPC2-BLANK-SPIKE	Q1526-12	03/12/2025	10:37	PD087953.D	0.00	0.00
PIBLK492	PIBLK141	03/12/2025	11:46	PD087954.D	9.09	3.56
CCAL01	INDA320	03/12/2025	12:00	PD087955.D	9.08	3.56
CCAL01	INDB320	03/12/2025	12:15	PD087956.D	9.08	3.56

Analytical Sequence

Client: Chemtech Consulting Group

SDG No.: Q1526

Project: Weekly Storage Blanks

Instrument ID: ECD_D

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 02/18/2025 02/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
RESC013	RESC058	02/18/2025	09:29	PD087889.D	8.08	2.88
PEM052	PEM067	02/18/2025	09:42	PD087890.D	8.08	2.88
PIBLK486	PIBLK134	02/18/2025	13:20	PD087906.D	8.08	2.88
PEM053	PEM068	02/18/2025	13:34	PD087907.D	8.08	2.88
PIBLK491	PIBLK140	03/12/2025	08:49	PD087950.D	8.08	2.88
PEM056	PEM073	03/12/2025	09:03	PD087951.D	8.08	2.88
PEST-GPC2-BLANK	Q1526-11	03/12/2025	09:16	PD087952.D	0.00	0.00
PEST-GPC2-BLANK-SPIKE	Q1526-12	03/12/2025	10:37	PD087953.D	0.00	0.00
PIBLK492	PIBLK141	03/12/2025	11:46	PD087954.D	8.08	2.88
CCAL01	INDA320	03/12/2025	12:00	PD087955.D	8.08	2.88
CCAL01	INDB320	03/12/2025	12:15	PD087956.D	8.08	2.88

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PEST-GPC2-BLANK-SPIKE

Contract: CHEM02

Lab Code: CHEM Case No.: Q1526

SAS No.: Q1526 SDG NO.: Q1526

Lab Sample ID: Q1526-12

Date(s) Analyzed: 03/12/2025 03/12/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	7.03	6.98	7.08	35.0	16
	2	6.19	6.14	6.24	40.0	
Aldrin	1	5.28	5.23	5.33	16.0	21.4
	2	4.37	4.32	4.42	20.0	
gamma-BHC (Lindane)	1	4.34	4.29	4.39	16.0	21.6
	2	3.73	3.68	3.78	20.0	
Heptachlor	1	4.94	4.89	4.99	16.0	15.5
	2	4.09	4.04	4.14	19.0	
Dieldrin	1	6.36	6.31	6.41	34.0	13.3
	2	5.52	5.47	5.57	38.0	
Endrin	1	6.58	6.53	6.63	34.0	11.2
	2	5.80	5.75	5.85	37.0	



QC SAMPLE DATA

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	03/12/25
Project:	Weekly Storage Blanks	Date Received:	03/12/25
Client Sample ID:	PIBLK-PD087950.D	SDG No.:	Q1526
Lab Sample ID:	I.BLK-PD087950.D	Matrix:	WATER
Analytical Method:	SFAM_PEST	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
PH :			
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD087950.D	1		03/12/25	PD031225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.0055	U	0.0055	0.050	ug/L
319-85-7	beta-BHC	0.0054	U	0.0054	0.050	ug/L
319-86-8	delta-BHC	0.0099	U	0.0099	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.0059	U	0.0059	0.050	ug/L
76-44-8	Heptachlor	0.0053	U	0.0053	0.050	ug/L
309-00-2	Aldrin	0.0054	U	0.0054	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0053	U	0.0053	0.050	ug/L
959-98-8	Endosulfan I	0.0058	U	0.0058	0.050	ug/L
60-57-1	Dieldrin	0.0056	U	0.0056	0.10	ug/L
72-55-9	4,4-DDE	0.0061	U	0.0061	0.10	ug/L
72-20-8	Endrin	0.0066	U	0.0066	0.10	ug/L
33213-65-9	Endosulfan II	0.0069	U	0.0069	0.10	ug/L
72-54-8	4,4-DDD	0.0072	U	0.0072	0.10	ug/L
1031-07-8	Endosulfan Sulfate	0.0060	U	0.0060	0.10	ug/L
50-29-3	4,4-DDT	0.0057	U	0.0057	0.10	ug/L
72-43-5	Methoxychlor	0.0069	U	0.0069	0.50	ug/L
53494-70-5	Endrin ketone	0.0061	U	0.0061	0.10	ug/L
7421-93-4	Endrin Aldehyde	0.0064	U	0.0064	0.10	ug/L
5103-71-9	cis-Chlordane	0.0059	U	0.0059	0.050	ug/L
5103-74-2	trans-Chlordane	0.0054	U	0.0054	0.050	ug/L
8001-35-2	Toxaphene	0.23	U	0.23	5.00	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0.22		30 - 150	108%	SPK: 20
2051-24-3	Decachlorobiphenyl	0.33		30 - 150	81%	SPK: 40

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	03/12/25
Project:	Weekly Storage Blanks	Date Received:	03/12/25
Client Sample ID:	PIBLK-PD087950.D	SDG No.:	Q1526
Lab Sample ID:	I.BLK-PD087950.D	Matrix:	WATER
Analytical Method:	SFAM_PEST	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
PH :			
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD087950.D	1		03/12/25	PD031225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
Data File : PD087950.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 08:49
Operator : AR\AJ
Sample : PIBLK140
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PIBLK491

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 10:03:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 15:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.553	2.883	45248874	352.0E6	21.685	26.170
27) SA Decachlor...	9.082	8.080	107.2E6	598.1E6	32.547	37.828

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
Data File : PD087950.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 08:49
Operator : AR\AJ
Sample : PIBLK140
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PIBLK491


Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 10:03:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 15:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_

Signal: PD087950.D\IECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 3.553 min
Delta R.T.: 0.000 min
Response: 45248874
Conc: 21.68 ng/ml

Instrument :
ECD_D
ClientSampleId :
PIBLK491

Time

Response_

Signal: PD087950.D\IECD2B.ch




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: -0.001 min
Response: 352016652
Conc: 26.17 ng/ml

Time

Response_

Signal: PD087950.D\IECD1A.ch




#27 Decachlorobiphenyl

R.T.: 9.082 min
Delta R.T.: 0.003 min
Response: 107240691
Conc: 32.55 ng/ml

Time

Response_

Signal: PD087950.D\IECD2B.ch



#27 Decachlorobiphenyl

R.T.: 8.080 min
Delta R.T.: 0.000 min
Response: 598052416
Conc: 37.83 ng/ml

Time

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	03/12/25
Project:	Weekly Storage Blanks	Date Received:	03/12/25
Client Sample ID:	PIBLK-PD087954.D	SDG No.:	Q1526
Lab Sample ID:	I.BLK-PD087954.D	Matrix:	WATER
Analytical Method:	SFAM_PEST	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
PH :			
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD087954.D	1		03/12/25	PD031225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.0055	U	0.0055	0.050	ug/L
319-85-7	beta-BHC	0.0054	U	0.0054	0.050	ug/L
319-86-8	delta-BHC	0.0099	U	0.0099	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.0059	U	0.0059	0.050	ug/L
76-44-8	Heptachlor	0.0053	U	0.0053	0.050	ug/L
309-00-2	Aldrin	0.0054	U	0.0054	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0053	U	0.0053	0.050	ug/L
959-98-8	Endosulfan I	0.0058	U	0.0058	0.050	ug/L
60-57-1	Dieldrin	0.0056	U	0.0056	0.10	ug/L
72-55-9	4,4-DDE	0.0061	U	0.0061	0.10	ug/L
72-20-8	Endrin	0.0066	U	0.0066	0.10	ug/L
33213-65-9	Endosulfan II	0.0069	U	0.0069	0.10	ug/L
72-54-8	4,4-DDD	0.0072	U	0.0072	0.10	ug/L
1031-07-8	Endosulfan Sulfate	0.0060	U	0.0060	0.10	ug/L
50-29-3	4,4-DDT	0.0057	U	0.0057	0.10	ug/L
72-43-5	Methoxychlor	0.0069	U	0.0069	0.50	ug/L
53494-70-5	Endrin ketone	0.0061	U	0.0061	0.10	ug/L
7421-93-4	Endrin Aldehyde	0.0064	U	0.0064	0.10	ug/L
5103-71-9	cis-Chlordane	0.0059	U	0.0059	0.050	ug/L
5103-74-2	trans-Chlordane	0.0054	U	0.0054	0.050	ug/L
8001-35-2	Toxaphene	0.23	U	0.23	5.00	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0.21		30 - 150	106%	SPK: 20
2051-24-3	Decachlorobiphenyl	0.32		30 - 150	81%	SPK: 40

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	03/12/25
Project:	Weekly Storage Blanks	Date Received:	03/12/25
Client Sample ID:	PIBLK-PD087954.D	SDG No.:	Q1526
Lab Sample ID:	I.BLK-PD087954.D	Matrix:	WATER
Analytical Method:	SFAM_PEST	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	Pesticide-TCL
GPC Factor :	1.0	Injection Volume :	
PH :			
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD087954.D	1		03/12/25	PD031225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
Data File : PD087954.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 11:46
Operator : AR\AJ
Sample : PIBLK141
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PIBLK492

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 10:49:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 15:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.561	2.883	44406954	342.6E6	21.281	25.471
27) SA Decachlor...	9.092	8.083	107.0E6	588.8E6	32.471	37.243

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031225\
Data File : PD087954.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 11:46
Operator : AR\AJ
Sample : PIBLK141
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PIBLK492


Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 10:49:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 15:08:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_

Signal: PD087954.D\IECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 3.561 min
Delta R.T.: 0.008 min
Response: 44406954
Conc: 21.28 ng/ml

Instrument :
ECD_D
ClientSampleId :
PIBLK492

Time

Response_

Signal: PD087954.D\IECD2B.ch




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: -0.001 min
Response: 342604086
Conc: 25.47 ng/ml

Time

Response_

Signal: PD087954.D\IECD1A.ch




#27 Decachlorobiphenyl

R.T.: 9.092 min
Delta R.T.: 0.014 min
Response: 106990290
Conc: 32.47 ng/ml

Time

Response_

Signal: PD087954.D\IECD2B.ch



#27 Decachlorobiphenyl

R.T.: 8.083 min
Delta R.T.: 0.003 min
Response: 588810459
Conc: 37.24 ng/ml

Time

Manual Integration Report

Sequence:

PD021825

Instrument

ECD_d

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM067	PD087890.D	alpha-BHC	Abdul	2/26/2025 12:00:39 PM	Ankita	2/26/2025 12:33:03	Peak Integrated by Software
TOXAPH101	PD087891.D	Toxaphene-1 #2	Abdul	2/26/2025 12:00:42 PM	Ankita	2/26/2025 12:33:04	Peak Integrated by Software
PEM068	PD087907.D	alpha-BHC	Abdul	2/26/2025 12:00:56 PM	Ankita	2/26/2025 12:33:06	Peak Integrated by Software

Manual Integration Report

Sequence:	PD031225	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM073	PD087951.D	4,4"-DDE #2	Abdul	3/13/2025 11:54:09 AM	Ankita	3/13/2025 11:55:11	Peak Integrated by Software
INDB320	PD087956.D	beta-BHC #2	Abdul	3/13/2025 11:54:02 AM	Ankita	3/13/2025 11:55:13	Peak Integrated by Software
INDB320	PD087956.D	delta-BHC	Abdul	3/13/2025 11:54:02 AM	Ankita	3/13/2025 11:55:13	Peak Integrated by Software

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QCBatch ID # PD021825

Review By	Abdul	Review On	2/26/2025 12:01:13 PM
Supervise By	Ankita	Supervise On	2/26/2025 12:33:10 PM
SubDirectory	PD021825	HP Acquire Method	HP Processing Method pd021825clp
STD. NAME	STD REF.#		
Tune/Reschk	PP23732,PP23975		
Initial Calibration Stds	PP23731,PP23730,PP23729,PP23728,PP23727,PP23718,PP23724,PP23717,PP23723,PP23716,PP23722,PP23715,PP23721,PP23714,PP23720		
CCC	PP23716,PP23722,PP23729		
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD087888.D	18 Feb 2025 09:15	AR\AJ	Ok
2	RESC058	PD087889.D	18 Feb 2025 09:29	AR\AJ	Ok
3	PEM067	PD087890.D	18 Feb 2025 09:42	AR\AJ	Ok,M
4	TOXAPH101	PD087891.D	18 Feb 2025 09:56	AR\AJ	Ok,M
5	TOXAPH201	PD087892.D	18 Feb 2025 10:10	AR\AJ	Ok
6	TOXAPH301	PD087893.D	18 Feb 2025 10:23	AR\AJ	Ok
7	TOXAPH401	PD087894.D	18 Feb 2025 10:37	AR\AJ	Ok
8	TOXAPH501	PD087895.D	18 Feb 2025 10:51	AR\AJ	Ok
9	INDA101	PD087896.D	18 Feb 2025 11:04	AR\AJ	Ok
10	INDB101	PD087897.D	18 Feb 2025 11:18	AR\AJ	Ok
11	INDA201	PD087898.D	18 Feb 2025 11:31	AR\AJ	Ok
12	INDB201	PD087899.D	18 Feb 2025 11:45	AR\AJ	Ok
13	INDA301	PD087900.D	18 Feb 2025 11:59	AR\AJ	Ok
14	INDB301	PD087901.D	18 Feb 2025 12:12	AR\AJ	Ok
15	INDA401	PD087902.D	18 Feb 2025 12:26	AR\AJ	Ok
16	INDB401	PD087903.D	18 Feb 2025 12:39	AR\AJ	Ok
17	INDA501	PD087904.D	18 Feb 2025 12:53	AR\AJ	Ok
18	INDB501	PD087905.D	18 Feb 2025 13:07	AR\AJ	Ok
19	PIBLK134	PD087906.D	18 Feb 2025 13:20	AR\AJ	Ok
20	PEM068	PD087907.D	18 Feb 2025 13:34	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QCBatch ID # PD031225

Review By	Abdul	Review On	3/13/2025 11:54:25 AM
Supervise By	Ankita	Supervise On	3/13/2025 11:55:18 AM
SubDirectory	PD031225	HP Acquire Method	HP Processing Method pd021825clp
STD. NAME	STD REF.#		
Tune/Reschk	PP23732,PP23975		
Initial Calibration Stds	PP23731,PP23730,PP23729,PP23728,PP23727,PP23718,PP23724,PP23717,PP23723,PP23716,PP23722,PP23715,PP23721,PP23714,PP23720		
CCC	PP23716,PP23722,PP23729		
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD087949.D	12 Mar 2025 08:36	AR\AJ	Ok
2	PIBLK140	PD087950.D	12 Mar 2025 08:49	AR\AJ	Ok
3	PEM073	PD087951.D	12 Mar 2025 09:03	AR\AJ	Ok,M
4	Q1526-11	PD087952.D	12 Mar 2025 09:16	AR\AJ	Ok
5	Q1526-12	PD087953.D	12 Mar 2025 10:37	AR\AJ	Ok
6	PIBLK141	PD087954.D	12 Mar 2025 11:46	AR\AJ	Ok
7	INDA320	PD087955.D	12 Mar 2025 12:00	AR\AJ	Ok
8	INDB320	PD087956.D	12 Mar 2025 12:15	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD021825

Review By	Abdul	Review On	2/26/2025 12:01:13 PM
Supervise By	Ankita	Supervise On	2/26/2025 12:33:10 PM
SubDirectory	PD021825	HP Acquire Method	HP Processing Method pd021825clp

STD. NAME	STD REF.#
Tune/Reschk	PP23732,PP23975
Initial Calibration Stds	PP23731,PP23730,PP23729,PP23728,PP23727,PP23718,PP23724,PP23717,PP23723,PP23716,PP23722,PP23715,PP23721,PP23714,PP23720
CCC	PP23716,PP23722,PP23729
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD087888.D	18 Feb 2025 09:15		AR\AJ	Ok
2	RESC058	RESC013	PD087889.D	18 Feb 2025 09:29		AR\AJ	Ok
3	PEM067	PEM052	PD087890.D	18 Feb 2025 09:42		AR\AJ	Ok,M
4	TOXAPH101	TOXAPH1075	PD087891.D	18 Feb 2025 09:56		AR\AJ	Ok,M
5	TOXAPH201	TOXAPH2076	PD087892.D	18 Feb 2025 10:10		AR\AJ	Ok
6	TOXAPH301	TOXAPH3077	PD087893.D	18 Feb 2025 10:23		AR\AJ	Ok
7	TOXAPH401	TOXAPH4078	PD087894.D	18 Feb 2025 10:37		AR\AJ	Ok
8	TOXAPH501	TOXAPH5079	PD087895.D	18 Feb 2025 10:51		AR\AJ	Ok
9	INDA101	INDA1080	PD087896.D	18 Feb 2025 11:04		AR\AJ	Ok
10	INDB101	INDB1081	PD087897.D	18 Feb 2025 11:18		AR\AJ	Ok
11	INDA201	INDA2082	PD087898.D	18 Feb 2025 11:31		AR\AJ	Ok
12	INDB201	INDB2083	PD087899.D	18 Feb 2025 11:45		AR\AJ	Ok
13	INDA301	INDA3084	PD087900.D	18 Feb 2025 11:59		AR\AJ	Ok
14	INDB301	INDB3085	PD087901.D	18 Feb 2025 12:12		AR\AJ	Ok
15	INDA401	INDA4086	PD087902.D	18 Feb 2025 12:26		AR\AJ	Ok
16	INDB401	INDB4087	PD087903.D	18 Feb 2025 12:39		AR\AJ	Ok
17	INDA501	INDA5088	PD087904.D	18 Feb 2025 12:53		AR\AJ	Ok
18	INDB501	INDB5089	PD087905.D	18 Feb 2025 13:07		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD021825

Review By	Abdul	Review On	2/26/2025 12:01:13 PM
Supervise By	Ankita	Supervise On	2/26/2025 12:33:10 PM
SubDirectory	PD021825	HP Acquire Method	HP Processing Method pd021825clp

STD. NAME	STD REF.#
Tune/Reschk	PP23732,PP23975
Initial Calibration Stds	PP23731,PP23730,PP23729,PP23728,PP23727,PP23718,PP23724,PP23717,PP23723,PP23716,PP23722,PP23715,PP23721,PP23714,PP23720
CCC	PP23716,PP23722,PP23729
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	PIBLK134	PIBLK486	PD087906.D	18 Feb 2025 13:20		AR\AJ	Ok
20	PEM068	PEM053	PD087907.D	18 Feb 2025 13:34		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD031225

Review By	Abdul	Review On	3/13/2025 11:54:25 AM
Supervise By	Ankita	Supervise On	3/13/2025 11:55:18 AM
SubDirectory	PD031225	HP Acquire Method	HP Processing Method pd021825clp

STD. NAME	STD REF.#
Tune/Reschk	PP23732,PP23975
Initial Calibration Stds	PP23731,PP23730,PP23729,PP23728,PP23727,PP23718,PP23724,PP23717,PP23723,PP23716,PP23722,PP23715,PP23721,PP23714,PP23720
CCC	PP23716,PP23722,PP23729
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD087949.D	12 Mar 2025 08:36		AR\AJ	Ok
2	PIBLK140	PIBLK491	PD087950.D	12 Mar 2025 08:49		AR\AJ	Ok
3	PEM073	PEM056	PD087951.D	12 Mar 2025 09:03		AR\AJ	Ok,M
4	Q1526-11	PEST-GPC2-BLANK	PD087952.D	12 Mar 2025 09:16		AR\AJ	Ok
5	Q1526-12	PEST-GPC2-BLANK-S	PD087953.D	12 Mar 2025 10:37		AR\AJ	Ok
6	PIBLK141	PIBLK492	PD087954.D	12 Mar 2025 11:46		AR\AJ	Ok
7	INDA320	INDA3234	PD087955.D	12 Mar 2025 12:00		AR\AJ	Ok
8	INDB320	INDB3235	PD087956.D	12 Mar 2025 12:15		AR\AJ	Ok,M

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 3/10/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 03/07/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 03/08/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB134943

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1523-02	WC-AI-01-C	1	1.15	8.56	9.71	7.88	78.6	
Q1523-05	WC-AI-02-C	2	1.15	8.73	9.88	7.96	78.0	
Q1524-01	72-11930-343-COMP	3	1.15	8.69	9.84	8.84	88.5	
Q1524-02	VNJ-241	4	1.16	8.62	9.78	8.82	88.9	
Q1526-05	SVOC-GPC-BLANK	5	1.00	1.00	2.00	2.00	100.0	
Q1526-06	PEST-GPC-BLANK	6	1.00	1.00	2.00	2.00	100.0	
Q1526-07	PEST-GPC-BLANK-SPIKE	7	1.00	1.00	2.00	2.00	100.0	
Q1526-08	PCB-GPC-BLANK	8	1.00	1.00	2.00	2.00	100.0	
Q1526-09	PCB-GPC-BLANK-SPIKE	9	1.00	1.00	2.00	2.00	100.0	
Q1526-10	SVOC-GPC2-BLANK	10	1.00	1.00	2.00	2.00	100.0	
Q1526-11	PEST-GPC2-BLANK	11	1.00	1.00	2.00	2.00	100.0	
Q1526-12	PEST-GPC2-BLANK-SPIKE	12	1.00	1.00	2.00	2.00	100.0	
Q1526-13	PCB-GPC2-BLANK	13	1.00	1.00	2.00	2.00	100.0	
Q1526-14	PCB-GPC2-BLANK-SPIKE	14	1.00	1.00	2.00	2.00	100.0	
Q1527-01	022825-SEALPOT	15	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1527-02	030425-A	16	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1527-03	030425-B	17	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1529-01	60300	18	1.15	8.43	9.58	9.48	98.8	
Q1529-02	60301	19	1.15	8.83	9.98	6.71	63.0	
Q1529-03	60303	20	1.12	8.71	9.83	7.43	72.4	
Q1529-04	60289	21	1.00	1.00	2.00	2.00	100.0	debris
Q1530-01	1-WIPE	22	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1530-02	2-WIPE	23	1.00	1.00	2.00	2.00	100.0	wipe sample

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

134443

WorkList Name : %1-030725 WorkList ID : 188107 Department : Wet-Chemistry Date : 03-07-2025 09:05:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1523-02	WC-AI-01-C	Solid	Percent Solids	Cool 4 deg C	ENTA05	I31	03/06/2025	Chemtech -SO
Q1523-05	WC-AI-02-C	Solid	Percent Solids	Cool 4 deg C	ENTA05	I31	03/06/2025	Chemtech -SO
Q1524-01	72-11930-343-COMP	Solid	Percent Solids	Cool 4 deg C	PSEG03	K31	03/07/2025	Chemtech -SO
Q1524-02	VNJ-241	Solid	Percent Solids	Cool 4 deg C	PSEG03	K31	03/07/2025	Chemtech -SO
Q1526-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	03/07/2025	Chemtech -SO
Q1526-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	03/07/2025	Chemtech -SO
Q1526-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	03/07/2025	Chemtech -SO
Q1526-08	PCB-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	03/07/2025	Chemtech -SO
Q1526-09	PCB-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	03/07/2025	Chemtech -SO
Q1526-10	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	03/07/2025	Chemtech -SO
Q1526-11	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	03/07/2025	Chemtech -SO
Q1526-12	PEST-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	03/07/2025	Chemtech -SO
Q1526-13	PCB-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	03/07/2025	Chemtech -SO
Q1526-14	PCB-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	F11	03/07/2025	Chemtech -SO
Q1527-01	022825-SEALPOT	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	03/07/2025	Chemtech -SO
Q1527-02	030425-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	03/07/2025	Chemtech -SO
Q1527-03	030425-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	03/07/2025	Chemtech -SO
Q1529-01	60300	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	03/07/2025	Chemtech -SO
Q1529-02	60301	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	03/07/2025	Chemtech -SO
Q1529-03	60303	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	03/07/2025	Chemtech -SO
Q1529-04	60289	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	03/07/2025	Chemtech -SO

Date/Time 03/07/25 15:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]
 Date/Time 03/07/25 17:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

✓ 17134943

WorkList Name : %1-030725 WorkList ID : 188107 Department : Wet-Chemistry Date : 03-07-2025 09:05:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1530-01	1-WIPE	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	03/07/2025	Chemtech -SO
Q1530-02	2-WIPE	Solid	Percent Solids	Cool 4 deg C	PSEG03	I31	03/07/2025	Chemtech -SO

Date/Time 03/07/25 15:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 03/07/25 17:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

SOP ID: M3620C-Florisol cleanup-6, M3640A-GPC cleanup-6

Clean Up SOP #: GPC Cleanup, Florisol

Matrix: Solid

Weigh By: N/A

Balance check: N/A

Balance ID: N/A

pH Strip Lot#: N/A

Extraction By: N/A

Filter By: N/A

pH Meter ID: N/A

Hood ID: 6,7

Extraction Start Date: 03/10/2025

Extraction Start Time: 10:34

Extraction End Date: 03/11/2025

Extraction End Time: 10:00

Concentration By: RS

Supervisor By: RUPESH

Extraction Method: ☐ Separatory Funnel

☐ Continuous Liquid/Liquid

☐ Sonication

☐ Waste Dilution

☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
GPC Pest Spike	10.0ML	20/40 PPB	PP24060
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
MeCl2/Acetone/1:1	N/A	EP2591
Baked Na2SO4	N/A	EP2590
Methylene Chloride	N/A	E3878
Hexane	N/A	E3877
Florisol	N/A	E3806
9:1 Hexane:Acetone Mixture	N/A	EP2545
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

No Extraction only GPC Check prepared.

KD Bath ID: WATER BATH-1

KD Bath Temperature: 60 °C

Envap ID: NEVAP-02


Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
3/11/25	RS (Ext lab)	7-P-pest 1PUB
10:05	Preparation Group	Analysis Group

Analytical Method: M3620C-Florisil cleanup-6, M3640A-GPC cleanup-6

Concentration Date: 03/10/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
Q1526-11	PEST-GPC2-BLANK	Pesticide-TCL	10.0	N/A	RUPESH	ritesh	5			
Q1526-12	PEST-GPC2-BLANK-SPIKE	Pesticide-TCL	10.0	N/A	RUPESH	ritesh	5			



Rs
3/11

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1526 WorkList ID : 188157 Department : Extraction Date : 03-10-2025 10:28:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1526-10	SVOC-GPC2-BLANK	Solid	SVOC-Chemtech Full -25	Cool 4 deg C	CHEM02	F11	03/07/2025	SFAM_SVOC
Q1526-11	PEST-GPC2-BLANK	Solid	Pesticide-TCL	Cool 4 deg C	CHEM02	F11	03/07/2025	SFAM_PEST
Q1526-12	PEST-GPC2-BLANK-SPIKE	Solid	Pesticide-TCL	Cool 4 deg C	CHEM02	F11	03/07/2025	SFAM_PEST
Q1526-13	PCB-GPC2-BLANK	Solid	PCB	Cool 4 deg C	CHEM02	F11	03/07/2025	SFAM_PCB
Q1526-14	PCB-GPC2-BLANK-SPIKE	Solid	PCB	Cool 4 deg C	CHEM02	F11	03/07/2025	SFAM_PCB

Date/Time 3/10/25 10:29
Raw Sample Received by: RS (Ext-Lab)
Raw Sample Relinquished by: *CRS*

Date/Time 3/10/25 10:50
Raw Sample Received by: *CRS*
Raw Sample Relinquished by: RS (Ext-Lab)

Prep Standard - Chemical Standard Summary

Order ID : Q1526

Test : Pesticide-TCL

Prepbatch ID : PB167055,

Sequence ID/Qc Batch ID: PD031225,

Standard ID :

EP2590,EP2591,PP23713,PP23714,PP23715,PP23716,PP23717,PP23718,PP23719,PP23720,PP23721,PP23722,PP23723,PP23724,PP23725,PP23726,PP23727,PP23728,PP23729,PP23730,PP23731,PP23732,PP23975,PP24060,

Chemical ID :

E3551,E3788,E3805,E3806,E3826,E3829,E3876,E3877,E3878,P11763,P11797,P11945,P13191,P13360,P13383,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2590	02/26/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 02/26/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2017	1:1 ACETONE/METHYLENE CHLORIDE	EP2591	02/26/2025	08/14/2025	RUPESHKUMAR SHAH	None	None	Riteshkumar Patel 02/26/2025

FROM 8000.00000ml of E3876 + 8000.00000ml of E3878 = Final Quantity: 16000.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3921	CS6 IND STD MIX A	PP23713	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 1.00000ml of P11945 + 49.00000ml of E3805 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
679	CS5 IND STD MIX A	PP23714	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23713 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
680	CS4 IND STD MIX A	PP23715	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23714 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
681	CS3 IND STD MIX A	PP23716	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23715 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
682	CS2 IND STD MIX A	PP23717	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23716 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
683	CS1 IND STD MIX A	PP23718	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23717 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3920	CS6 IND STD MIX B	PP23719	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 1.00000ml of P11763 + 49.00000ml of E3805 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
684	CS5 IND STD MIX B	PP23720	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23719 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
685	CS4 IND STD MIX B	PP23721	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23720 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
686	CS3 IND STD MIX B	PP23722	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23721 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
687	CS2 IND STD MIX B	PP23723	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23722 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
688	CS1 IND STD MIX B	PP23724	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23723 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
70	10/20 PPM Pest/PCB SOM01.2 Surg Stock	PP23725	10/02/2024	04/01/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 1.00000ml of P13383 + 9.00000ml of E3788 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3922	Toxaphene CS6	PP23726	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.80000ml of P13360 + 48.40000ml of E3805 + 0.80000ml of PP23725 = Final Quantity: 50.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
674	Toxaphene CS5	PP23727	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23726 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
675	Toxaphene CS4	PP23728	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23727 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
676	Toxaphene CS3	PP23729	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23727 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
677	Toxaphene CS2	PP23730	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23729 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
678	Toxaphene CS1	PP23731	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of E3805 + 0.50000ml of PP23730 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
776	EPA S0M01.2 RESCHK	PP23732	10/02/2024	03/30/2025	Abdul Mirza	None	None	Yogesh Patel
								10/03/2024

FROM 0.50000ml of PP23716 + 0.50000ml of PP23722 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
758	PEM Mix w/Surr	PP23975	11/14/2024	05/09/2025	Ankita Jodhani	None	None	Yogesh Patel
								11/18/2024

FROM 1.00000ml of P11797 + 99.00000ml of E3826 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3793	20/40 PPB PEST GPC spike solution	PP24060	11/25/2024	05/18/2025	Ankita Jodhani	None	None	Yogesh Patel
								11/27/2024

FROM 0.08000ml of P13191 + 99.92000ml of E3829 = Final Quantity: 100.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	23H1462005	04/23/2025	08/13/2024 / Rajesh	08/13/2024 / Rajesh	E3788

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24C1862008	03/30/2025	09/30/2024 / Rajesh	09/25/2024 / Rajesh	E3805

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agela Technologies Inc.	FS0006 / Cleanert Florisil cartridge	M06518	03/25/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3806

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	05/09/2025	11/09/2024 / Rajesh	11/07/2024 / Rajesh	E3826

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24J0862003	05/18/2025	11/18/2024 / Rajesh	11/04/2024 / Rajesh	E3829

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	08/25/2025	02/25/2025 /	02/12/2025 / Rajesh	E3876

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24K1762005	08/14/2025	02/14/2025 / Rajesh	12/27/2024 / Rajesh	E3878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32004 / Pesticide Mix, CLP method, Standard Mix B 3/90 SOW, Hexane, 1mL/ampul	A0176477	04/02/2025	10/02/2024 / Abdul	05/27/2022 / Sohil	P11763

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32074 / Pesticide Performance Evaluation Mix w/Surrogate	A0183168	05/14/2025	11/14/2024 / Ankita	05/27/2022 / Sohil	P11797

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32003 / Pesticide Mix, CLP method, Standard Mix A 3/90 SOW, Hexane, 1mL/ampul	A0176384	04/02/2025	10/02/2024 / Abdul	06/22/2022 / Abdul	P11945

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32018 / Pesticide Matrix Spike Mix	A0203053	05/25/2025	11/25/2024 / Ankita	01/15/2024 / Abdul	P13191

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0203830	04/02/2025	10/02/2024 / Abdul	05/03/2024 / Abdul	P13360

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32453 / SOM01.1 Pesticide Surrogate Standard	A0194530	04/02/2025	10/02/2024 / Abdul	12/06/2023 / Abdul	P13383



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03
Batch No.: 23H1462005
Manufactured Date: 2023-07-26
Expiration Date: 2026-07-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.4 %	99.7 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.6	< 0.1
Water (H ₂ O)	≤ 0.5 %	0.3 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 8/13/24

E 3788

Ken Koehnlein
Sr. Manager, Quality Assurance

Hexanes (95% n-hexane)
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9262-03
Batch No.: 24C1862008
Manufactured Date: 2024-01-30
Expiration Date: 2025-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.4 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 9/25/24

E 3805

Jamie Croak
Director Quality Operations, Bioscience Production

Cleanert Florisil

1g/6ml 30/pkg

固相萃取产品

LOT#: M06518

MFG#: F04074



Made in China

CAT# FS0006

 Agela Technologies

E 3806



n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3826

Rec'd by RP on 11/7/24

Jamie Croak
Director Quality Operations, Bioscience Production

11/6/24

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

 **avantors[™]**

Material No.: 9266-A4

Batch No.: 24J0862003

Manufactured Date: 2024-09-12

Expiration Date: 2025-12-12

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	2
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3829



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

 **avantors**™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 2/12/25

E 3876



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

 E3877

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24K1762005
Manufactured Date: 2024-10-08
Expiration Date: 2026-01-07
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.5 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	0.0
Chloride (Cl)	$\leq 10 \text{ ppm}$	$< 5 \text{ ppm}$
Water (by KF, coulometric)	$\leq 0.02 \%$	0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3878

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087 U.S.A. Phone 610.386.1700



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis

P11759 to P11763

Received by : SJ 5/27/2022



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32004 Lot No.: A0176477

Description : Pesticide Standard Mix B (3/90)

Pesticide Standard Mix B (3/90) 8-16 µg/mL, Hexane/Toluene (90:10), 1mL/ampul

Container Size : 2 mL Pkg Amt: > 1 mL

Expiration Date : September 30, 2025 Storage: 10°C or colder

Handling: Contains PCBs - sonicate prior to use. Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	2,4,5,6-Tetrachloro-m-xylene CAS # 877-09-8 (Lot 0052481) Purity 98%	8.0 µg/mL	+/- 0.1453 +/- 0.3910 +/- 0.5461	µg/mL Gravimetric Unstressed Stressed
2	beta-BHC CAS # 319-85-7 (Lot SL210106) Purity 99%	8.0 µg/mL	+/- 0.1447 +/- 0.3892 +/- 0.5436	µg/mL Gravimetric Unstressed Stressed
3	delta-BHC CAS # 319-86-8 (Lot 8-TAH-175-1) Purity 98%	8.0 µg/mL	+/- 0.1453 +/- 0.3910 +/- 0.5461	µg/mL Gravimetric Unstressed Stressed
4	Aldrin CAS # 309-00-2 (Lot 12044700) Purity 99%	8.0 µg/mL	+/- 0.1447 +/- 0.3892 +/- 0.5436	µg/mL Gravimetric Unstressed Stressed
5	Heptachlor epoxide (isomer B) CAS # 1024-57-3 (Lot 11129300) Purity 99%	8.0 µg/mL	+/- 0.1447 +/- 0.3892 +/- 0.5436	µg/mL Gravimetric Unstressed Stressed
6	trans-Chlordane CAS # 5103-74-2 (Lot 32095) Purity 99%	8.0 µg/mL	+/- 0.1447 +/- 0.3892 +/- 0.5436	µg/mL Gravimetric Unstressed Stressed
7	cis-Chlordane CAS # 5103-71-9 (Lot 31707) Purity 99%	8.0 µg/mL	+/- 0.1447 +/- 0.3892 +/- 0.5436	µg/mL Gravimetric Unstressed Stressed

8	4,4'-DDE			16.1	µg/mL	+/-	0.2906	µg/mL	Gravimetric
	CAS #	72-55-9	(Lot GHYQG)			+/-	0.7817	µg/mL	Unstressed
	Purity	99%				+/-	1.0918	µg/mL	Stressed
9	Endosulfan II			16.0	µg/mL	+/-	0.2894	µg/mL	Gravimetric
	CAS #	33213-65-9	(Lot 10861900)			+/-	0.7785	µg/mL	Unstressed
	Purity	99%				+/-	1.0873	µg/mL	Stressed
10	Endrin aldehyde			16.0	µg/mL	+/-	0.2894	µg/mL	Gravimetric
	CAS #	7421-93-4	(Lot 30606)			+/-	0.7785	µg/mL	Unstressed
	Purity	99%				+/-	1.0873	µg/mL	Stressed
11	Endosulfan sulfate			16.0	µg/mL	+/-	0.2894	µg/mL	Gravimetric
	CAS #	1031-07-8	(Lot BCCB0424)			+/-	0.7785	µg/mL	Unstressed
	Purity	99%				+/-	1.0873	µg/mL	Stressed
12	Endrin ketone			16.0	µg/mL	+/-	0.2895	µg/mL	Gravimetric
	CAS #	53494-70-5	(Lot 11058900)			+/-	0.7788	µg/mL	Unstressed
	Purity	98%				+/-	1.0877	µg/mL	Stressed
13	Decachlorobiphenyl (BZ# 209)			16.1	µg/mL	+/-	0.2906	µg/mL	Gravimetric
	CAS #	2051-24-3	(Lot 30679)			+/-	0.7817	µg/mL	Unstressed
	Purity	99%				+/-	1.0918	µg/mL	Stressed
Solvent: Hexane/Toluene (90:10)									
CAS # 110-54-3/108-88-3									
Purity 99%									

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)


Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Aurelia B. Confer

Aurelia Confer - Operations Tech I

Date Mixed: 16-Sep-2021

Balance: B707717271

Amanda Miller

Amanda Miller - Operations Tech-ARM QC

Date Passed: 20-Sep-2021

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis

P11789 to P11793

Received by SJ 5/27/2022



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32074 **Lot No.:** A0183168

Description : Pesticide Performance Eval Mix w/Surrogate

Performance Evaluation Std. 3/90 SOW w/surrogates 1-25µg/mL,
Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2026 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	2,4,5,6-Tetrachloro-m-xylene CAS # 877-09-8 (Lot 0052481) Purity 98%	2.0 µg/mL	+/- 0.1220 +/- 0.1523 +/- 0.1799	µg/mL Gravimetric Unstressed Stressed
2	alpha-BHC CAS # 319-84-6 (Lot 12469000) Purity 99%	1.0 µg/mL	+/- 0.0610 +/- 0.0762 +/- 0.0900	µg/mL Gravimetric Unstressed Stressed
3	gamma-BHC (Lindane) CAS # 58-89-9 (Lot 12642100) Purity 99%	1.0 µg/mL	+/- 0.0610 +/- 0.0762 +/- 0.0900	µg/mL Gravimetric Unstressed Stressed
4	beta-BHC CAS # 319-85-7 (Lot BCCC6425) Purity 99%	1.0 µg/mL	+/- 0.0610 +/- 0.0762 +/- 0.0900	µg/mL Gravimetric Unstressed Stressed
5	Endrin CAS # 72-20-8 (Lot 13000500) Purity 99%	5.1 µg/mL	+/- 0.3045 +/- 0.3805 +/- 0.4496	µg/mL Gravimetric Unstressed Stressed
6	4,4'-DDT CAS # 50-29-3 (Lot 210916JLM) Purity 99%	10.1 µg/mL	+/- 0.6090 +/- 0.7609 +/- 0.8992	µg/mL Gravimetric Unstressed Stressed
7	Methoxychlor CAS # 72-43-5 (Lot 12555700) Purity 98%	25.2 µg/mL	+/- 1.5221 +/- 1.9018 +/- 2.2475	µg/mL Gravimetric Unstressed Stressed

8	Decachlorobiphenyl (BZ# 209)	2.0 µg/mL	+/- 0.1221	µg/mL	Gravimetric
	CAS # 2051-24-3 (Lot 30679)		+/- 0.1524	µg/mL	Unstressed
	Purity 99%		+/- 0.1800	µg/mL	Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)


Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)

Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Brittany Federinko - Operations Tech I

Date Mixed: 22-Mar-2022

Balance: 1128360905


John Lidgett - AD Chemist

Date Passed: 24-Mar-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

P11794 to P11798

Received by SJ 5/27/2022



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32074 **Lot No.:** A0183168

Description : Pesticide Performance Eval Mix w/Surrogate

Performance Evaluation Std. 3/90 SOW w/surrogates 1-25µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2026 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene CAS # 877-09-8 (Lot 0052481) Purity 98%	2.0 µg/mL	+/- 0.1220 µg/mL Gravimetric +/- 0.1523 µg/mL Unstressed +/- 0.1799 µg/mL Stressed
2	alpha-BHC CAS # 319-84-6 (Lot 12469000) Purity 99%	1.0 µg/mL	+/- 0.0610 µg/mL Gravimetric +/- 0.0762 µg/mL Unstressed +/- 0.0900 µg/mL Stressed
3	gamma-BHC (Lindane) CAS # 58-89-9 (Lot 12642100) Purity 99%	1.0 µg/mL	+/- 0.0610 µg/mL Gravimetric +/- 0.0762 µg/mL Unstressed +/- 0.0900 µg/mL Stressed
4	beta-BHC CAS # 319-85-7 (Lot BCCC6425) Purity 99%	1.0 µg/mL	+/- 0.0610 µg/mL Gravimetric +/- 0.0762 µg/mL Unstressed +/- 0.0900 µg/mL Stressed
5	Endrin CAS # 72-20-8 (Lot 13000500) Purity 99%	5.1 µg/mL	+/- 0.3045 µg/mL Gravimetric +/- 0.3805 µg/mL Unstressed +/- 0.4496 µg/mL Stressed
6	4,4'-DDT CAS # 50-29-3 (Lot 210916JLM) Purity 99%	10.1 µg/mL	+/- 0.6090 µg/mL Gravimetric +/- 0.7609 µg/mL Unstressed +/- 0.8992 µg/mL Stressed
7	Methoxychlor CAS # 72-43-5 (Lot 12555700) Purity 98%	25.2 µg/mL	+/- 1.5221 µg/mL Gravimetric +/- 1.9018 µg/mL Unstressed +/- 2.2475 µg/mL Stressed

8	Decachlorobiphenyl (BZ# 209)	2.0 µg/mL	+/- 0.1221	µg/mL	Gravimetric
	CAS # 2051-24-3 (Lot 30679)		+/- 0.1524	µg/mL	Unstressed
	Purity 99%		+/- 0.1800	µg/mL	Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Column:
 30m x .25mm x .2um
 Rtx-CLP II (cat.# 11323)


Carrier Gas:
 helium-constant pressure 20 psi.

Temp. Program:
 150°C to 300°C
 @ 4°C/min. (hold 5 min.)

Inj. Temp:
 200°C

Det. Temp:
 300°C

Det. Type:
 ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Brittany Federinko
 Brittany Federinko - Operations Tech I

Date Mixed: 22-Mar-2022 **Balance:** 1128360905

John Lidgett
 John Lidgett - AD Chemist

Date Passed: 24-Mar-2022

Manufactured under Restek's ISO 9001:2015
 Registered Quality System
 Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32018 **Lot No.:** A0203053
Description : Pesticide Matrix Spike Mix
Pesticide Matrix Spike Mix 25-50 µg/mL, Acetone, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : October 31, 2027 **Storage:** 10°C or colder
Ship: Ambient

P13191
↓
P13194
9/15/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	gamma-BHC (Lindane)	58-89-9	14646400	99%	25.0 µg/mL	+/- 1.3149
2	Heptachlor	76-44-8	813251	99%	25.0 µg/mL	+/- 1.3149
3	Aldrin	309-00-2	14389400	98%	25.0 µg/mL	+/- 1.3164
4	Dieldrin	60-57-1	14515000	98%	50.0 µg/mL	+/- 2.6297
5	Endrin	72-20-8	14485300	98%	50.0 µg/mL	+/- 2.6286
6	4,4'-DDT	50-29-3	230410JLMA	98%	50.1 µg/mL	+/- 2.6317

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0203038

Description : Toxaphene Standard

Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.0 µg/mL	+/- 55.9920

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13358
↓
P 13369
(12)

[Signature]
05-06-2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 10-Oct-2023

Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13358
↓
P13369
↓
(12)


RAUF
05-06-2024



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32453 **Lot No.:** A0194530

Description : SOM01.1 Pesticide Surrogate Standard

Pesticide Surrogate Mix 100-200µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2029 **Storage:** 0°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	100.6 µg/mL	+/- 5.5961
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.3 µg/mL	+/- 11.1978

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

P 13382 } (5)
↓
P 13386

[Signature]
05/01/2024

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Jess Hoy - Operations Tech I

Date Mixed: 09-Feb-2023

Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 15-Feb-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13382 } (5)
↓
P13386 }


05/01/2024



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32003 **Lot No.:** A0176384

Description : Pesticide Standard Mix A (3/90)

Standard Mix A 3/90 SOW 8-80µg/mL, Hexane/Toluene (90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P 11945
↓
P 11946
②
UR
06/24/2022

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	2,4,5,6-Tetrachloro-m-xylene CAS # 877-09-8 (Lot 0052481) Purity 98%	8.0 µg/mL	+/- 0.1442 +/- 0.3878 +/- 0.5417	µg/mL Gravimetric µg/mL Unstressed µg/mL Stressed
2	alpha-BHC CAS # 319-84-6 (Lot 3324600) Purity 99%	8.0 µg/mL	+/- 0.1447 +/- 0.3892 +/- 0.5436	µg/mL Gravimetric µg/mL Unstressed µg/mL Stressed
3	gamma-BHC (Lindane) CAS # 58-89-9 (Lot 11837800) Purity 98%	8.0 µg/mL	+/- 0.1442 +/- 0.3878 +/- 0.5417	µg/mL Gravimetric µg/mL Unstressed µg/mL Stressed
4	Heptachlor CAS # 76-44-8 (Lot 0006617486) Purity 99%	8.0 µg/mL	+/- 0.1447 +/- 0.3892 +/- 0.5436	µg/mL Gravimetric µg/mL Unstressed µg/mL Stressed
5	Endosulfan I CAS # 959-98-8 (Lot BCBS8631) Purity 99%	8.0 µg/mL	+/- 0.1447 +/- 0.3892 +/- 0.5436	µg/mL Gravimetric µg/mL Unstressed µg/mL Stressed
6	Dieldrin CAS # 60-57-1 (Lot 12074700) Purity 97%	16.0 µg/mL	+/- 0.2889 +/- 0.7772 +/- 1.0854	µg/mL Gravimetric µg/mL Unstressed µg/mL Stressed
7	Endrin CAS # 72-20-8 (Lot 11773800) Purity 98%	16.0 µg/mL	+/- 0.2895 +/- 0.7788 +/- 1.0877	µg/mL Gravimetric µg/mL Unstressed µg/mL Stressed

8	4,4'-DDD			16.0	µg/mL	+/-	0.2894	µg/mL	Gravimetric
	CAS #	72-54-8	(Lot HAN02)			+/-	0.7785	µg/mL	Unstressed
	Purity	99%				+/-	1.0873	µg/mL	Stressed
9	4,4'-DDT			16.0	µg/mL	+/-	0.2895	µg/mL	Gravimetric
	CAS #	50-29-3	(Lot 210823JLM)			+/-	0.7788	µg/mL	Unstressed
	Purity	98%				+/-	1.0877	µg/mL	Stressed
10	Methoxychlor			80.0	µg/mL	+/-	1.4132	µg/mL	Gravimetric
	CAS #	72-43-5	(Lot 12325400)			+/-	3.8798	µg/mL	Unstressed
	Purity	97%				+/-	5.4271	µg/mL	Stressed
11	Decachlorobiphenyl (BZ# 209)			16.0	µg/mL	+/-	0.2894	µg/mL	Gravimetric
	CAS #	2051-24-3	(Lot 30638)			+/-	0.7785	µg/mL	Unstressed
	Purity	99%				+/-	1.0873	µg/mL	Stressed
Solvent: Hexane/Toluene (90:10)									
	CAS #	110-54-3/108-88-3							
	Purity	99%							

P 11945 } (2)
↓
P 11946
↑
MR
06/24/2022

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)


Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C


Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Erik Strommer - Operations Tech I

Date Mixed: 14-Sep-2021 **Balance:** 1128353505


Amanda Miller - Operations Tech-ARM QC

Date Passed: 17-Sep-2021

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P11945
↓
P11946
②
06/24/2022

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.



SHIPPING DOCUMENTS



284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 78-8922

www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q1526

COC Number:

CLIENT INFORMATION

COMPANY: Alliance Technical Group

ADDRESS: 284 Sheffield Street

CITY Mountainside STATE: NJ ZIP: 07092

ATTENTION: QA Officer

PHONE: 908-789-8900 FAX: 908-788-9222

PROJECT INFORMATION

PROJECT NAME: Weekly Storage Blanks

PROJECT #: LOCATION:

PROJECT MANAGER:

E-MAIL:

PHONE:

FAX:

BILLING INFORMATION

BILL TO: Same

PO#

ADDRESS:

CITY:

STATE: ZIP:

ATTENTION:

PHONE:

DATA TURNAROUND INFORMATION

FAX: 10 DAYS*

HARD COPY: 10 DAYS*

EDD DAYS*

* TO BE APPROVED BY ALLIANCE

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

- ☐ RESULTS ONLY (L1) ☐ US EPA CLP (L4)
☐ RESULTS + QC (L2) ☐ NYS ASP "A" (L2)
☐ NJ REDUCED (L3) ☐ NYS ASP "B" (L4)
☐ NJ FULL (L4) ☐ Other _____
☐ EDD Format _____

ANALYSIS

VOC-Drinking H2O	SVOC-Full 25-8270	PESTICIDE-TCL	PCB						
1	2	3	4	5	6	7	8	9	

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles	A/E									<- Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-ICE F-Other
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	Storage-Blank-SOIL-REF-6	AQ		X			2	X									
2.	Storage-Blank-WATER-REF-5	AQ		X			2	X									
3.	Storage-Blank-WATER-REF-4	AQ		X			2	X									
4.	Storage-Blank-SAMPLE-MANAGEMENT	AQ		X			2	X									
5.	SVOC-GPC-BLANK	SO		X			1		X								
6.	PEST-GPC-BLANK	SO		X			1			X							
7.	PEST-GPC-BLANK-SPIKE	SO		X			1			X							
8.	PCB-GPC-BLANK	SO		X			1				X						
9.	PCB-GPC-BLANK-SPIKE	SO		X			1				X						
10.	SVOC-GPC2-BLANK	SO		X			1		X								

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	Conditions of bottles or coolers at receipt: ☒ Compliant ☐ Non Compliant ☐ Cooler Temp _____ MeOH extraction requires an additional 4oz. Jar for percent solid ☐ Ice in Cooler?: _____
1. <i>Relinquished</i>		1.	Comments:
RELINQUISHED BY	DATE/TIME	RECEIVED BY	
2.		2.	
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	
3.		3.	

Page 1 of 2

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
ALLIANCE: ☐ Picked Up ☐ Overnight

Shipment Complete
☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT

YELLOW - ALLIANCE COPY

PINK - SAMPLER COPY

3/7



284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 78-8922

www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q1526

COC Number:

CLIENT INFORMATION

COMPANY: Alliance Technical Group

ADDRESS: 284 Sheffield Street

CITY Mountainside STATE: NJ ZIP: 07092

ATTENTION: QA Officer

PHONE: 908-789-8900

FAX: 908-788-9222

PROJECT INFORMATION

PROJECT NAME: Weekly Storage Blanks

PROJECT #:

LOCATION:

PROJECT MANAGER:

E-MAIL:

PHONE:

FAX:

BILLING INFORMATION

BILL TO: Same

PO#

ADDRESS:

CITY:

STATE: ZIP:

ATTENTION:

PHONE:

DATA TURNAROUND INFORMATION

FAX: 10 DAYS*

HARD COPY: 10 DAYS*

EDD DAYS*

* TO BE APPROVED BY ALLIANCE

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ RESULTS ONLY (L1) ☐ US EPA CLP (L4)☐ RESULTS + QC (L2) ☐ NYS ASP "A" (L2)☐ NJ REDUCED (L3) ☐ NYS ASP "B" (L4)☐ NJ FULL (L4) ☐ Other☐ EDD Format

ANALYSIS

VOC-Drinking H2O

SVOC-Full+25-8270

PESTICIDE-TCL

PCB

1

2

3

4

5

6

7

8

9

PRESERVATIVES

COMMENTS

<- Specify Preservatives

A-HCl B-HNO3

C-H2SO4 D-NaOH

E-ICE F-Other

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles	PRESERVATIVES									COMMENTS ← Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-ICE F-Other	
			COMP	GRAB	DATE	TIME		A/E										
								1	2	3	4	5	6	7	8	9		
1.	PEST-GPC2-BLANK	SO		X			1				X							
2.	PEST-GPC2-BLANK-SPIKE	SO		X			1				X							
3.	PCB-GPC2-BLANK	SO		X			1					X						
4.	PCB-GPC2-BLANK-SPIKE	SO		X			1					X						
5.																		
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp _____
1. <i>Feb 2nd</i>		1.	MeOH extraction requires an additional 4oz. Jar for percent solid
RELINQUISHED BY	DATE/TIME	RECEIVED BY	Comments:
2.		2.	
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	
3.		3.	

Page 2 of 2

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
ALLIANCE: ☐ Picked Up ☐ Overnight

Shipment Complete
☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT

YELLOW - ALLIANCE COPY

PINK - SAMPLER COPY

3/7

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488