

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

EPA Sample No.: SSTD020665 Date Analyzed: 03/11/2025

Lab File ID: BP024155.D Time Analyzed: 14:18

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	293967	7.746	1241390	10.53	796171	14.39
UPPER LIMIT	587934	8.246	2482780	11.028	1592340	14.887
LOWER LIMIT	146984	7.246	620695	10.028	398086	13.887
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	291310	7.75	1178450	10.53	764143	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.