

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526

EPA Sample No.: SSTD020665 Date Analyzed: 03/11/2025

Lab File ID: BP024155.D Time Analyzed: 14:18

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1627840	17.198	1856170	21.657	2005100	25.057
UPPER LIMIT	3255680	17.698	3712340	22.157	4010200	25.557
LOWER LIMIT	813920	16.698	928085	21.157	1002550	24.557
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	1659160	17.20	1784880	21.66	1922730	25.06

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.