

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1526 SAS No.: Q1526 SDG NO.: Q1526
 Lab File ID: VX045257.D Date Analyzed: 03/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:24
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	92968	5.54	162999	6.76	140183	10.05
UPPER LIMIT	185936	6.044	325998	7.257	280366	10.549
LOWER LIMIT	46484	5.044	81499.5	6.257	70091.5	9.549
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	70415	5.55	135839	6.76	124026	10.06
Storage-Blank-WATER-REF-5	69522	5.55	134711	6.76	125045	10.05
Storage-Blank-WATER-REF-4	64762	5.55	126801	6.76	117950	10.06
Storage-Blank-SAMPLE-MANAGEMENT	65435	5.54	128551	6.76	119631	10.06
VX0313WBL01	67746	5.54	133586	6.76	123621	10.06
VX0313WBS01	97410	5.55	174469	6.76	153760	10.06
VX0313WBSD01	90235	5.55	159386	6.76	144055	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.