

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1531 SAS No.: Q1531 SDG No.: Q1531
 Instrument ID: BNA_N Calibration Date/Time: 03/12/2025 14:39
 Lab File ID: BN036599.D Init. Calib. Date(s): 03/10/2025 03/10/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:42 15:19
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.595	0.568		-4.5	50.0
Fluoranthene-d10	1.025	1.089		6.2	50.0
2-Fluorophenol	0.932	0.958		2.8	50.0
Phenol-d6	1.152	1.129		-2.0	50.0
Nitrobenzene-d5	0.435	0.442		1.6	50.0
2-Fluorobiphenyl	2.327	2.369		1.8	50.0
2,4,6-Tribromophenol	0.182	0.191		4.9	50.0
Terphenyl-d14	0.958	1.019		6.4	50.0
1,4-Dioxane	0.444	0.504		13.5	50.0

All other compounds must meet a minimum RRF of 0.010.