

### Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/07/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/07/25
Client Sample ID:	RE104D2-20250307	SDG No.:	Q1531
Lab Sample ID:	Q1531-19	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036594.D	1	03/10/25 11:35	03/12/25 11:02	PB167057

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
<b>TARGETS</b>							
123-91-1	1,4-Dioxane	2.30		0.070	0.20	0.20	ug/L
<b>SURROGATES</b>							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		74%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.46		30 - 150		116%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		76%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.32		53 - 106		81%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.54	*	58 - 132		136%	SPK: 0.4
<b>INTERNAL STANDARDS</b>							
3855-82-1	1,4-Dichlorobenzene-d4	1940		7.724			
1146-65-2	Naphthalene-d8	4250		10.509			
15067-26-2	Acenaphthene-d10	2600		14.366			
1517-22-2	Phenanthrene-d10	5270		17.111			
1719-03-5	Chrysene-d12	3680		21.295			
1520-96-3	Perylene-d12	3140		23.551			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products