

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031125\
 Data File : BN036589.D
 Acq On : 11 Mar 2025 19:46
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Mar 12 01:09:44 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	1964	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	4774	0.400	ng	0.00
13) Acenaphthene-d10	14.356	164	2543	0.400	ng	-0.01
19) Phenanthrene-d10	17.111	188	5156	0.400	ng	0.00
29) Chrysene-d12	21.295	240	3498	0.400	ng	0.00
35) Perylene-d12	23.554	264	2998	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.312	112	1585	0.346	ng	0.00
5) Phenol-d6	6.894	99	1840	0.325	ng	0.00
8) Nitrobenzene-d5	8.875	82	1860	0.358	ng	0.00
11) 2-Methylnaphthalene-d10	12.106	152	2744	0.386	ng	0.00
14) 2,4,6-Tribromophenol	15.858	330	450	0.390	ng	0.00
15) 2-Fluorobiphenyl	12.988	172	6396	0.432	ng	0.00
27) Fluoranthene-d10	19.141	212	5306	0.402	ng	0.00
31) Terphenyl-d14	19.745	244	3310	0.395	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.239	88	897	0.412	ng	98
3) n-Nitrosodimethylamine	3.550	42	1761	0.400	ng	95
6) bis(2-Chloroethyl)ether	7.147	93	1905	0.326	ng	99
9) Naphthalene	10.562	128	5442	0.388	ng	100
10) Hexachlorobutadiene	10.851	225	1336	0.404	ng	# 97
12) 2-Methylnaphthalene	12.182	142	3430	0.384	ng	100
16) Acenaphthylene	14.078	152	4983	0.415	ng	99
17) Acenaphthene	14.420	154	3052	0.389	ng	99
18) Fluorene	15.414	166	4246	0.400	ng	100
20) 4,6-Dinitro-2-methylph...	15.499	198	375	0.430	ng	95
21) 4-Bromophenyl-phenylether	16.305	248	1311	0.406	ng	95
22) Hexachlorobenzene	16.416	284	1646	0.422	ng	98
23) Atrazine	16.578	200	1043	0.403	ng	98
24) Pentachlorophenol	16.764	266	661	0.372	ng	100
25) Phenanthrene	17.149	178	6304	0.408	ng	99
26) Anthracene	17.248	178	5453	0.391	ng	99
28) Fluoranthene	19.169	202	7017	0.404	ng	99
30) Pyrene	19.536	202	7087	0.414	ng	100
32) Benzo(a)anthracene	21.286	228	4721	0.388	ng	100
33) Chrysene	21.331	228	5431	0.409	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	3858	0.445	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.835	276	4281	0.396	ng	95
37) Benzo(b)fluoranthene	22.920	252	4544	0.416	ng	97
38) Benzo(k)fluoranthene	22.920	252	4544	0.397	ng	98
39) Benzo(a)pyrene	23.455	252	3929	0.428	ng	# 92
40) Dibenzo(a,h)anthracene	25.852	278	3325	0.395	ng	98
41) Benzo(g,h,i)perylene	26.528	276	3963	0.411	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031125\
Data File : BN036589.D
Acq On : 11 Mar 2025 19:46
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Quant Time: Mar 12 01:09:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 16:06:28 2025
Response via : Initial Calibration







#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.724 min Scan# 64
 Delta R.T. 0.000 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC







Tgt Ion:152 Resp: 1964
 Ion Ratio Lower Upper
 152 100
 150 151.7 123.7 185.5
 115 66.3 54.3 81.5



#2
 1,4-Dioxane
 Concen: 0.412 ng
 RT: 3.239 min Scan# 22
 Delta R.T. 0.000 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46

Tgt Ion: 88 Resp: 897
 Ion Ratio Lower Upper
 88 100
 43 48.8 37.8 56.8
 58 86.1 67.4 101.2








#3
 n-Nitrosodimethylamine
 Concen: 0.400 ng
 RT: 3.550 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	42	Resp:	1761
Ion	Ratio	Lower	Upper
42	100		
74	71.6	60.6	90.8
44	8.3	6.3	9.5



#4
 2-Fluorophenol
 Concen: 0.346 ng
 RT: 5.312 min Scan# 309
 Delta R.T. 0.000 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46

Tgt Ion:	112	Resp:	1585
Ion	Ratio	Lower	Upper
112	100		
64	68.8	53.1	79.7
63	42.3	31.8	47.8








#5
Phenol-d6
Concen: 0.325 ng
RT: 6.894 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 99 Resp: 1840
Ion Ratio Lower Upper
99 100
42 36.8 26.5 39.7
71 46.3 34.1 51.1



#6
bis(2-Chloroethyl)ether
Concen: 0.326 ng
RT: 7.147 min Scan# 563
Delta R.T. -0.007 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Tgt Ion: 93 Resp: 1905
Ion Ratio Lower Upper
93 100
63 85.9 67.7 101.5
95 31.6 25.6 38.4








#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.509 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:	136	Resp:	4774
Ion	Ratio	Lower	Upper
136	100		
137	12.9	10.3	15.5
54	13.1	11.5	17.3
68	8.0	7.0	10.4



#8
Nitrobenzene-d5
Concen: 0.358 ng
RT: 8.875 min Scan# 777
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Tgt Ion:	82	Resp:	1860
Ion	Ratio	Lower	Upper
82	100		
128	39.2	30.6	45.8
54	65.3	52.2	78.4








#9
Naphthalene
Concen: 0.388 ng
RT: 10.562 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:	128	Resp:	5442
Ion Ratio	Lower	Upper	
128	100		
129	12.1	9.8	14.6
127	14.6	11.8	17.8



#10
Hexachlorobutadiene
Concen: 0.404 ng
RT: 10.851 min Scan# 962
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Tgt Ion:	225	Resp:	1336
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	62.6	51.8	77.8








#11
2-Methylnaphthalene-d10
Concen: 0.386 ng
RT: 12.106 min Scan# 1141
Delta R.T. -0.005 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:152 Resp: 2744
Ion Ratio Lower Upper
152 100
151 21.1 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.384 ng
RT: 12.182 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Tgt Ion:142 Resp: 3430
Ion Ratio Lower Upper
142 100
141 90.1 71.7 107.5
115 48.3 38.3 57.5








#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.356 min Scan# 1445
 Delta R.T. -0.011 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:164 Resp: 2543
 Ion Ratio Lower Upper
 164 100
 162 106.4 84.2 126.2
 160 56.8 42.2 63.2



#14
 2,4,6-Tribromophenol
 Concen: 0.390 ng
 RT: 15.858 min Scan# 1581
 Delta R.T. 0.000 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46

Tgt Ion:330 Resp: 450
 Ion Ratio Lower Upper
 330 100
 332 98.7 75.2 112.8
 141 54.2 43.4 65.2









#15
2-Fluorobiphenyl
Concen: 0.432 ng
RT: 12.988 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:172 Resp: 6396
Ion Ratio Lower Upper
172 100
171 36.6 29.5 44.3
170 25.9 20.2 30.4



#16
Acenaphthylene
Concen: 0.415 ng
RT: 14.078 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Tgt Ion:152 Resp: 4983
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.4
153 12.7 10.6 15.8








#17
Acenaphthene
Concen: 0.389 ng
RT: 14.420 min Scan# 1452
Delta R.T. -0.011 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:154 Resp: 3052
Ion Ratio Lower Upper
154 100
153 118.9 94.1 141.1
152 63.0 49.8 74.6



#18
Fluorene
Concen: 0.400 ng
RT: 15.414 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Tgt Ion:166 Resp: 4246
Ion Ratio Lower Upper
166 100
165 100.0 79.8 119.8
167 13.3 10.6 15.8








#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 5156
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.1 8.8 13.2



#20
4,6-Dinitro-2-methylphenol
Concen: 0.430 ng
RT: 15.499 min Scan# 1552
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Tgt Ion:198 Resp: 375
Ion Ratio Lower Upper
198 100
51 126.6 107.9 161.9
105 69.0 56.2 84.2









#21
4-Bromophenyl-phenylether
Concen: 0.406 ng
RT: 16.305 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 1311
Ion Ratio Lower Upper
248 100
250 95.8 73.0 109.6
141 81.4 68.6 103.0



#22
Hexachlorobenzene
Concen: 0.422 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Tgt Ion:284 Resp: 1646
Ion Ratio Lower Upper
284 100
142 48.7 37.0 55.4
249 34.7 28.1 42.1





#23

Atrazine

Concen: 0.403 ng

RT: 16.578 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036589.D

Acq: 11 Mar 2025 19:46

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:200 Resp: 1043

Ion Ratio Lower Upper

200 100

173 35.1 27.3 40.9

215 47.0 36.8 55.2



#24

Pentachlorophenol

Concen: 0.372 ng

RT: 16.764 min Scan# 1654

Delta R.T. -0.012 min

Lab File: BN036589.D

Acq: 11 Mar 2025 19:46




Tgt Ion:266 Resp: 661


Ion Ratio Lower Upper

266 100

264 62.6 49.6 74.4

268 63.5 50.9 76.3









#25
Phenanthrene
Concen: 0.408 ng
RT: 17.149 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:178 Resp: 6304
Ion Ratio Lower Upper
178 100
176 19.4 15.9 23.9
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.391 ng
RT: 17.248 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Tgt Ion:178 Resp: 5453
Ion Ratio Lower Upper
178 100
176 19.4 15.4 23.2
179 15.3 12.6 18.8









#27
Fluoranthene-d10
Concen: 0.402 ng
RT: 19.141 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 5306
Ion Ratio Lower Upper
212 100
106 13.5 11.8 17.6
104 8.7 7.3 10.9



#28
Fluoranthene
Concen: 0.404 ng
RT: 19.169 min Scan# 1936
Delta R.T. -0.005 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Tgt Ion:202 Resp: 7017
Ion Ratio Lower Upper
202 100
101 10.6 9.4 14.0
203 16.7 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.295 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036589.D

Acq: 11 Mar 2025 19:46

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:240 Resp: 3498

Ion Ratio Lower Upper

240 100

120 18.6 14.6 22.0

236 30.4 24.1 36.1



#30

Pyrene

Concen: 0.414 ng

RT: 19.536 min Scan# 2015

Delta R.T. 0.000 min

Lab File: BN036589.D

Acq: 11 Mar 2025 19:46




Tgt Ion:202 Resp: 7087

Ion Ratio Lower Upper

202 100

200 21.4 17.1 25.7

203 17.4 14.1 21.1








#31
 Terphenyl-d14
 Concen: 0.395 ng
 RT: 19.745 min Scan# 2060
 Delta R.T. 0.000 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244 Resp: 3310
 Ion Ratio Lower Upper
 244 100
 212 12.7 9.6 14.4
 122 17.0 13.9 20.9



#32
 Benzo(a)anthracene
 Concen: 0.388 ng
 RT: 21.286 min Scan# 2259
 Delta R.T. 0.000 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46

Tgt Ion:228 Resp: 4721
 Ion Ratio Lower Upper
 228 100
 226 28.3 22.5 33.7
 229 20.7 16.6 25.0





#33

Chrysene

Concen: 0.409 ng

RT: 21.331 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036589.D

Acq: 11 Mar 2025 19:46

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:228 Resp: 5431

Ion Ratio Lower Upper

228 100

226 32.3 25.3 37.9

229 20.6 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.445 ng

RT: 21.214 min Scan# 2251

Delta R.T. 0.000 min

Lab File: BN036589.D

Acq: 11 Mar 2025 19:46




Tgt Ion:149 Resp: 3858

Ion Ratio Lower Upper

149 100

167 26.0 20.7 31.1

279 3.2 3.6 5.4#








#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.554 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	264	Resp:	2998
Ion Ratio	Lower	Upper	
264	100		
260	28.5	22.6	33.8
265	101.7	88.1	132.1



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.396 ng
 RT: 25.835 min Scan# 3557
 Delta R.T. -0.003 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46

Tgt Ion:	276	Resp:	4281
Ion Ratio	Lower	Upper	
276	100		
138	25.6	23.4	35.2
277	23.5	20.0	30.0








#37
Benzo(b)fluoranthene
Concen: 0.416 ng
RT: 22.920 min Scan# 2560
Delta R.T. 0.047 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:252 Resp: 4544
Ion Ratio Lower Upper
252 100
253 30.5 23.9 35.9
125 19.7 17.4 26.2



#38
Benzo(k)fluoranthene
Concen: 0.397 ng
RT: 22.920 min Scan# 2560
Delta R.T. 0.003 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46


Tgt Ion:252 Resp: 4544
Ion Ratio Lower Upper
252 100
253 30.5 24.6 36.8
125 19.7 17.8 26.8







#39
Benzo(a)pyrene
Concen: 0.428 ng
RT: 23.455 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC






Tgt Ion:252 Resp: 3929
Ion Ratio Lower Upper
252 100
253 36.3 27.8 41.8
125 36.4 22.7 34.1#



#40
Dibenzo(a,h)anthracene
Concen: 0.395 ng
RT: 25.852 min Scan# 3563
Delta R.T. -0.003 min
Lab File: BN036589.D
Acq: 11 Mar 2025 19:46


Tgt Ion:278 Resp: 3325
Ion Ratio Lower Upper
278 100
139 24.3 20.8 31.2
279 36.9 28.8 43.2





#41
 Benzo(g,h,i)perylene
 Concen: 0.411 ng
 RT: 26.528 min Scan# 31
 Delta R.T. -0.003 min
 Lab File: BN036589.D
 Acq: 11 Mar 2025 19:46

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC



Tgt Ion:	276	Resp:	3963
Ion	Ratio	Lower	Upper
276	100		
277	26.8	22.2	33.4
138	28.0	24.1	36.1

