



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1531

Client: AECOM Technical Services, Inc.

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1531-08MS	Client Sample ID:	RE103D1-20250306MS					DataFile:	BN036579.D		
1,4-Dioxane	0.41	6.50	7.20	ug/L	171	*			70	130	