

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1531 SAS No.: Q1531 SDG NO.: Q1531
Lab File ID: VX045183.D Date Analyzed: 03/10/2025
Instrument ID: MSVOA_X Time Analyzed: 10:06
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	77718	12.018				
UPPER LIMIT	155436	12.518				
LOWER LIMIT	38859	11.518				
EPA SAMPLE NO.						
RE122D2-20250305	58126	12.02				
RE122D1-20250305	53642	12.02				
RE126D1-20250306	54115	12.02				
RE126D2-20250306	56943	12.02				
RE103D1-20250306	54358	12.02				
RE103D1-20250306MS	65455	12.02				
RE103D1-20250306MSD	60207	12.02				
RE103D3-20250306	51747	12.02				
RE103D2-20250306	50043	12.02				
RE108D1-20250306	58912	12.02				
RE105D1-20250306	57421	12.02				
RE125D1-20250307	54697	12.02				
RE125D2-20250307	52991	12.02				
VX0310WBL01	59432	12.02				
VX0310WBS01	74969	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.