

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15  
 Lab Code: CHEM Case No.: Q1531 SAS No.: Q1531 SDG NO.: Q1531  
 Lab File ID: VX045211.D Date Analyzed: 03/11/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:59  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                    | IS4<br>AREA # | RT #   |  |  |  |  |
|--------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD        | 70136         | 12.018 |  |  |  |  |
| UPPER LIMIT        | 140272        | 12.518 |  |  |  |  |
| LOWER LIMIT        | 35068         | 11.518 |  |  |  |  |
| EPA SAMPLE NO.     |               |        |  |  |  |  |
| RE122D3-20250305   | 53763         | 12.02  |  |  |  |  |
| RE122D1-20250305DL | 48152         | 12.02  |  |  |  |  |
| RE126D2-20250306DL | 53967         | 12.02  |  |  |  |  |
| RE103D1-20250306DL | 52396         | 12.02  |  |  |  |  |
| RE103D2-20250306DL | 53980         | 12.02  |  |  |  |  |
| RE108D2-20250306   | 51036         | 12.02  |  |  |  |  |
| RE105D2-20250306   | 51223         | 12.02  |  |  |  |  |
| RE125D1-20250307DL | 47303         | 12.02  |  |  |  |  |
| RE125D2-20250307DL | 52578         | 12.02  |  |  |  |  |
| FB01               | 56102         | 12.02  |  |  |  |  |
| RE104D2-20250307   | 53817         | 12.02  |  |  |  |  |
| TB                 | 51680         | 12.02  |  |  |  |  |
| VX0311WBL01        | 50068         | 12.02  |  |  |  |  |
| VX0311WBS01        | 71836         | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.