

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1531 SAS No.: Q1531 SDG NO.: Q1531
 Lab File ID: VX045238.D Date Analyzed: 03/12/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:58
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	69105	12.018				
UPPER LIMIT	138210	12.518				
LOWER LIMIT	34552.5	11.518				
EPA SAMPLE NO.						
DUP01-20250306	51386	12.02				
DUP01-20250306DL	48860	12.02				
RE108D2-20250306DL	43787	12.02				
RE105D2-20250306DL	48309	12.02				
RE104D2-20250307DL	50133	12.02				
RE104D3-20250307	50164	12.02				
VX0312WBL01	43950	12.02				
VX0312WBS01	69454	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.