

LAB CHRONICLE

OrderID:	Q1532	OrderDate:	3/10/2025 8:40:33 AM
Client:	Environmental Management Services	Project:	36-12 37th St. Queens, NY
Contact:	Kosta Kamberis, PE	Location:	F11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1532-01	1	Air	TO-15	TO-15	03/06/25		03/10/25	03/07/25
Q1532-02	2	Air	TO-15	TO-15	03/06/25		03/10/25	03/07/25
Q1532-03	3	Air	TO-15	TO-15	03/06/25		03/10/25	03/07/25
Q1532-04	4	Air	TO-15	TO-15	03/06/25		03/10/25	03/07/25

Hit Summary Sheet SW-846

SDG No.: Q1532

Client: Environmental Management Services

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	1							
Q1532-01	1	Air	Dichlorodifluoromethane	2.52		1.04	2.47	ug/m3
Q1532-01	1	Air	Chloromethane	1.09		0.21	1.03	ug/m3
Q1532-01	1	Air	Trichlorofluoromethane	1.46	J	0.90	2.81	ug/m3
Q1532-01	1	Air	Acetone	17.1		0.57	1.19	ug/m3
Q1532-01	1	Air	Methylene Chloride	5.91		0.66	1.74	ug/m3
Q1532-01	1	Air	2-Butanone	1.92		0.35	1.47	ug/m3
Q1532-01	1	Air	Carbon Tetrachloride	0.44		0.060	0.19	ug/m3
Q1532-01	1	Air	Benzene	0.42	J	0.29	1.60	ug/m3
Q1532-01	1	Air	Toluene	1.51	J	0.41	1.88	ug/m3
Q1532-01	1	Air	Tetrachloroethene	0.81		0.14	0.20	ug/m3
Q1532-01	1	Air	m/p-Xylene	1.48	J	0.91	4.34	ug/m3
			Total Voc :		34.7			
			Total Concentration:		34.7			
Client ID:	2							
Q1532-02	2	Air	Dichlorodifluoromethane	2.52		1.04	2.47	ug/m3
Q1532-02	2	Air	Chloromethane	2.07		0.21	1.03	ug/m3
Q1532-02	2	Air	Trichlorofluoromethane	1.35	J	0.90	2.81	ug/m3
Q1532-02	2	Air	Acetone	18.3		0.57	1.19	ug/m3
Q1532-02	2	Air	Methylene Chloride	4.17		0.66	1.74	ug/m3
Q1532-02	2	Air	2-Butanone	2.89		0.35	1.47	ug/m3
Q1532-02	2	Air	Carbon Tetrachloride	0.50		0.060	0.19	ug/m3
Q1532-02	2	Air	Chloroform	0.49	J	0.39	2.44	ug/m3
Q1532-02	2	Air	Benzene	0.45	J	0.29	1.60	ug/m3
Q1532-02	2	Air	Toluene	1.81	J	0.41	1.88	ug/m3
Q1532-02	2	Air	Tetrachloroethene	0.27		0.14	0.20	ug/m3
Q1532-02	2	Air	m/p-Xylene	1.35	J	0.91	4.34	ug/m3
			Total Voc :		36.2			
			Total Concentration:		36.2			
Client ID:	3							
Q1532-03	3	Air	Dichlorodifluoromethane	2.52		1.04	2.47	ug/m3
Q1532-03	3	Air	Chloromethane	1.20		0.21	1.03	ug/m3
Q1532-03	3	Air	Trichlorofluoromethane	1.40	J	0.90	2.81	ug/m3
Q1532-03	3	Air	Acetone	15.4		0.57	1.19	ug/m3
Q1532-03	3	Air	Methylene Chloride	5.21		0.66	1.74	ug/m3
Q1532-03	3	Air	2-Butanone	2.04		0.35	1.47	ug/m3
Q1532-03	3	Air	Carbon Tetrachloride	0.44		0.060	0.19	ug/m3

Hit Summary Sheet SW-846

SDG No.: Q1532

Client: Environmental Management Services

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q1532-03	3	Air	Benzene	0.45	J	0.29	1.60	ug/m3
Q1532-03	3	Air	Toluene	1.43	J	0.41	1.88	ug/m3
Q1532-03	3	Air	Tetrachloroethene	0.47		0.14	0.20	ug/m3
Q1532-03	3	Air	m/p-Xylene	1.17	J	0.91	4.34	ug/m3
Total Voc :				31.8				
Total Concentration:				31.8				
Client ID: 4	4							
Q1532-04	4	Air	Dichlorodifluoromethane	2.52		1.04	2.47	ug/m3
Q1532-04	4	Air	Chloromethane	1.18		0.21	1.03	ug/m3
Q1532-04	4	Air	Trichlorofluoromethane	1.46	J	0.90	2.81	ug/m3
Q1532-04	4	Air	Acetone	13.8		0.57	1.19	ug/m3
Q1532-04	4	Air	Methylene Chloride	11.8		0.66	1.74	ug/m3
Q1532-04	4	Air	2-Butanone	1.62		0.35	1.47	ug/m3
Q1532-04	4	Air	Carbon Tetrachloride	0.44		0.060	0.19	ug/m3
Q1532-04	4	Air	Benzene	0.48	J	0.29	1.60	ug/m3
Q1532-04	4	Air	Toluene	1.66	J	0.41	1.88	ug/m3
Q1532-04	4	Air	Tetrachloroethene	0.41		0.14	0.20	ug/m3
Q1532-04	4	Air	m/p-Xylene	1.35	J	0.91	4.34	ug/m3
Q1532-04	4	Air	Hexane	1.52	J	0.39	1.76	ug/m3
Total Voc :				38.2				
Total Concentration:				38.2				

Project : 36-12 37th St. Queens, NY

Sampling Date : 03/06/25

Field Id Number : 1

Analysis Date : 03/10/25

Laboratory Id Number : Q1532-01

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.51		2.52		
Chloromethane	74-87-3	50.49	0.53		1.09		
Vinyl Chloride	75-01-4	62.5	0.03	U	0.08		
Bromomethane	74-83-9	94.94	0.5	U	1.94		
Chloroethane	75-00-3	64.52	0.5	U	1.32		
Tetrahydrofuran	109-99-9	72.11	0.5	U	1.47		
Trichlorofluoromethane	75-69-4	137.4	0.26	J	1.46		
Dichlorotetrafluoroethane	76-14-2	170.9	0.5	U	3.49		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.5	U	3.83		
tert-Butyl alcohol	75-65-0	74.12	0.5	U	1.52		
Heptane	142-82-5	100.2	0.5	U	2.05		
1,1-Dichloroethene	75-35-4	96.94	0.5	U	1.98		
Acetone	67-64-1	58.08	7.2		17.1		
Carbon Disulfide	75-15-0	76.14	0.5	U	1.56		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.5	U	1.8		
Methylene Chloride	75-09-2	84.94	1.7		5.91		
trans-1,2-Dichloroethene	156-60-5	96.94	0.5	U	1.98		
1,1-Dichloroethane	75-34-3	98.96	0.5	U	2.02		
Cyclohexane	110-82-7	84.16	0.5	U	1.72		
2-Butanone	78-93-3	72.11	0.65		1.92		
Carbon Tetrachloride	56-23-5	153.8	0.07		0.44		
cis-1,2-Dichloroethene	156-59-2	96.94	0.5	U	1.98		
Chloroform	67-66-3	119.4	0.5	U	2.44		
1,1,1-Trichloroethane	71-55-6	133.4	0.03	U	0.16		
2,2,4-Trimethylpentane	540-84-1	114.2	0.5	U	2.34		
Benzene	71-43-2	78.11	0.13	J	0.42		
1,2-Dichloroethane	107-06-2	98.96	0.5	U	2.02		
Trichloroethene	79-01-6	131.4	0.03	U	0.16		
1,2-Dichloropropane	78-87-5	113	0.5	U	2.31		
Bromodichloromethane	75-27-4	163.8	0.5	U	3.35		
4-Methyl-2-Pentanone	108-10-1	100.2	0.5	U	2.05		
Toluene	108-88-3	92.14	0.4	J	1.51		
t-1,3-Dichloropropene	10061-02-6	111	0.5	U	2.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.5	U	2.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.5	U	2.73		

Project : 36-12 37th St. Queens, NY

Sampling Date : 03/06/25

Field Id Number : 1

Analysis Date : 03/10/25

Laboratory Id Number : Q1532-01

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.5	U	4.26		
1,2-Dibromoethane	106-93-4	187.9	0.1	U	0.77		
Tetrachloroethene	127-18-4	165.8	0.12		0.81		
Chlorobenzene	108-90-7	112.6	0.5	U	2.3		
Ethyl Benzene	100-41-4	106.2	0.5	U	2.17		
m/p-Xylene	179601-23-1	106.2	0.34	J	1.48		
o-Xylene	95-47-6	106.2	0.5	U	2.17		
Styrene	100-42-5	104.1	0.5	U	2.13		
Bromoform	75-25-2	252.8	0.5	U	5.17		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.03	U	0.21		
2-Chlorotoluene	95-49-8	126.6	0.5	U	2.59		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.5	U	2.46		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.5	U	2.46		
1,3-Dichlorobenzene	541-73-1	147	0.5	U	3.01		
1,4-Dichlorobenzene	106-46-7	147	0.5	U	3.01		
1,2-Dichlorobenzene	95-50-1	147	0.5	U	3.01		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.5	U	3.71		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.5	U	5.33		
Naphthalene	91-20-3	128.17	0.1	U	0.52		
1,3-Butadiene	106-99-0	54.09	0.5	U	1.11		
4-Ethyltoluene	622-96-8	120.2	0.5	U	2.46		
Hexane	110-54-3	86.17	0.5	U	1.76		
Allyl Chloride	107-05-1	76.53	0.5	U	1.57		
1,4-Dioxane	123-91-1	88.12	0.5	U	1.8		
Methyl Methacrylate	80-62-6	100.117	0.5	U	2.05		

Project : 36-12 37th St. Queens, NY

Sampling Date : 03/06/25

Field Id Number : 2

Analysis Date : 03/10/25

Laboratory Id Number : Q1532-02

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.51		2.52		
Chloromethane	74-87-3	50.49	1		2.07		
Vinyl Chloride	75-01-4	62.5	0.03	U	0.08		
Bromomethane	74-83-9	94.94	0.5	U	1.94		
Chloroethane	75-00-3	64.52	0.5	U	1.32		
Tetrahydrofuran	109-99-9	72.11	0.5	U	1.47		
Trichlorofluoromethane	75-69-4	137.4	0.24	J	1.35		
Dichlorotetrafluoroethane	76-14-2	170.9	0.5	U	3.49		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.5	U	3.83		
tert-Butyl alcohol	75-65-0	74.12	0.5	U	1.52		
Heptane	142-82-5	100.2	0.5	U	2.05		
1,1-Dichloroethene	75-35-4	96.94	0.5	U	1.98		
Acetone	67-64-1	58.08	7.7		18.3		
Carbon Disulfide	75-15-0	76.14	0.5	U	1.56		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.5	U	1.8		
Methylene Chloride	75-09-2	84.94	1.2		4.17		
trans-1,2-Dichloroethene	156-60-5	96.94	0.5	U	1.98		
1,1-Dichloroethane	75-34-3	98.96	0.5	U	2.02		
Cyclohexane	110-82-7	84.16	0.5	U	1.72		
2-Butanone	78-93-3	72.11	0.98		2.89		
Carbon Tetrachloride	56-23-5	153.8	0.08		0.5		
cis-1,2-Dichloroethene	156-59-2	96.94	0.5	U	1.98		
Chloroform	67-66-3	119.4	0.1	J	0.49		
1,1,1-Trichloroethane	71-55-6	133.4	0.03	U	0.16		
2,2,4-Trimethylpentane	540-84-1	114.2	0.5	U	2.34		
Benzene	71-43-2	78.11	0.14	J	0.45		
1,2-Dichloroethane	107-06-2	98.96	0.5	U	2.02		
Trichloroethene	79-01-6	131.4	0.03	U	0.16		
1,2-Dichloropropane	78-87-5	113	0.5	U	2.31		
Bromodichloromethane	75-27-4	163.8	0.5	U	3.35		
4-Methyl-2-Pentanone	108-10-1	100.2	0.5	U	2.05		
Toluene	108-88-3	92.14	0.48	J	1.81		
t-1,3-Dichloropropene	10061-02-6	111	0.5	U	2.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.5	U	2.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.5	U	2.73		

Project : 36-12 37th St. Queens, NY

Sampling Date : 03/06/25

Field Id Number : 2

Analysis Date : 03/10/25

Laboratory Id Number : Q1532-02

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.5	U	4.26		
1,2-Dibromoethane	106-93-4	187.9	0.1	U	0.77		
Tetrachloroethene	127-18-4	165.8	0.04		0.27		
Chlorobenzene	108-90-7	112.6	0.5	U	2.3		
Ethyl Benzene	100-41-4	106.2	0.5	U	2.17		
m/p-Xylene	179601-23-1	106.2	0.31	J	1.35		
o-Xylene	95-47-6	106.2	0.5	U	2.17		
Styrene	100-42-5	104.1	0.5	U	2.13		
Bromoform	75-25-2	252.8	0.5	U	5.17		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.03	U	0.21		
2-Chlorotoluene	95-49-8	126.6	0.5	U	2.59		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.5	U	2.46		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.5	U	2.46		
1,3-Dichlorobenzene	541-73-1	147	0.5	U	3.01		
1,4-Dichlorobenzene	106-46-7	147	0.5	U	3.01		
1,2-Dichlorobenzene	95-50-1	147	0.5	U	3.01		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.5	U	3.71		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.5	U	5.33		
Naphthalene	91-20-3	128.17	0.1	U	0.52		
1,3-Butadiene	106-99-0	54.09	0.5	U	1.11		
4-Ethyltoluene	622-96-8	120.2	0.5	U	2.46		
Hexane	110-54-3	86.17	0.5	U	1.76		
Allyl Chloride	107-05-1	76.53	0.5	U	1.57		
1,4-Dioxane	123-91-1	88.12	0.5	U	1.8		
Methyl Methacrylate	80-62-6	100.117	0.5	U	2.05		

Project : 36-12 37th St. Queens, NY

Sampling Date : 03/06/25

Field Id Number : 3

Analysis Date : 03/10/25

Laboratory Id Number : Q1532-03

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.51		2.52		
Chloromethane	74-87-3	50.49	0.58		1.2		
Vinyl Chloride	75-01-4	62.5	0.03	U	0.08		
Bromomethane	74-83-9	94.94	0.5	U	1.94		
Chloroethane	75-00-3	64.52	0.5	U	1.32		
Tetrahydrofuran	109-99-9	72.11	0.5	U	1.47		
Trichlorofluoromethane	75-69-4	137.4	0.25	J	1.4		
Dichlorotetrafluoroethane	76-14-2	170.9	0.5	U	3.49		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.5	U	3.83		
tert-Butyl alcohol	75-65-0	74.12	0.5	U	1.52		
Heptane	142-82-5	100.2	0.5	U	2.05		
1,1-Dichloroethene	75-35-4	96.94	0.5	U	1.98		
Acetone	67-64-1	58.08	6.5		15.4		
Carbon Disulfide	75-15-0	76.14	0.5	U	1.56		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.5	U	1.8		
Methylene Chloride	75-09-2	84.94	1.5		5.21		
trans-1,2-Dichloroethene	156-60-5	96.94	0.5	U	1.98		
1,1-Dichloroethane	75-34-3	98.96	0.5	U	2.02		
Cyclohexane	110-82-7	84.16	0.5	U	1.72		
2-Butanone	78-93-3	72.11	0.69		2.04		
Carbon Tetrachloride	56-23-5	153.8	0.07		0.44		
cis-1,2-Dichloroethene	156-59-2	96.94	0.5	U	1.98		
Chloroform	67-66-3	119.4	0.5	U	2.44		
1,1,1-Trichloroethane	71-55-6	133.4	0.03	U	0.16		
2,2,4-Trimethylpentane	540-84-1	114.2	0.5	U	2.34		
Benzene	71-43-2	78.11	0.14	J	0.45		
1,2-Dichloroethane	107-06-2	98.96	0.5	U	2.02		
Trichloroethene	79-01-6	131.4	0.03	U	0.16		
1,2-Dichloropropane	78-87-5	113	0.5	U	2.31		
Bromodichloromethane	75-27-4	163.8	0.5	U	3.35		
4-Methyl-2-Pentanone	108-10-1	100.2	0.5	U	2.05		
Toluene	108-88-3	92.14	0.38	J	1.43		
t-1,3-Dichloropropene	10061-02-6	111	0.5	U	2.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.5	U	2.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.5	U	2.73		

Project : 36-12 37th St. Queens, NY

Sampling Date : 03/06/25

Field Id Number : 3

Analysis Date : 03/10/25

Laboratory Id Number : Q1532-03

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.5	U	4.26		
1,2-Dibromoethane	106-93-4	187.9	0.1	U	0.77		
Tetrachloroethene	127-18-4	165.8	0.07		0.47		
Chlorobenzene	108-90-7	112.6	0.5	U	2.3		
Ethyl Benzene	100-41-4	106.2	0.5	U	2.17		
m/p-Xylene	179601-23-1	106.2	0.27	J	1.17		
o-Xylene	95-47-6	106.2	0.5	U	2.17		
Styrene	100-42-5	104.1	0.5	U	2.13		
Bromoform	75-25-2	252.8	0.5	U	5.17		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.03	U	0.21		
2-Chlorotoluene	95-49-8	126.6	0.5	U	2.59		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.5	U	2.46		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.5	U	2.46		
1,3-Dichlorobenzene	541-73-1	147	0.5	U	3.01		
1,4-Dichlorobenzene	106-46-7	147	0.5	U	3.01		
1,2-Dichlorobenzene	95-50-1	147	0.5	U	3.01		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.5	U	3.71		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.5	U	5.33		
Naphthalene	91-20-3	128.17	0.1	U	0.52		
1,3-Butadiene	106-99-0	54.09	0.5	U	1.11		
4-Ethyltoluene	622-96-8	120.2	0.5	U	2.46		
Hexane	110-54-3	86.17	0.5	U	1.76		
Allyl Chloride	107-05-1	76.53	0.5	U	1.57		
1,4-Dioxane	123-91-1	88.12	0.5	U	1.8		
Methyl Methacrylate	80-62-6	100.117	0.5	U	2.05		

Project : 36-12 37th St. Queens, NY

Sampling Date : 03/06/25

Field Id Number : 4

Analysis Date : 03/10/25

Laboratory Id Number : Q1532-04

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.51		2.52		
Chloromethane	74-87-3	50.49	0.57		1.18		
Vinyl Chloride	75-01-4	62.5	0.03	U	0.08		
Bromomethane	74-83-9	94.94	0.5	U	1.94		
Chloroethane	75-00-3	64.52	0.5	U	1.32		
Tetrahydrofuran	109-99-9	72.11	0.5	U	1.47		
Trichlorofluoromethane	75-69-4	137.4	0.26	J	1.46		
Dichlorotetrafluoroethane	76-14-2	170.9	0.5	U	3.49		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.5	U	3.83		
tert-Butyl alcohol	75-65-0	74.12	0.5	U	1.52		
Heptane	142-82-5	100.2	0.5	U	2.05		
1,1-Dichloroethene	75-35-4	96.94	0.5	U	1.98		
Acetone	67-64-1	58.08	5.8		13.8		
Carbon Disulfide	75-15-0	76.14	0.5	U	1.56		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.5	U	1.8		
Methylene Chloride	75-09-2	84.94	3.4		11.8		
trans-1,2-Dichloroethene	156-60-5	96.94	0.5	U	1.98		
1,1-Dichloroethane	75-34-3	98.96	0.5	U	2.02		
Cyclohexane	110-82-7	84.16	0.5	U	1.72		
2-Butanone	78-93-3	72.11	0.55		1.62		
Carbon Tetrachloride	56-23-5	153.8	0.07		0.44		
cis-1,2-Dichloroethene	156-59-2	96.94	0.5	U	1.98		
Chloroform	67-66-3	119.4	0.5	U	2.44		
1,1,1-Trichloroethane	71-55-6	133.4	0.03	U	0.16		
2,2,4-Trimethylpentane	540-84-1	114.2	0.5	U	2.34		
Benzene	71-43-2	78.11	0.15	J	0.48		
1,2-Dichloroethane	107-06-2	98.96	0.5	U	2.02		
Trichloroethene	79-01-6	131.4	0.03	U	0.16		
1,2-Dichloropropane	78-87-5	113	0.5	U	2.31		
Bromodichloromethane	75-27-4	163.8	0.5	U	3.35		
4-Methyl-2-Pentanone	108-10-1	100.2	0.5	U	2.05		
Toluene	108-88-3	92.14	0.44	J	1.66		
t-1,3-Dichloropropene	10061-02-6	111	0.5	U	2.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.5	U	2.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.5	U	2.73		

Project : 36-12 37th St. Queens, NY

Sampling Date : 03/06/25

Field Id Number : 4

Analysis Date : 03/10/25

Laboratory Id Number : Q1532-04

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.5	U	4.26		
1,2-Dibromoethane	106-93-4	187.9	0.1	U	0.77		
Tetrachloroethene	127-18-4	165.8	0.06		0.41		
Chlorobenzene	108-90-7	112.6	0.5	U	2.3		
Ethyl Benzene	100-41-4	106.2	0.5	U	2.17		
m/p-Xylene	179601-23-1	106.2	0.31	J	1.35		
o-Xylene	95-47-6	106.2	0.5	U	2.17		
Styrene	100-42-5	104.1	0.5	U	2.13		
Bromoform	75-25-2	252.8	0.5	U	5.17		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.03	U	0.21		
2-Chlorotoluene	95-49-8	126.6	0.5	U	2.59		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.5	U	2.46		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.5	U	2.46		
1,3-Dichlorobenzene	541-73-1	147	0.5	U	3.01		
1,4-Dichlorobenzene	106-46-7	147	0.5	U	3.01		
1,2-Dichlorobenzene	95-50-1	147	0.5	U	3.01		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.5	U	3.71		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.5	U	5.33		
Naphthalene	91-20-3	128.17	0.1	U	0.52		
1,3-Butadiene	106-99-0	54.09	0.5	U	1.11		
4-Ethyltoluene	622-96-8	120.2	0.5	U	2.46		
Hexane	110-54-3	86.17	0.43	J	1.52		
Allyl Chloride	107-05-1	76.53	0.5	U	1.57		
1,4-Dioxane	123-91-1	88.12	0.5	U	1.8		
Methyl Methacrylate	80-62-6	100.117	0.5	U	2.05		

Method	Matrix	Parameter	Spike_Amt_ Added	Calculated MDL	MB Data	MB Datafile	New MDL TO Be Used	MDL Unit
TO-15	Air	1,1,1,2-Tetrachloroethane	0.4	0.090683682	0		0.091	ppbv
TO-15	Air	1,1,1-Trichloroethane	0.03	0.00947589	0		0.0095	ppbv
TO-15	Air	1,1,2,2-Tetrachloroethane	0.03	0.010183321	0		0.011	ppbv
TO-15	Air	1,1,2-Trichloroethane	0.4	0.070112514	0		0.071	ppbv
TO-15	Air	1,1,2-Trichlorotrifluoroethane	0.4	0.155560652	0		0.16	ppbv
TO-15	Air	1,1-Dichloroethane	0.4	0.122378743	0		0.13	ppbv
TO-15	Air	1,1-Dichloroethene	0.4	0.139735617	0		0.14	ppbv
TO-15	Air	1,2,4-Trichlorobenzene	0.4	0.084677012	0		0.085	ppbv
TO-15	Air	1,2,4-Trimethylbenzene	0.4	0.077496367	0		0.078	ppbv
TO-15	Air	1,2-Dibromoethane	0.4	0.071318853	0		0.072	ppbv
TO-15	Air	1,2-Dichlorobenzene	0.4	0.074250332	0		0.075	ppbv
TO-15	Air	1,2-Dichloroethane	0.4	0.090890005	0		0.091	ppbv
TO-15	Air	1,2-Dichloropropane	0.4	0.073346868	0.109	VL038919.D	0.11	ppbv
TO-15	Air	1,3,5-Trimethylbenzene	0.4	0.101198372	0		0.11	ppbv
TO-15	Air	1,3-Butadiene	0.4	0.124009143	0		0.13	ppbv
TO-15	Air	1,3-Dichlorobenzene	0.4	0.079013305	0		0.08	ppbv
TO-15	Air	1,4-Dichlorobenzene	0.4	0.052746151	0		0.053	ppbv
TO-15	Air	1,4-Dioxane	0.4	0.202123585	0		0.21	ppbv
TO-15	Air	2,2,4-Trimethylpentane	0.4	0.09988427	0		0.10	ppbv
TO-15	Air	2-Butanone	0.4	0.118687999	0		0.12	ppbv
TO-15	Air	2-Chlorotoluene	0.4	0.104141644	0		0.11	ppbv
TO-15	Air	2-Hexanone	0.4	0.080656229	0		0.081	ppbv
TO-15	Air	4-Ethyltoluene	0.4	0.112037051	0		0.12	ppbv
TO-15	Air	4-Methyl-2-Pentanone	0.4	0.092836933	0		0.093	ppbv
TO-15	Air	Acetone	0.4	0.234416891	0		0.24	ppbv
TO-15	Air	Allyl Chloride	0.4	0.145618654	0		0.15	ppbv
TO-15	Air	Benzene	0.4	0.086407689	0		0.087	ppbv
TO-15	Air	Benzyl Chloride	0.4	0.173895169	0		0.18	ppbv
TO-15	Air	Bromodichloromethane	0.4	0.076780313	0		0.077	ppbv
TO-15	Air	Bromoethene	0.4	0.115560124	0		0.12	ppbv
TO-15	Air	Bromoform	0.4	0.056947551	0		0.057	ppbv
TO-15	Air	Bromomethane	0.4	0.137678239	0		0.14	ppbv
TO-15	Air	Carbon Disulfide	0.4	0.157343925	0		0.16	ppbv
TO-15	Air	Carbon Tetrachloride	0.03	0.010617308	0		0.011	ppbv
TO-15	Air	Chlorobenzene	0.4	0.075202602	0		0.076	ppbv
TO-15	Air	Chlorodifluoromethane	0.4	0.087466815	0		0.088	ppbv
TO-15	Air	Chloroethane	0.4	0.163530318	0		0.17	ppbv
TO-15	Air	Chloroform	0.4	0.082232394	0		0.083	ppbv
TO-15	Air	Chloromethane	0.4	0.097292998	0		0.098	ppbv
TO-15	Air	cis-1,2-Dichloroethene	0.4	0.088332535	0		0.089	ppbv
TO-15	Air	cis-1,3-Dichloropropene	0.4	0.063248763	0		0.064	ppbv
TO-15	Air	Cyclohexane	0.4	0.217999816	0		0.22	ppbv
TO-15	Air	Dibromochloromethane	0.4	0.065703935	0		0.066	ppbv
TO-15	Air	Dichlorodifluoromethane	0.4	0.20957733	0		0.21	ppbv
TO-15	Air	Dichlorotetrafluoroethane	0.4	0.087348046	0		0.088	ppbv
TO-15	Air	Ethyl Acetate	0.4	0.078161123	0		0.079	ppbv
TO-15	Air	Ethyl Benzene	0.4	0.115413562	0		0.12	ppbv
TO-15	Air	Heptane	0.4	0.100800637	0		0.11	ppbv
TO-15	Air	Hexachloro-1,3-Butadiene	0.4	0.084169773	0		0.085	ppbv
TO-15	Air	Hexane	0.4	0.106485082	0		0.11	ppbv
TO-15	Air	Isopropyl Alcohol	0.4	0.068329134	0		0.069	ppbv
TO-15	Air	Isopropylbenzene	0.4	0.086744608	0		0.087	ppbv
TO-15	Air	m/p-Xylene	0.4	0.205831125	0		0.21	ppbv
TO-15	Air	Methyl Methacrylate	0.4	0.085528753	0		0.086	ppbv
TO-15	Air	Methyl tert-Butyl Ether	0.4	0.119546728	0		0.12	ppbv
TO-15	Air	Methylene Chloride	0.4	0.18405265	0.11	VL039022.D	0.19	ppbv
TO-15	Air	Naphthalene	0.4	0.074961199	0		0.075	ppbv
TO-15	Air	Naphthalene,2-methyl-	0.4	0.148496389	0		0.15	ppbv
TO-15	Air	n-Butylbenzene	0.4	0.08201719	0		0.083	ppbv
TO-15	Air	n-propylbenzene	0.4	0.0778963	0		0.078	ppbv
TO-15	Air	o-Xylene	0.4	0.117630235	0		0.12	ppbv
TO-15	Air	p-Isopropyltoluene	0.4	0.085704866	0		0.086	ppbv
TO-15	Air	Propene	0.4	0.143790814	0.14	VL039337.D	0.15	ppbv
TO-15	Air	sec-Butylbenzene	0.4	0.093164076	0		0.094	ppbv
TO-15	Air	Styrene	0.4	0.109444442	0		0.11	ppbv
TO-15	Air	t-1,3-Dichloropropene	0.4	0.059734415	0		0.06	ppbv
TO-15	Air	tert-Butyl alcohol	0.4	0.187711954	0		0.19	ppbv
TO-15	Air	tert-Butylbenzene	0.4	0.092760327	0		0.093	ppbv
TO-15	Air	Tetrachloroethene	0.03	0.014462962	0		0.015	ppbv
TO-15	Air	Tetrahydrofuran	0.4	0.122655048	0		0.13	ppbv
TO-15	Air	Toluene	0.4	0.109284	0		0.11	ppbv
TO-15	Air	trans-1,2-Dichloroethene	0.4	0.185264638	0		0.19	ppbv
TO-15	Air	Trichloroethene	0.03	0.016558175	0		0.017	ppbv
TO-15	Air	Trichlorofluoromethane	0.4	0.157465794	0		0.16	ppbv
TO-15	Air	Vinyl Acetate	0.4	0.152371485	0		0.16	ppbv
TO-15	Air	Vinyl Chloride	0.03	0.013585954	0		0.014	ppbv
TO-15	Air	Ethanol	0.4	0.317031286	0		0.32	ug/L

StddevVal	StudentsT Value	Instrument ID	Datafile1	Concentration1	Adate1	Datafile2	Concentration2
0.03485153	2.602	MSVOA_L	VL038602.D	0.51	03142022T18:52:00	VL038645.D	0.494
0.003641772	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.029
0.003913651	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.035
0.026945624	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.439
0.059785032	2.602	MSVOA_L	VL038602.D	0.47	03142022T18:52:00	VL038645.D	0.466
0.047032569	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.456
0.053703158	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.411
0.032543048	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.336
0.029783385	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.396
0.027409244	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.352
0.028535869	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.407
0.034930824	2.602	MSVOA_L	VL038602.D	0.41	03142022T18:52:00	VL038645.D	0.445
0.02818865	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.406
0.038892534	2.602	MSVOA_L	VL038602.D	0.33	03142022T18:52:00	VL038645.D	0.374
0.047659163	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.419
0.030366374	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.428
0.020271388	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.418
0.077680086	2.602	MSVOA_L	VL038602.D	0.46	03142022T18:52:00	VL038645.D	0.285
0.038387498	2.602	MSVOA_L	VL038602.D	0.39	03142022T18:52:00	VL038645.D	0.383
0.045614143	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.406
0.040023691	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.375
0.030997782	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.283
0.043058052	2.602	MSVOA_L	VL038602.D	0.34	03142022T18:52:00	VL038645.D	0.356
0.035679067	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.282
0.090091042	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.468
0.055964125	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.392
0.033208182	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.407
0.066831349	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.454
0.029508191	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.408
0.044412039	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.433
0.021886069	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.31
0.052912467	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.404
0.060470378	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.357
0.004080441	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.026
0.028901845	2.602	MSVOA_L	VL038602.D	0.48	03142022T18:52:00	VL038645.D	0.467
0.033615225	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.468
0.062847932	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.453
0.031603534	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.464
0.037391621	2.602	MSVOA_L	VL038602.D	0.47	03142022T18:52:00	VL038645.D	0.439
0.033947938	2.602	MSVOA_L	VL038602.D	0.36	03142022T18:52:00	VL038645.D	0.391
0.024307749	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.317
0.083781636	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.414
0.02525132	2.602	MSVOA_L	VL038602.D	0.39	03142022T18:52:00	VL038645.D	0.359
0.080544708	2.602	MSVOA_L	VL038602.D	0.49	03142022T18:52:00	VL038645.D	0.43
0.033569579	2.602	MSVOA_L	VL038602.D	0.5	03142022T18:52:00	VL038645.D	0.472
0.030038864	2.602	MSVOA_L	VL038602.D	0.39	03142022T18:52:00	VL038645.D	0.412
0.044355712	2.602	MSVOA_L	VL038602.D	0.33	03142022T18:52:00	VL038645.D	0.357
0.038739676	2.602	MSVOA_L	VL038602.D	0.35	03142022T18:52:00	VL038645.D	0.349
0.032348107	2.602	MSVOA_L	VL038602.D	0.48	03142022T18:52:00	VL038645.D	0.5
0.04092432	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.411
0.026260236	2.602	MSVOA_L	VL038602.D	0.47	03142022T18:52:00	VL038645.D	0.44
0.033337666	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.387
0.079104967	2.602	MSVOA_L	VL038602.D	0.76	03142022T18:52:00	VL038645.D	0.762
0.032870389	2.602	MSVOA_L	VL038602.D	0.31	03142022T18:52:00	VL038645.D	0.334
0.045944169	2.602	MSVOA_L	VL038602.D	0.41	03142022T18:52:00	VL038645.D	0.418
0.070735069	2.602	MSVOA_L	VL038602.D	0.49	03142022T18:52:00	VL038645.D	0.479
0.02880907	2.602	MSVOA_L	VL038602.D	0.25	03142022T18:52:00	VL038645.D	0.289
0.057070096	2.602	MSVOA_L	VL038602.D	0.14	03142022T18:52:00	VL038645.D	0.074
0.031520826	2.602	MSVOA_L	VL038602.D	0.35	03142022T18:52:00	VL038645.D	0.366
0.029937087	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.365
0.045207623	2.602	MSVOA_L	VL038602.D	0.36	03142022T18:52:00	VL038645.D	0.368
0.032938073	2.602	MSVOA_L	VL038602.D	0.35	03142022T18:52:00	VL038645.D	0.366
0.05526165	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.443
0.035804795	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.392
0.042061661	2.602	MSVOA_L	VL038602.D	0.26	03142022T18:52:00	VL038645.D	0.275
0.022957116	2.602	MSVOA_L	VL038602.D	0.27	03142022T18:52:00	VL038645.D	0.29
0.072141412	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.459
0.035649626	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.405
0.005558402	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.031
0.047138758	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.324
0.042	2.602	MSVOA_L	VL038602.D	0.31	03142022T18:52:00	VL038645.D	0.332
0.07120086	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.396
0.006363634	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.035
0.060517215	2.602	MSVOA_L	VL038602.D	0.46	03142022T18:52:00	VL038645.D	0.472
0.058559372	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.389
0.00522135	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.031
0.105747594	2.998	MSVOA_L	VL040210.D	0.438	02222023T00:52:00	VL040220.D	0.43

- A
- B
- C
- D**
- E
- F
- G
- H

Concentration5	Adate5	Datafile6	Concentration6	Adate6	Datafile7	Concentration7	Adate7
0.433	08182022T21:04:00	VL039600.D	0.463	08262022T12:33:00	VL039709.D	0.402	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.031	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.03	08182022T22:24:00	VL039586.D	0.029	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.404	08182022T21:04:00	VL039600.D	0.424	08262022T12:33:00	VL039709.D	0.412	10102022T23:30:00
0.448	08182022T21:04:00	VL039600.D	0.454	08262022T12:33:00	VL039709.D	0.438	10102022T23:30:00
0.426	08182022T21:04:00	VL039600.D	0.451	08262022T12:33:00	VL039709.D	0.445	10102022T23:30:00
0.394	08182022T21:04:00	VL039600.D	0.387	08262022T12:33:00	VL039709.D	0.427	10102022T23:30:00
0.314	08182022T21:04:00	VL039600.D	0.367	08262022T12:33:00	VL039709.D	0.373	10102022T23:30:00
0.342	08182022T21:04:00	VL039600.D	0.389	08262022T12:33:00	VL039709.D	0.389	10102022T23:30:00
0.361	08182022T21:04:00	VL039600.D	0.402	08262022T12:33:00	VL039709.D	0.375	10102022T23:30:00
0.408	08182022T21:04:00	VL039600.D	0.431	08262022T12:33:00	VL039709.D	0.448	10102022T23:30:00
0.411	08182022T21:04:00	VL039600.D	0.455	08262022T12:33:00	VL039709.D	0.418	10102022T23:30:00
0.444	08182022T21:04:00	VL039600.D	0.429	08262022T12:33:00	VL039709.D	0.373	10102022T23:30:00
0.331	08182022T21:04:00	VL039600.D	0.357	08262022T12:33:00	VL039709.D	0.342	10102022T23:30:00
0.405	08182022T21:04:00	VL039600.D	0.449	08262022T12:33:00	VL039709.D	0.422	10102022T23:30:00
0.368	08182022T21:04:00	VL039600.D	0.414	08262022T12:33:00	VL039709.D	0.399	10102022T23:30:00
0.404	08182022T21:04:00	VL039600.D	0.384	08262022T12:33:00	VL039709.D	0.439	10102022T23:30:00
0.373	08182022T21:04:00	VL039600.D	0.341	08262022T12:33:00	VL039709.D	0.418	10102022T23:30:00
0.348	08182022T21:04:00	VL039600.D	0.379	08262022T12:33:00	VL039709.D	0.392	10102022T23:30:00
0.314	08182022T21:04:00	VL039600.D	0.451	08262022T12:33:00	VL039709.D	0.354	10102022T23:30:00
0.312	08182022T21:04:00	VL039600.D	0.359	08262022T12:33:00	VL039709.D	0.348	10102022T23:30:00
0.27	08182022T21:04:00	VL039600.D	0.294	08262022T12:33:00	VL039709.D	0.326	10102022T23:30:00
0.284	08182022T21:04:00	VL039600.D	0.348	08262022T12:33:00	VL039709.D	0.355	10102022T23:30:00
0.322	08182022T21:04:00	VL039600.D	0.335	08262022T12:33:00	VL039709.D	0.32	10102022T23:30:00
0.363	08182022T21:04:00	VL039600.D	0.484	08262022T12:33:00	VL039709.D	0.5	10102022T23:30:00
0.418	08182022T21:04:00	VL039600.D	0.459	08262022T12:33:00	VL039709.D	0.379	10102022T23:30:00
0.345	08182022T21:04:00	VL039600.D	0.376	08262022T12:33:00	VL039709.D	0.398	10102022T23:30:00
0.394	08182022T21:04:00	VL039600.D	0.336	08262022T12:33:00	VL039709.D	0.356	10102022T23:30:00
0.382	08182022T21:04:00	VL039600.D	0.367	08262022T12:33:00	VL039709.D	0.377	10102022T23:30:00
0.408	08182022T21:04:00	VL039600.D	0.441	08262022T12:33:00	VL039709.D	0.388	10102022T23:30:00
0.305	08182022T21:04:00	VL039600.D	0.303	08262022T12:33:00	VL039709.D	0.293	10102022T23:30:00
0.364	08182022T21:04:00	VL039600.D	0.494	08262022T12:33:00	VL039709.D	0.464	10102022T23:30:00
0.399	08182022T21:04:00	VL039600.D	0.268	08262022T12:33:00	VL039709.D	0.332	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.027	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.431	08182022T21:04:00	VL039600.D	0.466	08262022T12:33:00	VL039709.D	0.46	10102022T23:30:00
0.456	08182022T21:04:00	VL039600.D	0.504	08262022T12:33:00	VL039709.D	0.446	10102022T23:30:00
0.462	08182022T21:04:00	VL039600.D	0.504	08262022T12:33:00	VL039709.D	0.531	10102022T23:30:00
0.423	08182022T21:04:00	VL039600.D	0.46	08262022T12:33:00	VL039709.D	0.411	10102022T23:30:00
0.392	08182022T21:04:00	VL039600.D	0.472	08262022T12:33:00	VL039709.D	0.485	10102022T23:30:00
0.36	08182022T21:04:00	VL039600.D	0.407	08262022T12:33:00	VL039709.D	0.438	10102022T23:30:00
0.277	08182022T21:04:00	VL039600.D	0.291	08262022T12:33:00	VL039709.D	0.284	10102022T23:30:00
0.376	08182022T21:04:00	VL039600.D	0.395	08262022T12:33:00	VL039709.D	0.383	10102022T23:30:00
0.339	08182022T21:04:00	VL039600.D	0.335	08262022T12:33:00	VL039709.D	0.301	10102022T23:30:00
0.62	08182022T21:04:00	VL039600.D	0.643	08262022T12:33:00	VL039709.D	0.694	10102022T23:30:00
0.454	08182022T21:04:00	VL039600.D	0.508	08262022T12:33:00	VL039709.D	0.473	10102022T23:30:00
0.341	08182022T21:04:00	VL039600.D	0.432	08262022T12:33:00	VL039709.D	0.395	10102022T23:30:00
0.297	08182022T21:04:00	VL039600.D	0.338	08262022T12:33:00	VL039709.D	0.334	10102022T23:30:00
0.347	08182022T21:04:00	VL039600.D	0.382	08262022T12:33:00	VL039709.D	0.407	10102022T23:30:00
0.447	08182022T21:04:00	VL039600.D	0.502	08262022T12:33:00	VL039709.D	0.468	10102022T23:30:00
0.385	08182022T21:04:00	VL039600.D	0.405	08262022T12:33:00	VL039709.D	0.435	10102022T23:30:00
0.458	08182022T21:04:00	VL039600.D	0.434	08262022T12:33:00	VL039709.D	0.481	10102022T23:30:00
0.361	08182022T21:04:00	VL039600.D	0.388	08262022T12:33:00	VL039709.D	0.405	10102022T23:30:00
0.673	08182022T21:04:00	VL039600.D	0.719	08262022T12:33:00	VL039709.D	0.688	10102022T23:30:00
0.3	08182022T21:04:00	VL039600.D	0.323	08262022T12:33:00	VL039709.D	0.346	10102022T23:30:00
0.337	08182022T21:04:00	VL039600.D	0.355	08262022T12:33:00	VL039709.D	0.396	10102022T23:30:00
0.536	08182022T21:04:00	VL039600.D	0.634	08262022T12:33:00	VL039709.D	0.499	10102022T23:30:00
0.28	08182022T21:04:00	VL039600.D	0.302	08262022T12:33:00	VL039709.D	0.291	10102022T23:30:00
0.158	08182022T21:04:00	VL039600.D	0.094	08262022T12:33:00	VL039709.D	0.149	10102022T23:30:00
0.321	08182022T21:04:00	VL039600.D	0.354	08262022T12:33:00	VL039709.D	0.358	10102022T23:30:00
0.332	08182022T21:04:00	VL039600.D	0.338	08262022T12:33:00	VL039709.D	0.347	10102022T23:30:00
0.315	08182022T21:04:00	VL039600.D	0.347	08262022T12:33:00	VL039709.D	0.354	10102022T23:30:00
0.316	08182022T21:04:00	VL039600.D	0.362	08262022T12:33:00	VL039709.D	0.358	10102022T23:30:00
0.499	08182022T21:04:00	VL039600.D	0.578	08262022T12:33:00	VL039709.D	0.533	10102022T23:30:00
0.359	08182022T21:04:00	VL039600.D	0.382	08262022T12:33:00	VL039709.D	0.371	10102022T23:30:00
0.23	08182022T21:04:00	VL039600.D	0.257	08262022T12:33:00	VL039709.D	0.278	10102022T23:30:00
0.252	08182022T21:04:00	VL039600.D	0.252	08262022T12:33:00	VL039709.D	0.273	10102022T23:30:00
0.384	08182022T21:04:00	VL039600.D	0.382	08262022T12:33:00	VL039709.D	0.454	10102022T23:30:00
0.345	08182022T21:04:00	VL039600.D	0.399	08262022T12:33:00	VL039709.D	0.371	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.038	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.299	08182022T21:04:00	VL039600.D	0.344	08262022T12:33:00	VL039709.D	0.389	10102022T23:30:00
0.288	08182022T21:04:00	VL039600.D	0.296	08262022T12:33:00	VL039709.D	0.33	10102022T23:30:00
0.375	08182022T21:04:00	VL039600.D	0.466	08262022T12:33:00	VL039709.D	0.437	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.031	08252022T20:23:00	VL039711.D	0.04	10112022T00:45:00
0.436	08182022T21:04:00	VL039600.D	0.47	08262022T12:33:00	VL039709.D	0.385	10102022T23:30:00
0.42	08182022T21:04:00	VL039600.D	0.483	08262022T12:33:00	VL039709.D	0.363	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.029	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.636	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.607	10042023T19:33:00

Datafile8	Concentration8	Adate8	Datafile9	Concentration9	Adate9	Datafile10	Concentration10
VL040134.D	0.42	01302023T15:54:00	VL040210.D	0.383	02222023T00:52:00	VL040220.D	0.42
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.038	02232023T14:26:00	VL040223.D	0.033
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.032	02232023T14:26:00	VL040223.D	0.027
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.395	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.476	02222023T00:52:00	VL040220.D	0.47
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.436	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.452	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.32	01302023T15:54:00	VL040210.D	0.299	02222023T00:52:00	VL040220.D	0.32
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.375	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.38	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.378	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.434	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.432	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.344	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.396	02222023T00:52:00	VL040220.D	0.41
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.395	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.402	02222023T00:52:00	VL040220.D	0.4
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.387	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.352	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.416	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.34	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.32	01302023T15:54:00	VL040210.D	0.318	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.321	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.337	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.45	01302023T15:54:00	VL040210.D	0.482	02222023T00:52:00	VL040220.D	0.5
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.411	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.359	02222023T00:52:00	VL040220.D	0.39
VL040134.D	0.29	01302023T15:54:00	VL040210.D	0.311	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.378	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.413	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.32	01302023T15:54:00	VL040210.D	0.32	02222023T00:52:00	VL040220.D	0.33
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.408	02222023T00:52:00	VL040220.D	0.49
VL040134.D	0.31	01302023T15:54:00	VL040210.D	0.391	02222023T00:52:00	VL040220.D	0.37
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.033	02232023T14:26:00	VL040223.D	0.037
VL040134.D	0.44	01302023T15:54:00	VL040210.D	0.414	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.474	02222023T00:52:00	VL040220.D	0.46
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.416	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.47	01302023T15:54:00	VL040210.D	0.445	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.46	01302023T15:54:00	VL040210.D	0.443	02222023T00:52:00	VL040220.D	0.47
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.435	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.33	01302023T15:54:00	VL040210.D	0.284	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.49	01302023T15:54:00	VL040210.D	0.521	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.343	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.48	01302023T15:54:00	VL040210.D	0.497	02222023T00:52:00	VL040220.D	0.49
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.431	02222023T00:52:00	VL040220.D	0.46
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.398	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.309	02222023T00:52:00	VL040220.D	0.33
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.336	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.421	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.4	02222023T00:52:00	VL040220.D	0.41
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.479	02222023T00:52:00	VL040220.D	0.48
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.361	02222023T00:52:00	VL040220.D	0.4
VL040134.D	0.71	01302023T15:54:00	VL040210.D	0.699	02222023T00:52:00	VL040220.D	0.75
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.314	02222023T00:52:00	VL040220.D	0.35
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.421	02222023T00:52:00	VL040220.D	0.41
VL040134.D	0.44	01302023T15:54:00	VL040210.D	0.558	02222023T00:52:00	VL040220.D	0.57
VL040134.D	0.27	01302023T15:54:00	VL040210.D	0.246	02222023T00:52:00	VL040220.D	0.27
VL040134.D	0.19	01302023T15:54:00	VL040210.D	0.116	02222023T00:52:00	VL040220.D	0.21
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.349	02222023T00:52:00	VL040220.D	0.35
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.354	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.331	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.33	01302023T15:54:00	VL040210.D	0.329	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.42	01302023T15:54:00	VL040210.D	0.432	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.352	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.29	01302023T15:54:00	VL040210.D	0.262	02222023T00:52:00	VL040220.D	0.27
VL040134.D	0.3	01302023T15:54:00	VL040210.D	0.292	02222023T00:52:00	VL040220.D	0.3
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.395	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.349	02222023T00:52:00	VL040220.D	0.36
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.029	02232023T14:26:00	VL040223.D	0.031
VL040134.D	0.42	01302023T15:54:00	VL040210.D	0.368	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.297	02222023T00:52:00	VL040220.D	0.33
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.397	02222023T00:52:00	VL040220.D	0.42
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.027	02232023T14:26:00	VL040223.D	0.028
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.451	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.438	02222023T00:52:00	VL040220.D	0.4
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.033	02232023T14:26:00	VL040223.D	0.038
VL040805.D	0.675	10252023T18:12:00					

Adate10	Datafile11	Concentration11	Adate11	Datafile12	Concentration1 2	Adate12	Datafile13
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.434	04122023T23:23:00	VL040559.D
02232023T15:04:00	VL040332.D	0.028	04102023T19:33:00	VL040346.D	0.027	04122023T19:50:00	VL040576.D
02232023T15:04:00	VL040332.D	0.023	04102023T19:33:00	VL040346.D	0.03	04122023T19:50:00	VL040576.D
02232023T13:12:00	VL040329.D	0.41	04102023T17:39:00	VL040349.D	0.403	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.42	04102023T17:39:00	VL040349.D	0.429	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.43	04102023T17:39:00	VL040349.D	0.4	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.44	04102023T17:39:00	VL040349.D	0.419	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.3	04102023T17:39:00	VL040349.D	0.302	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.34	04102023T17:39:00	VL040349.D	0.362	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.398	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.402	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.427	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.41	04102023T17:39:00	VL040349.D	0.385	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.33	04102023T17:39:00	VL040349.D	0.311	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.448	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.392	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.39	04102023T17:39:00	VL040349.D	0.402	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.33	04102023T17:39:00	VL040349.D	0.33	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.35	04102023T17:39:00	VL040349.D	0.356	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.451	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.32	04102023T17:39:00	VL040349.D	0.31	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.28	04102023T17:39:00	VL040349.D	0.29	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.32	04102023T17:39:00	VL040349.D	0.303	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.33	04102023T17:39:00	VL040349.D	0.339	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.58	04102023T17:39:00	VL040349.D	0.582	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.39	04102023T17:39:00	VL040349.D	0.402	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.365	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.421	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.397	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.41	04102023T17:39:00	VL040349.D	0.415	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.31	04102023T17:39:00	VL040349.D	0.337	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.44	04102023T17:39:00	VL040349.D	0.51	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.34	04102023T17:39:00	VL040349.D	0.342	04122023T23:23:00	VL040559.D
02232023T15:04:00	VL040332.D	0.028	04102023T19:33:00	VL040346.D	0.025	04122023T19:50:00	VL040576.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.433	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.453	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.4	04102023T17:39:00	VL040349.D	0.487	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.43	04102023T17:39:00	VL040349.D	0.456	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.47	04102023T17:39:00	VL040349.D	0.455	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.376	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.29	04102023T17:39:00	VL040349.D	0.287	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.29	04102023T17:39:00	VL040349.D	0.35	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.35	04102023T17:39:00	VL040349.D	0.339	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.51	04102023T17:39:00	VL040349.D	0.483	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.46	04102023T17:39:00	VL040349.D	0.467	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.39	04102023T17:39:00	VL040349.D	0.394	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.32	04102023T17:39:00	VL040349.D	0.31	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.33	04102023T17:39:00	VL040349.D	0.336	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.441	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.365	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.47	04102023T17:39:00	VL040349.D	0.442	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.363	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.66	04102023T17:39:00	VL040349.D	0.659	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.31	04102023T17:39:00	VL040349.D	0.305	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.41	04102023T17:39:00	VL040349.D	0.362	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.47	04102023T17:39:00	VL040349.D	0.533	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.26	04102023T17:39:00	VL040349.D	0.261	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.21	04102023T17:39:00	VL040349.D	0.19	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.33	04102023T17:39:00	VL040349.D	0.334	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.35	04102023T17:39:00	VL040349.D	0.345	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.31	04102023T17:39:00	VL040349.D	0.327	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.32	04102023T17:39:00	VL040349.D	0.314	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.42	04102023T17:39:00	VL040349.D	0.465	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.35	04102023T17:39:00	VL040349.D	0.335	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.26	04102023T17:39:00	VL040349.D	0.257	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.27	04102023T17:39:00	VL040349.D	0.256	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.46	04102023T17:39:00	VL040349.D	0.427	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.34	04102023T17:39:00	VL040349.D	0.327	04122023T23:23:00	VL040559.D
02232023T15:04:00	VL040332.D	0.022	04102023T19:33:00	VL040346.D	0.029	04122023T19:50:00	VL040576.D
02232023T13:12:00	VL040329.D	0.3	04102023T17:39:00	VL040349.D	0.347	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.29	04102023T17:39:00	VL040349.D	0.291	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.391	04122023T23:23:00	VL040559.D
02232023T15:04:00	VL040332.D	0.029	04102023T19:33:00	VL040346.D	0.028	04122023T19:50:00	VL040576.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.47	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.58	04122023T23:23:00	VL040559.D
02232023T15:04:00	VL040332.D	0.024	04102023T19:33:00	VL040346.D	0.029	04122023T19:50:00	VL040576.D

Concentration13	Adate13	Datafile14	Concentration14	Adate14	Datafile15	Concentration15	Adate15
0.48	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.488	10042023T19:33:00
0.03	07312023T17:18:00	VL040581.D	0.031	07312023T20:41:00	VL040772.D	0.035	10122023T01:48:00
0.031	07312023T17:18:00	VL040581.D	0.026	07312023T20:41:00	VL040772.D	0.029	10122023T01:48:00
0.459	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.405	10042023T19:33:00
0.573	07272023T03:27:00	VL040579.D	0.5	07312023T19:20:00	VL040747.D	0.557	10042023T19:33:00
0.535	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.527	10042023T19:33:00
0.527	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.481	10042023T19:33:00
0.33	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.305	10042023T19:33:00
0.389	07272023T03:27:00	VL040579.D	0.3	07312023T19:20:00	VL040747.D	0.361	10042023T19:33:00
0.425	07272023T03:27:00	VL040579.D	0.32	07312023T19:20:00	VL040747.D	0.359	10042023T19:33:00
0.442	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.45	10042023T19:33:00
0.519	07272023T03:27:00	VL040579.D	0.41	07312023T19:20:00	VL040747.D	0.426	10042023T19:33:00
0.452	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.42	10042023T19:33:00
0.33	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.321	10042023T19:33:00
0.412	07272023T03:27:00	VL040579.D	0.41	07312023T19:20:00	VL040747.D	0.466	10042023T19:33:00
0.388	07272023T03:27:00	VL040579.D	0.33	07312023T19:20:00	VL040747.D	0.412	10042023T19:33:00
0.438	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.382	10042023T19:33:00
0.373	07272023T03:27:00	VL040579.D	0.29	07312023T19:20:00	VL040747.D	0.257	10042023T19:33:00
0.398	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.335	10042023T19:33:00
0.448	07272023T03:27:00	VL040579.D	0.37	07312023T19:20:00	VL040747.D	0.443	10042023T19:33:00
0.35	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.325	10042023T19:33:00
0.303	07272023T03:27:00	VL040579.D	0.22	07312023T19:20:00	VL040747.D	0.283	10042023T19:33:00
0.32	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.317	10042023T19:33:00
0.348	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.302	10042023T19:33:00
0.674	07272023T03:27:00	VL040579.D	0.48	07312023T19:20:00	VL040747.D	0.65	10042023T19:33:00
0.436	07272023T03:27:00	VL040579.D	0.37	07312023T19:20:00	VL040747.D	0.511	10042023T19:33:00
0.397	07272023T03:27:00	VL040579.D	0.3	07312023T19:20:00	VL040747.D	0.386	10042023T19:33:00
0.425	07272023T03:27:00	VL040579.D	0.34	07312023T19:20:00	VL040747.D	0.307	10042023T19:33:00
0.467	07272023T03:27:00	VL040579.D	0.4	07312023T19:20:00	VL040747.D	0.409	10042023T19:33:00
0.508	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.506	10042023T19:33:00
0.352	07272023T03:27:00	VL040579.D	0.33	07312023T19:20:00	VL040747.D	0.349	10042023T19:33:00
0.543	07272023T03:27:00	VL040579.D	0.46	07312023T19:20:00	VL040747.D	0.496	10042023T19:33:00
0.46	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.477	10042023T19:33:00
0.036	07312023T17:18:00	VL040581.D	0.033	07312023T20:41:00	VL040772.D	0.031	10122023T01:48:00
0.481	07272023T03:27:00	VL040579.D	0.38	07312023T19:20:00	VL040747.D	0.489	10042023T19:33:00
0.526	07272023T03:27:00	VL040579.D	0.45	07312023T19:20:00	VL040747.D	0.55	10042023T19:33:00
0.516	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.576	10042023T19:33:00
0.521	07272023T03:27:00	VL040579.D	0.41	07312023T19:20:00	VL040747.D	0.501	10042023T19:33:00
0.486	07272023T03:27:00	VL040579.D	0.39	07312023T19:20:00	VL040747.D	0.544	10042023T19:33:00
0.405	07272023T03:27:00	VL040579.D	0.32	07312023T19:20:00	VL040747.D	0.431	10042023T19:33:00
0.295	07272023T03:27:00	VL040579.D	0.25	07312023T19:20:00	VL040747.D	0.298	10042023T19:33:00
0.301	07272023T03:27:00	VL040579.D	0.21	07312023T19:20:00	VL040747.D	0.317	10042023T19:33:00
0.403	07272023T03:27:00	VL040579.D	0.35	07312023T19:20:00	VL040747.D	0.381	10042023T19:33:00
0.54	07272023T03:27:00	VL040579.D	0.47	07312023T19:20:00	VL040747.D	0.562	10042023T19:33:00
0.516	07272023T03:27:00	VL040579.D	0.46	07312023T19:20:00	VL040747.D	0.509	10042023T19:33:00
0.424	07272023T03:27:00	VL040579.D	0.34	07312023T19:20:00	VL040747.D	0.441	10042023T19:33:00
0.32	07272023T03:27:00	VL040579.D	0.21	07312023T19:20:00	VL040747.D	0.299	10042023T19:33:00
0.341	07272023T03:27:00	VL040579.D	0.25	07312023T19:20:00	VL040747.D	0.355	10042023T19:33:00
0.466	07272023T03:27:00	VL040579.D	0.37	07312023T19:20:00	VL040747.D	0.479	10042023T19:33:00
0.38	07272023T03:27:00	VL040579.D	0.31	07312023T19:20:00	VL040747.D	0.465	10042023T19:33:00
0.451	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.426	10042023T19:33:00
0.379	07272023T03:27:00	VL040579.D	0.29	07312023T19:20:00	VL040747.D	0.386	10042023T19:33:00
0.735	07272023T03:27:00	VL040579.D	0.54	07312023T19:20:00	VL040747.D	0.682	10042023T19:33:00
0.283	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.324	10042023T19:33:00
0.468	07272023T03:27:00	VL040579.D	0.38	07312023T19:20:00	VL040747.D	0.472	10042023T19:33:00
0.619	07272023T03:27:00	VL040579.D	0.62	07312023T19:20:00	VL040747.D	0.625	10042023T19:33:00
0.274	07272023T03:27:00	VL040579.D	0.19	07312023T19:20:00	VL040747.D	0.258	10042023T19:33:00
0.107	07272023T03:27:00	VL040579.D	0.06	07312023T19:20:00	VL040747.D	0.133	10042023T19:33:00
0.361	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.343	10042023T19:33:00
0.355	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.369	10042023T19:33:00
0.355	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.301	10042023T19:33:00
0.357	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.313	10042023T19:33:00
0.436	07272023T03:27:00	VL040579.D	0.34	07312023T19:20:00	VL040747.D	0.471	10042023T19:33:00
0.393	07272023T03:27:00	VL040579.D	0.29	07312023T19:20:00	VL040747.D	0.354	10042023T19:33:00
0.242	07272023T03:27:00	VL040579.D	0.17	07312023T19:20:00	VL040747.D	0.266	10042023T19:33:00
0.283	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.312	10042023T19:33:00
0.539	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.394	10042023T19:33:00
0.366	07272023T03:27:00	VL040579.D	0.28	07312023T19:20:00	VL040747.D	0.381	10042023T19:33:00
0.025	07312023T17:18:00	VL040581.D	0.028	07312023T20:41:00	VL040772.D	0.044	10122023T01:48:00
0.353	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.351	10042023T19:33:00
0.285	07272023T03:27:00	VL040579.D	0.2	07312023T19:20:00	VL040747.D	0.288	10042023T19:33:00
0.477	07272023T03:27:00	VL040579.D	0.4	07312023T19:20:00	VL040747.D	0.507	10042023T19:33:00
0.027	07312023T17:18:00	VL040581.D	0.027	07312023T20:41:00	VL040772.D	0.039	10122023T01:48:00
0.589	07272023T03:27:00	VL040579.D	0.52	07312023T19:20:00	VL040747.D	0.561	10042023T19:33:00
0.487	07272023T03:27:00	VL040579.D	0.46	07312023T19:20:00	VL040747.D	0.456	10042023T19:33:00
0.03	07312023T17:18:00	VL040581.D	0.033	07312023T20:41:00	VL040772.D	0.041	10122023T01:48:00

Datafile16	Concentration16	Adate16	Department	AVG Concentration	Recovery
VL040805.D	0.428	10252023T18:12:00	VOC	0.45	110
VL040787.D	0.04	10232023T20:32:00	VOC	0.032	110
VL040787.D	0.04	10232023T20:32:00	VOC	0.03	100
VL040805.D	0.471	10252023T18:12:00	VOC	0.42	110
VL040805.D	0.476	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.427	10252023T18:12:00	VOC	0.46	110
VL040805.D	0.47	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.288	10252023T18:12:00	VOC	0.31	79
VL040805.D	0.435	10252023T18:12:00	VOC	0.37	93
VL040805.D	0.433	10252023T18:12:00	VOC	0.38	95
VL040805.D	0.405	10252023T18:12:00	VOC	0.41	100
VL040805.D	0.501	10252023T18:12:00	VOC	0.44	110
VL040805.D	0.456	10252023T18:12:00	VOC	0.42	100
VL040805.D	0.403	10252023T18:12:00	VOC	0.35	87
VL040805.D	0.455	10252023T18:12:00	VOC	0.44	110
VL040805.D	0.412	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.414	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.486	10252023T18:12:00	VOC	0.38	96
VL040805.D	0.434	10252023T18:12:00	VOC	0.37	93
VL040805.D	0.467	10252023T18:12:00	VOC	0.42	100
VL040805.D	0.42	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.353	10252023T18:12:00	VOC	0.3	75
VL040805.D	0.397	10252023T18:12:00	VOC	0.34	84
VL040805.D	0.403	10252023T18:12:00	VOC	0.34	84
VL040805.D	0.576	10252023T18:12:00	VOC	0.53	130
VL040805.D	0.422	10252023T18:12:00	VOC	0.43	110
VL040805.D	0.445	10252023T18:12:00	VOC	0.38	96
VL040805.D	0.205	10252023T18:12:00	VOC	0.36	91
VL040805.D	0.458	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.436	10252023T18:12:00	VOC	0.44	110
VL040805.D	0.34	10252023T18:12:00	VOC	0.33	82
VL040805.D	0.508	10252023T18:12:00	VOC	0.47	120
VL040805.D	0.361	10252023T18:12:00	VOC	0.38	95
VL040787.D	0.04	10232023T20:32:00	VOC	0.031	100
VL040805.D	0.454	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.517	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.435	10252023T18:12:00	VOC	0.47	120
VL040805.D	0.487	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.458	10252023T18:12:00	VOC	0.46	120
VL040805.D	0.435	10252023T18:12:00	VOC	0.4	99
VL040805.D	0.335	10252023T18:12:00	VOC	0.3	75
VL040805.D	0.416	10252023T18:12:00	VOC	0.39	96
VL040805.D	0.384	10252023T18:12:00	VOC	0.36	89
VL040805.D	0.516	10252023T18:12:00	VOC	0.54	140
VL040805.D	0.505	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.404	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.406	10252023T18:12:00	VOC	0.33	82
VL040805.D	0.404	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.446	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.452	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.442	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.435	10252023T18:12:00	VOC	0.38	96
VL040805.D	0.888	10252023T18:12:00	VOC	0.72	180
VL040805.D	0.348	10252023T18:12:00	VOC	0.32	80
VL040805.D	0.353	10252023T18:12:00	VOC	0.41	100
VL040805.D	0.619	10252023T18:12:00	VOC	0.56	140
VL040805.D	0.226	10252023T18:12:00	VOC	0.27	66
VL040805.D	0	10252023T18:12:00	VOC	0.14	35
VL040805.D	0.415	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.389	10252023T18:12:00	VOC	0.35	89
VL040805.D	0.439	10252023T18:12:00	VOC	0.35	87
VL040805.D	0.381	10252023T18:12:00	VOC	0.34	85
VL040805.D	0.467	10252023T18:12:00	VOC	0.46	110
VL040805.D	0.449	10252023T18:12:00	VOC	0.37	93
VL040805.D	0.35	10252023T18:12:00	VOC	0.27	67
VL040805.D	0.291	10252023T18:12:00	VOC	0.28	70
VL040805.D	0.37	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.427	10252023T18:12:00	VOC	0.37	92
VL040787.D	0.04	10232023T20:32:00	VOC	0.032	110
VL040805.D	0.416	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.38	10252023T18:12:00	VOC	0.31	78
VL040805.D	0.435	10252023T18:12:00	VOC	0.43	110
VL040787.D	0.05	10232023T20:32:00	VOC	0.033	110
VL040805.D	0.506	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.341	10252023T18:12:00	VOC	0.43	110
VL040787.D	0.04	10232023T20:32:00	VOC	0.033	110
			VOC	0.544	136

A
B
C
D
E
F
G
H



SAMPLE DATA

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	1	SDG No.:	Q1532
Lab Sample ID:	Q1532-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042121.D	1		03/10/25 12:59	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.51	2.52		1.04	2.47	ug/m3
74-87-3	Chloromethane	0.53	1.09		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.030	0.080	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.50	1.94	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.50	1.32	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.50	1.47	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.26	1.46	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	3.83	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.50	3.49	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.50	1.52	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.50	2.05	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.50	1.98	U	0.56	1.98	ug/m3
67-64-1	Acetone	7.20	17.1		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.50	1.56	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.50	1.80	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	1.70	5.91		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.50	1.98	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.50	2.02	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.50	1.72	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.65	1.92		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.070	0.44		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.50	1.98	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.50	2.44	U	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.030	0.16	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.50	2.34	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.13	0.42	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.50	2.02	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.030	0.16	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.50	2.31	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.50	3.35	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.50	2.05	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.40	1.51	J	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.50	2.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.50	2.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.50	2.73	U	0.38	2.73	ug/m3

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	1	SDG No.:	Q1532
Lab Sample ID:	Q1532-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042121.D	1		03/10/25 12:59	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	4.26	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.10	0.77	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.12	0.81		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.50	2.30	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.50	2.17	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.34	1.48	J	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.50	2.17	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.50	2.13	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.50	5.17	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.030	0.21	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.50	2.59	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.50	2.46	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.50	2.46	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.50	3.01	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.50	3.01	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.50	3.01	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.50	3.71	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.50	5.33	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.50	1.11	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.10	0.52	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.50	2.46	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.50	1.76	U	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.50	1.57	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.50	1.80	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.50	2.05	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.2			65 - 135	102%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	162000		2.8			
540-36-3	1,4-Difluorobenzene	375000		3.978			
3114-55-4	Chlorobenzene-d5	322000		8.901			

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	1	SDG No.:	Q1532
Lab Sample ID:	Q1532-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042121.D	1		03/10/25 12:59	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	2	SDG No.:	Q1532
Lab Sample ID:	Q1532-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042123.D	1		03/10/25 14:06	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.51	2.52		1.04	2.47	ug/m3
74-87-3	Chloromethane	1.00	2.07		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.030	0.080	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.50	1.94	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.50	1.32	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.50	1.47	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.24	1.35	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	3.83	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.50	3.49	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.50	1.52	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.50	2.05	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.50	1.98	U	0.56	1.98	ug/m3
67-64-1	Acetone	7.70	18.3		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.50	1.56	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.50	1.80	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	1.20	4.17		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.50	1.98	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.50	2.02	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.50	1.72	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.98	2.89		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.080	0.50		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.50	1.98	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.10	0.49	J	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.030	0.16	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.50	2.34	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.14	0.45	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.50	2.02	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.030	0.16	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.50	2.31	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.50	3.35	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.50	2.05	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.48	1.81	J	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.50	2.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.50	2.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.50	2.73	U	0.38	2.73	ug/m3

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	2	SDG No.:	Q1532
Lab Sample ID:	Q1532-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042123.D	1		03/10/25 14:06	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	4.26	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.10	0.77	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.040	0.27		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.50	2.30	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.50	2.17	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.31	1.35	J	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.50	2.17	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.50	2.13	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.50	5.17	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.030	0.21	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.50	2.59	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.50	2.46	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.50	2.46	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.50	3.01	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.50	3.01	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.50	3.01	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.50	3.71	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.50	5.33	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.50	1.11	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.10	0.52	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.50	2.46	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.50	1.76	U	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.50	1.57	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.50	1.80	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.50	2.05	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.1			65 - 135	101%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	167000		2.797			
540-36-3	1,4-Difluorobenzene	387000		3.978			
3114-55-4	Chlorobenzene-d5	330000		8.898			

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	2	SDG No.:	Q1532
Lab Sample ID:	Q1532-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042123.D	1		03/10/25 14:06	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	3	SDG No.:	Q1532
Lab Sample ID:	Q1532-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042125.D	1		03/10/25 15:10	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.51	2.52		1.04	2.47	ug/m3
74-87-3	Chloromethane	0.58	1.20		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.030	0.080	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.50	1.94	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.50	1.32	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.50	1.47	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.25	1.40	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	3.83	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.50	3.49	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.50	1.52	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.50	2.05	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.50	1.98	U	0.56	1.98	ug/m3
67-64-1	Acetone	6.50	15.4		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.50	1.56	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.50	1.80	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	1.50	5.21		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.50	1.98	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.50	2.02	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.50	1.72	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.69	2.04		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.070	0.44		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.50	1.98	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.50	2.44	U	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.030	0.16	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.50	2.34	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.14	0.45	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.50	2.02	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.030	0.16	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.50	2.31	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.50	3.35	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.50	2.05	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.38	1.43	J	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.50	2.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.50	2.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.50	2.73	U	0.38	2.73	ug/m3

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	3	SDG No.:	Q1532
Lab Sample ID:	Q1532-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042125.D	1		03/10/25 15:10	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	4.26	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.10	0.77	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.070	0.47		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.50	2.30	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.50	2.17	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.27	1.17	J	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.50	2.17	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.50	2.13	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.50	5.17	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.030	0.21	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.50	2.59	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.50	2.46	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.50	2.46	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.50	3.01	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.50	3.01	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.50	3.01	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.50	3.71	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.50	5.33	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.50	1.11	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.10	0.52	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.50	2.46	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.50	1.76	U	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.50	1.57	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.50	1.80	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.50	2.05	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.3			65 - 135	103%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	165000		2.8			
540-36-3	1,4-Difluorobenzene	383000		3.975			
3114-55-4	Chlorobenzene-d5	326000		8.898			

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	3	SDG No.:	Q1532
Lab Sample ID:	Q1532-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042125.D	1		03/10/25 15:10	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	4	SDG No.:	Q1532
Lab Sample ID:	Q1532-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042127.D	1		03/10/25 16:15	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.51	2.52		1.04	2.47	ug/m3
74-87-3	Chloromethane	0.57	1.18		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.030	0.080	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.50	1.94	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.50	1.32	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.50	1.47	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.26	1.46	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	3.83	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.50	3.49	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.50	1.52	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.50	2.05	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.50	1.98	U	0.56	1.98	ug/m3
67-64-1	Acetone	5.80	13.8		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.50	1.56	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.50	1.80	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	3.40	11.8		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.50	1.98	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.50	2.02	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.50	1.72	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.55	1.62		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.070	0.44		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.50	1.98	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.50	2.44	U	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.030	0.16	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.50	2.34	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.15	0.48	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.50	2.02	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.030	0.16	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.50	2.31	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.50	3.35	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.50	2.05	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.44	1.66	J	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.50	2.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.50	2.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.50	2.73	U	0.38	2.73	ug/m3

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	4	SDG No.:	Q1532
Lab Sample ID:	Q1532-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042127.D	1		03/10/25 16:15	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	4.26	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.10	0.77	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.060	0.41		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.50	2.30	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.50	2.17	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.31	1.35	J	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.50	2.17	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.50	2.13	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.50	5.17	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.030	0.21	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.50	2.59	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.50	2.46	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.50	2.46	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.50	3.01	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.50	3.01	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.50	3.01	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.50	3.71	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.50	5.33	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.50	1.11	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.10	0.52	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.50	2.46	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.43	1.52	J	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.50	1.57	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.50	1.80	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.50	2.05	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.3			65 - 135	103%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	163000		2.8			
540-36-3	1,4-Difluorobenzene	380000		3.978			
3114-55-4	Chlorobenzene-d5	324000		8.901			

Report of Analysis

Client:	Environmental Management Services	Date Collected:	03/06/25
Project:	36-12 37th St. Queens, NY	Date Received:	03/07/25
Client Sample ID:	4	SDG No.:	Q1532
Lab Sample ID:	Q1532-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042127.D	1		03/10/25 16:15	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements



QC SUMMARY

Surrogate Summary

SDG No.: Q1532

Client: Environmental Management Services

Analytical Method: SWTO-15

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1532-01	1	1-Bromo-4-Fluorobenzene	10	10.2	102	65	135
Q1532-01DUP	1DUP	1-Bromo-4-Fluorobenzene	10	10.2	101	65	135
Q1532-02	2	1-Bromo-4-Fluorobenzene	10	10.1	101	65	135
Q1532-03	3	1-Bromo-4-Fluorobenzene	10	10.3	103	65	135
Q1532-04	4	1-Bromo-4-Fluorobenzene	10	10.3	103	65	135
VL0310ABL01	VL0310ABL01	1-Bromo-4-Fluorobenzene	10	9.90	99	65	135
VL0310ABS01	VL0310ABS01	1-Bromo-4-Fluorobenzene	10	10.6	106	65	135

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1532

Client: Environmental Management Services

Analytical Method: SWTO-15

Datafile : VL042117.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0310ABS01	Dichlorodifluoromethane	10	10.2	ppbv	102			70	130	
	Chloromethane	10	10.1	ppbv	101			70	130	
	Vinyl Chloride	10	10.4	ppbv	104			70	130	
	Bromomethane	10	10.4	ppbv	104			70	130	
	Chloroethane	10	10.4	ppbv	104			70	130	
	Tetrahydrofuran	10	8.70	ppbv	87			70	130	
	Trichlorofluoromethane	10	10.7	ppbv	107			70	130	
	1,1,2-Trichlorotrifluoroethane	10	9.60	ppbv	96			70	130	
	Dichlorotetrafluoroethane	10	10.2	ppbv	102			70	130	
	tert-Butyl Alcohol	10	8.50	ppbv	85			70	130	
	Heptane	10	8.10	ppbv	81			70	130	
	1,1-Dichloroethene	10	9.40	ppbv	94			70	130	
	Acetone	10	9.80	ppbv	98			70	130	
	Carbon disulfide	10	9.70	ppbv	97			70	130	
	Methyl tert-butyl Ether	10	8.50	ppbv	85			70	130	
	Methylene Chloride	10	9.90	ppbv	99			70	130	
	trans-1,2-Dichloroethene	10	9.90	ppbv	99			70	130	
	1,1-Dichloroethane	10	10.1	ppbv	101			70	130	
	Cyclohexane	10	7.80	ppbv	78			70	130	
	2-Butanone	10	9.10	ppbv	91			70	130	
	Carbon Tetrachloride	10	9.50	ppbv	95			70	130	
	cis-1,2-Dichloroethene	10	7.90	ppbv	79			70	130	
	Chloroform	10	8.50	ppbv	85			70	130	
	1,1,1-Trichloroethane	10	7.90	ppbv	79			70	130	
	2,2,4-Trimethylpentane	10	9.10	ppbv	91			70	130	
	Benzene	10	8.90	ppbv	89			70	130	
	1,2-Dichloroethane	10	8.70	ppbv	87			70	130	
	Trichloroethene	10	8.40	ppbv	84			70	130	
	1,2-Dichloropropane	10	8.90	ppbv	89			70	130	
	Bromodichloromethane	10	9.10	ppbv	91			70	130	
	4-Methyl-2-Pentanone	10	9.60	ppbv	96			70	130	
	Toluene	10	8.50	ppbv	85			70	130	
	t-1,3-Dichloropropene	10	7.60	ppbv	76			70	130	
	cis-1,3-Dichloropropene	10	8.10	ppbv	81			70	130	
	1,1,2-Trichloroethane	10	8.90	ppbv	89			70	130	
	Dibromochloromethane	10	9.00	ppbv	90			70	130	
	1,2-Dibromoethane	10	8.70	ppbv	87			70	130	
	Tetrachloroethene	10	8.40	ppbv	84			70	130	
	Chlorobenzene	10	9.30	ppbv	93			70	130	
	Ethyl Benzene	10	8.60	ppbv	86			70	130	
	m/p-Xylene	20	18.0	ppbv	90			70	130	
	o-Xylene	10	9.10	ppbv	91			70	130	
	Styrene	10	8.50	ppbv	85			70	130	
	Bromoform	10	9.20	ppbv	92			70	130	
	1,1,2,2-Tetrachloroethane	10	9.50	ppbv	95			70	130	
	2-Chlorotoluene	10	8.60	ppbv	86			70	130	
	1,3,5-Trimethylbenzene	10	8.80	ppbv	88			70	130	
	1,2,4-Trimethylbenzene	10	8.70	ppbv	87			70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1532

Client: Environmental Management Services

Analytical Method: SWTO-15

Datafile : VL042117.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0310ABS01	1,3-Dichlorobenzene	10	9.00	ppbv	90			70	130	
	1,4-Dichlorobenzene	10	9.00	ppbv	90			70	130	
	1,2-Dichlorobenzene	10	9.30	ppbv	93			70	130	
	1,2,4-Trichlorobenzene	10	8.90	ppbv	89			70	130	
	Hexachloro-1,3-butadiene	10	8.20	ppbv	82			70	130	
	Naphthalene	10	11.2	ppbv	112			70	130	
	1,3-Butadiene	10	10.1	ppbv	101			70	130	
	4-Ethyltoluene	10	8.90	ppbv	89			70	130	
	Hexane	10	8.50	ppbv	85			70	130	
	Allyl Chloride	10	9.60	ppbv	96			70	130	
	1,4-Dioxane	10	8.30	ppbv	83			70	130	
	Methyl methacrylate	10	8.40	ppbv	84			70	130	

Duplicate Sample Summary

Lab Sample Id :	Q1532-01DUP	Q1532-01
Client Id :	1DUP	1
DF :	1	1
Datafile :	VL042122.D	VL042121.D
Anal Date & Time :	03/10/2025 13:32	03/10/2025 12:59

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0.05	0	200 *
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	0.68	0.65	4.5
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	7	7.2	2.8
Allyl Chloride	0	0	0
Benzene	0.13	0.13	0
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.08	0.07	13.3

Duplicate Sample Summary

Lab Sample Id :	Q1532-01DUP	Q1532-01
Client Id :	1DUP	1
DF :	1	1
Datafile :	VL042122.D	VL042121.D
Anal Date & Time :	03/10/2025 13:32	03/10/2025 12:59

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.55	0.53	3.7
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.5	0.51	2
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0	0	0
Heptane	0	0	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	0	0	0
m/p-Xylene	0.34	0.34	0
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	1.5	1.7	12.5
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0.14	0.12	15.4
Tetrahydrofuran	0	0	0
Toluene	0.39	0.4	2.5
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.27	0.26	3.8
Vinyl Chloride	0	0	0

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VL0310ABL01

Lab Name: CHEMTECH

Contract: ENVI46

Lab Code: CHEM Case No.: Q1532

SAS No.: Q1532 SDG NO.: Q1532

Lab File ID: VL042116.D

Lab Sample ID: VL0310ABL01

Date Analyzed: 03/10/2025

Time Analyzed: 09:54

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_L

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VL0310ABS01	VL0310ABS01	VL042117.D	03/10/2025
1	Q1532-01	VL042121.D	03/10/2025
1DUP	Q1532-01DUP	VL042122.D	03/10/2025
2	Q1532-02	VL042123.D	03/10/2025
3	Q1532-03	VL042125.D	03/10/2025
4	Q1532-04	VL042127.D	03/10/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: ENVI46
Lab Code: CHEM Case No.: Q1532 SAS No.: Q1532 SDG NO.: Q1532
Lab File ID: VL042040.D BFB Injection Date: 02/25/2025
Instrument ID: MSVOA_L BFB Injection Time: 07:56
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	22.6
75	30.0 - 66.0% of mass 95	55.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.7 (1.1) 1
174	50.0 - 120.0% of mass 95	63.3
175	4.0 - 9.0% of mass 174	4.9 (7.8) 1
176	93.0 - 101.0% of mass 174	60.9 (96.1) 1
177	5.0 - 9.0% of mass 176	3.4 (5.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICCC010	VSTDICCC010	VL042041.D	02/25/2025	08:56
VSTDICCC002	VSTDICCC002	VL042042.D	02/25/2025	09:29
VSTDICCC001	VSTDICCC001	VL042043.D	02/25/2025	10:00
VSTDICCC0.5	VSTDICCC0.5	VL042044.D	02/25/2025	10:31
VSTDICCC0.1	VSTDICCC0.1	VL042045.D	02/25/2025	11:03
VSTDICCC0.03	VSTDICCC0.03	VL042046.D	02/25/2025	11:34
VSTDICCC015	VSTDICCC015	VL042047.D	02/25/2025	12:07

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: ENVI46
Lab Code: CHEM Case No.: Q1532 SAS No.: Q1532 SDG NO.: Q1532
Lab File ID: VL042114.D BFB Injection Date: 03/10/2025
Instrument ID: MSVOA_L BFB Injection Time: 08:22
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	22.6
75	30.0 - 66.0% of mass 95	53.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.7 (1.1) 1
174	50.0 - 120.0% of mass 95	63.1
175	4.0 - 9.0% of mass 174	4.8 (7.6) 1
176	93.0 - 101.0% of mass 174	60.4 (95.6) 1
177	5.0 - 9.0% of mass 176	3.9 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC010	VSTDCCC010	VL042115.D	03/10/2025	09:08
VL0310ABL01	VL0310ABL01	VL042116.D	03/10/2025	09:54
VL0310ABS01	VL0310ABS01	VL042117.D	03/10/2025	10:52
1	Q1532-01	VL042121.D	03/10/2025	12:59
1DUP	Q1532-01DUP	VL042122.D	03/10/2025	13:32
2	Q1532-02	VL042123.D	03/10/2025	14:06
3	Q1532-03	VL042125.D	03/10/2025	15:10
4	Q1532-04	VL042127.D	03/10/2025	16:15

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENVI46
Lab Code: CHEM Case No.: Q1532 SAS No.: Q1532 SDG NO.: Q1532
Lab File ID: VL042115.D Date Analyzed: 03/10/2025
Instrument ID: MSVOA_L Time Analyzed: 09:08
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	165085	2.80	392891	3.98	342991	8.90
UPPER LIMIT	231119	3.13	550047	4.31	480187	9.23
LOWER LIMIT	99051	2.47	235735	3.65	205795	8.57
EPA SAMPLE NO.						
1	161577	2.80	375440	3.98	321830	8.90
1DUP	163759	2.80	380274	3.98	322839	8.90
2	166689	2.80	387215	3.98	330250	8.90
3	165184	2.80	383076	3.98	325600	8.90
4	162703	2.80	379717	3.98	323671	8.90
VL0310ABL01	162394	2.80	386417	3.97	325291	8.90
VL0310ABS01	168824	2.80	400351	3.98	350466	8.90

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Environmental Management Services	Date Collected:	
Project:	36-12 37th St. Queens, NY	Date Received:	
Client Sample ID:	VL0310ABL01	SDG No.:	Q1532
Lab Sample ID:	VL0310ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042116.D	1		03/10/25 09:54	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	2.47	U	1.04	2.47	ug/m3
74-87-3	Chloromethane	0.50	1.03	U	0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.030	0.080	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.50	1.94	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.50	1.32	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.50	1.47	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.50	2.81	U	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	3.83	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.50	3.49	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.50	1.52	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.50	2.05	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.50	1.98	U	0.56	1.98	ug/m3
67-64-1	Acetone	0.50	1.19	U	0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.50	1.56	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.50	1.80	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	0.50	1.74	U	0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.50	1.98	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.50	2.02	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.50	1.72	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.50	1.47	U	0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.030	0.19	U	0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.50	1.98	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.50	2.44	U	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.030	0.16	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.50	2.34	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.50	1.60	U	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.50	2.02	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.030	0.16	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.50	2.31	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.50	3.35	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.50	2.05	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.50	1.88	U	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.50	2.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.50	2.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.50	2.73	U	0.38	2.73	ug/m3

Report of Analysis

Client:	Environmental Management Services	Date Collected:	
Project:	36-12 37th St. Queens, NY	Date Received:	
Client Sample ID:	VL0310ABL01	SDG No.:	Q1532
Lab Sample ID:	VL0310ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042116.D	1		03/10/25 09:54	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	4.26	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.10	0.77	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.030	0.20	U	0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.50	2.30	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.50	2.17	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	1.00	4.34	U	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.50	2.17	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.50	2.13	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.50	5.17	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.030	0.21	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.50	2.59	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.50	2.46	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.50	2.46	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.50	3.01	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.50	3.01	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.50	3.01	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.50	3.71	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.50	5.33	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.50	1.11	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.10	0.52	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.50	2.46	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.50	1.76	U	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.50	1.57	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.50	1.80	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.50	2.05	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.90			65 - 135	99%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	162000		2.797			
540-36-3	1,4-Difluorobenzene	386000		3.972			
3114-55-4	Chlorobenzene-d5	325000		8.895			

Report of Analysis

Client:	Environmental Management Services	Date Collected:	
Project:	36-12 37th St. Queens, NY	Date Received:	
Client Sample ID:	VL0310ABL01	SDG No.:	Q1532
Lab Sample ID:	VL0310ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042116.D	1		03/10/25 09:54	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	Environmental Management Services	Date Collected:	
Project:	36-12 37th St. Queens, NY	Date Received:	
Client Sample ID:	VL0310ABS01	SDG No.:	Q1532
Lab Sample ID:	VL0310ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042117.D	1		03/10/25 10:52	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	10.2	50.4		1.04	2.47	ug/m3
74-87-3	Chloromethane	10.1	20.9		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	10.4	26.6		0.030	0.080	ug/m3
74-83-9	Bromomethane	10.4	40.4		0.54	1.94	ug/m3
75-00-3	Chloroethane	10.4	27.4		0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	8.70	25.7		0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	10.7	60.1		0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	9.60	73.6		1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	10.2	71.3		0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	8.50	25.8		0.58	1.52	ug/m3
142-82-5	Heptane	8.10	33.2		0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	9.40	37.3		0.56	1.98	ug/m3
67-64-1	Acetone	9.80	23.3		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	9.70	30.2		0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	8.50	30.6		0.43	1.80	ug/m3
75-09-2	Methylene Chloride	9.90	34.4		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	9.90	39.3		0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	10.1	40.9		0.53	2.02	ug/m3
110-82-7	Cyclohexane	7.80	26.9		0.76	1.72	ug/m3
78-93-3	2-Butanone	9.10	26.8		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	9.50	59.8		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	7.90	31.3		0.36	1.98	ug/m3
67-66-3	Chloroform	8.50	41.5		0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	7.90	43.1		0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	9.10	42.5		0.47	2.34	ug/m3
71-43-2	Benzene	8.90	28.4		0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	8.70	35.2		0.36	2.02	ug/m3
79-01-6	Trichloroethene	8.40	45.1		0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	8.90	41.1		0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	9.10	61.0		0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	9.60	39.3		0.37	2.05	ug/m3
108-88-3	Toluene	8.50	32.0		0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	7.60	34.5		0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	8.10	36.8		0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	8.90	48.6		0.38	2.73	ug/m3

Report of Analysis

Client:	Environmental Management Services	Date Collected:	
Project:	36-12 37th St. Queens, NY	Date Received:	
Client Sample ID:	VL0310ABS01	SDG No.:	Q1532
Lab Sample ID:	VL0310ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042117.D	1		03/10/25 10:52	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	9.00	76.7		0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	8.70	66.9		0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	8.40	57.0		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	9.30	42.8		0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	8.60	37.4		0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	18.0	78.2		0.91	4.34	ug/m3
95-47-6	o-Xylene	9.10	39.5		0.52	2.17	ug/m3
100-42-5	Styrene	8.50	36.2		0.47	2.13	ug/m3
75-25-2	Bromoform	9.20	95.1		0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	9.50	65.2		0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	8.60	44.5		0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	8.80	43.3		0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	8.70	42.8		0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	9.00	54.1		0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	9.00	54.1		0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	9.30	55.9		0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	8.90	66.1		0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	8.20	87.5		0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	10.1	22.3		0.29	1.11	ug/m3
91-20-3	Naphthalene	11.2	58.7		0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	8.90	43.8		0.59	2.46	ug/m3
110-54-3	Hexane	8.50	30.0		0.39	1.76	ug/m3
107-05-1	Allyl Chloride	9.60	30.1		0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	8.30	29.9		0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	8.40	34.4		0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.6			65 - 135	106%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	169000		2.797			
540-36-3	1,4-Difluorobenzene	400000		3.975			
3114-55-4	Chlorobenzene-d5	350000		8.898			

Report of Analysis

Client:	Environmental Management Services	Date Collected:	
Project:	36-12 37th St. Queens, NY	Date Received:	
Client Sample ID:	VL0310ABS01	SDG No.:	Q1532
Lab Sample ID:	VL0310ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042117.D	1		03/10/25 10:52	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements



CALIBRATION SUMMARY

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL022525AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Wed Feb 26 01:08:36 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL042046.D 0.1 =VL042045.D 0.5 =VL042044.D 1 =VL042043.D 2 =VL042042.D 10 =VL042041.D 15 =VL042047.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.109	1.084	1.091	0.943	0.973	1.040	7.31
3) Chlorodifluoro...			1.493	1.566	1.535	1.311	1.362	1.453	7.67
4) Chloromethane			0.438	0.398	0.399	0.344	0.344	0.385	10.53
5) T Vinyl Chloride	0.447	0.336	0.376	0.346	0.361	0.337	0.343	0.364	10.89
6) T Bromomethane			0.170	0.184	0.186	0.160	0.165	0.173	6.63
7) Chloroethane			0.151	0.158	0.156	0.135	0.140	0.148	6.83
8) T Dichlorotetra...			0.892	0.910	0.901	0.779	0.791	0.855	7.47
9) T Propene			0.614	0.630	0.636	0.522	0.558	0.592	8.47
10) T Heptane			1.971	1.946	1.960	1.662	1.679	1.844	8.59
11) T Trichlorofluor...			1.056	1.122	1.119	0.944	0.968	1.042	7.98
12) T 1,1,2-Trichlor...			0.919	0.902	0.865	0.697	0.773	0.831	11.28
13) Ethanol			0.053	0.057	0.040	0.020	0.018	0.038#	48.50
14) T Bromoethene			0.325	0.304	0.299	0.264	0.275	0.293	8.41
15) T Acetone			1.084	0.960	0.933	0.714	0.723	0.883	18.17
16) T 1,3-Butadiene			0.476	0.435	0.418	0.358	0.367	0.411	11.96
17) tert-Butyl alc...			1.184	1.185	1.201	0.950	1.004	1.105	10.71
18) T 1,1-Dichloroet...			0.398	0.437	0.386	0.320	0.338	0.376	12.57
19) T Isopropyl Alcohol			0.561	0.571	0.557	0.419	0.429	0.507	15.06
20) T Methylene Chlo...			0.429	0.374	0.329	0.261	0.278	0.334	20.72
21) T Allyl Chloride			0.707	0.717	0.639	0.538	0.590	0.638	11.94
22) T trans-1,2-Dich...			0.481	0.506	0.450	0.358	0.395	0.438	13.90
23) T Vinyl Acetate			0.689	0.614	0.695	0.520	0.586	0.621	11.82
24) T 1,1-Dichloroet...			0.927	0.856	0.863	0.669	0.758	0.815	12.42
25) T Ethyl Acetate			3.833	3.674	3.776	3.174	3.217	3.535	8.91
26) T Hexane			1.596	1.462	1.519	1.282	1.287	1.429	9.83
27) T Carbon Disulfide			1.042	0.972	1.032	0.856	0.942	0.969	7.78
28) T Methyl tert-Bu...			0.650	0.844	0.838	0.522	0.705	0.711	19.02
29) T Chloroform			2.165	2.044	1.971	1.722	1.716	1.923	10.36
30) T Cyclohexane			1.154	1.239	1.255	1.076	1.076	1.160	7.39
31) T cis-1,2-Dichlo...			1.421	1.401	1.450	1.237	1.255	1.353	7.33
32) T 1,1,1-Trichlor...	1.921	1.875	1.883	1.916	1.944	1.771	1.751	1.866	4.05

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.833	0.799	0.812	0.692	0.759	0.779	7.15
35) T Carbon Tetrach...	0.664	0.705	0.675	0.723	0.725	0.655	0.681	0.690	4.09
36) T Benzene			1.079	1.084	1.104	0.969	0.999	1.047	5.66
37) T 1,2-Dichloroet...			0.624	0.604	0.598	0.520	0.556	0.580	7.20
38) T Trichloroethene	0.447	0.462	0.433	0.447	0.447	0.378	0.379	0.427	8.06
39) T 1,2-Dichloropr...			0.421	0.415	0.414	0.357	0.379	0.397	7.02

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\										
Method File : VL022525AIR.M										
40)	T	1,4-Dioxane		0.192	0.196	0.181	0.164	0.173	0.181	7.31
41)	T	Tetrahydrofuran		0.469	0.492	0.473	0.418	0.450	0.460	6.04
42)	T	Bromodichlorom...		0.806	0.791	0.810	0.739	0.765	0.782	3.81
43)		Methyl Methacr...		0.400	0.418	0.426	0.392	0.409	0.409	3.35
44)	T	2,2,4-Trimethy...		1.928	1.935	1.955	1.670	1.750	1.848	6.99
45)	T	t-1,3-Dichloro...		0.414	0.392	0.418	0.393	0.414	0.406	3.12
46)	T	cis-1,3-Dichlo...		0.514	0.523	0.548	0.516	0.526	0.526	2.59
47)	T	1,1,2-Trichlor...		0.426	0.429	0.426	0.361	0.375	0.403	8.12
48)	T	Dibromochlorom...		0.618	0.667	0.667	0.594	0.613	0.632	5.25
49)	T	Bromoform		0.553	0.551	0.588	0.525	0.526	0.549	4.68
50)	T	4-Methyl-2-Pen...		0.996	0.974	0.975	1.014	1.079	1.008	4.27
51)	T	2-Hexanone		0.577	0.570	0.608	0.834	0.879	0.693	21.69
52)	T	Tetrachloroethene	0.322 0.374	0.380	0.383	0.385	0.327	0.337	0.358	7.88
53)	T	Toluene		1.150	1.210	1.284	1.088	1.118	1.170	6.67
54)	T	1,2-Dibromoethane	0.563	0.556	0.553	0.586	0.500	0.514	0.545	5.87
-----ISTD-----										
55)	I	Chlorobenzene-d5								
56)		1,1,1,2-Tetrac...		0.607	0.596	0.595	0.505	0.524	0.565	8.37
57)	T	Chlorobenzene		1.102	1.113	1.121	0.910	0.938	1.037	10.02
58)	T	Ethyl Benzene		1.886	1.906	1.974	1.694	1.749	1.842	6.31
59)	T	m/p-Xylene		1.488	1.550	1.580	1.330	1.380	1.466	7.35
60)	T	o-Xylene		1.572	1.550	1.577	1.317	1.351	1.473	8.68
61)	T	Styrene		0.659	0.719	0.738	0.658	0.690	0.693	5.18
62)		Isopropylbenzene		2.299	2.339	2.399	1.969	2.026	2.207	8.84
63)	T	1,1,2,2-Tetrac...	0.999 0.975	1.026	1.041	1.031	0.848	0.872	0.970	8.13
64)		n-propylbenzene		0.596	0.607	0.632	0.524	0.535	0.579	8.11
65)		tert-Butylbenzene		2.080	2.251	2.222	1.733	1.753	2.008	12.47
66)	T	Benzyl Chloride		0.141	0.159	0.163	0.153	0.150	0.153	5.40
67)		sec-Butylbenzene		2.973	3.229	3.182	2.482	2.494	2.872	12.65
68)	S	1-Bromo-4-Fluo...	0.773 0.760	0.768	0.777	0.765	0.764	0.779	0.769	0.92
69)		p-Isopropyltol...		2.240	2.533	2.575	2.021	2.056	2.285	11.38
70)		n-Butylbenzene		2.332	2.567	2.572	2.070	2.120	2.332	10.20
71)		2-Chlorotoluene		1.758	1.829	1.853	1.514	1.551	1.701	9.30
72)	T	4-Ethyltoluene		1.775	1.944	1.985	1.621	1.670	1.799	8.99
73)	T	1,3,5-Trimethy...		1.592	1.732	1.696	1.381	1.400	1.560	10.47
74)	T	1,2,4-Trimethy...		1.855	1.982	1.990	1.471	1.488	1.757	14.76
75)	T	1,3-Dichlorobe...		1.112	1.110	1.083	0.869	0.882	1.011	12.30
76)	T	1,4-Dichlorobe...		1.120	1.061	1.092	0.861	0.886	1.004	12.07
77)	T	1,2-Dichlorobe...		1.059	1.097	1.063	0.836	0.852	0.981	12.88
78)	T	Hexachloro-1,3...		0.840	0.832	0.844	0.637	0.666	0.764	13.53
79)	T	Naphthalene	0.841	0.978	1.237	1.352	1.371	1.450	1.205	20.15
80)	T	Naphthalene,2-...		0.090	0.142	0.240	0.562	0.615	0.330	73.74
81)	T	1,2,4-Trichlor...		0.589	0.687	0.765	0.644	0.673	0.671	9.58

(#) = Out of Range

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENVI46

Lab Code: CHEM Case No.: Q1532 SAS No.: Q1532 SDG No.: Q1532

Instrument ID: MSVOA_L Calibration Date/Time: 03/10/2025 09:08

Lab File ID: VL042115.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:56 12:07

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	1.040	1.070		2.88	30
Chloromethane	0.385	0.403		4.68	30
Vinyl Chloride	0.364	0.383		5.22	30
Bromomethane	0.173	0.189		9.25	30
Chloroethane	0.148	0.155		4.73	30
Tetrahydrofuran	0.460	0.409		-11.09	30
Trichlorofluoromethane	1.042	1.130		8.44	30
1,1,2-Trichlorotrifluoroethane	0.831	0.819		-1.44	30
Dichlorotetrafluoroethane	0.855	0.904		5.73	30
tert-Butyl alcohol	1.105	0.968		-12.4	30
Heptane	1.844	1.519		-17.63	30
1,1-Dichloroethene	0.376	0.364		-3.19	30
Acetone	0.883	0.868		-1.7	30
Carbon Disulfide	0.969	0.976		0.72	30
Methyl tert-Butyl Ether	0.711	0.754		6.05	30
Methylene Chloride	0.334	0.333		-0.3	30
trans-1,2-Dichloroethene	0.438	0.436		-0.46	30
1,1-Dichloroethane	0.815	0.843		3.44	30
Cyclohexane	1.160	0.936		-19.31	30
2-Butanone	0.779	0.714		-8.34	30
Carbon Tetrachloride	0.690	0.671		-2.75	30
cis-1,2-Dichloroethene	1.353	1.052		-22.25	30
Chloroform	1.923	1.687		-12.27	30
1,1,1-Trichloroethane	1.866	1.485		-20.42	30
2,2,4-Trimethylpentane	1.848	1.713		-7.3	30
Benzene	1.047	0.951		-9.17	30
1,2-Dichloroethane	0.580	0.503		-13.28	30
Trichloroethene	0.427	0.360		-15.69	30
1,2-Dichloropropane	0.397	0.366		-7.81	30
Bromodichloromethane	0.782	0.718		-8.18	30
4-Methyl-2-Pentanone	1.008	0.978		-2.98	30
Toluene	1.170	0.989		-15.47	30
t-1,3-Dichloropropene	0.406	0.307		-24.38	30
cis-1,3-Dichloropropene	0.526	0.430		-18.25	30
1,1,2-Trichloroethane	0.403	0.369		-8.44	30
Dibromochloromethane	0.632	0.569		-9.97	30
1,2-Dibromoethane	0.545	0.483		-11.38	30
Tetrachloroethene	0.358	0.306		-14.52	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENVI46

Lab Code: CHEM Case No.: Q1532 SAS No.: Q1532 SDG No.: Q1532

Instrument ID: MSVOA_L Calibration Date/Time: 03/10/2025 09:08

Lab File ID: VL042115.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:56 12:07

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Chlorobenzene	1.037	0.962		-7.23	30
Ethyl Benzene	1.842	1.602		-13.03	30
m/p-Xylene	1.466	1.356		-7.5	30
o-Xylene	1.473	1.360		-7.67	30
Styrene	0.693	0.595		-14.14	30
Bromoform	0.549	0.516		-6.01	30
1,1,2,2-Tetrachloroethane	0.970	0.955		-1.55	30
2-Chlorotoluene	1.701	1.503		-11.64	30
1,3,5-Trimethylbenzene	1.560	1.401		-10.19	30
1,2,4-Trimethylbenzene	1.757	1.604		-8.71	30
1,3-Dichlorobenzene	1.011	0.949		-6.13	30
1,4-Dichlorobenzene	1.004	0.938		-6.57	30
1,2-Dichlorobenzene	0.981	0.940		-4.18	30
1,2,4-Trichlorobenzene	0.671	0.625		-6.86	30
Hexachloro-1,3-Butadiene	0.764	0.645		-15.58	30
1,3-Butadiene	0.411	0.422		2.68	30
Naphthalene	1.205	1.371		13.78	30
4-Ethyltoluene	1.799	1.651		-8.23	30
1-Bromo-4-Fluorobenzene	0.769	0.815		5.98	30
Hexane	1.429	1.213		-15.11	30
Allyl Chloride	0.638	0.625		-2.04	30
1,4-Dioxane	181.383	168.887		-6.89	30
Methyl Methacrylate	0.409	0.336		-17.85	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.