

DATA PACKAGE

VOLATILE ORGANICS

PROJECT NAME : 3555 VICTORY BLVD, STATEN ISLAND NY

GFE LLC

58 Nokomis Ave

Lake Hiawatha, NJ - 07034

Phone No: 646-542-3465

ORDER ID : Q1533

ATTENTION : Frank Galdun



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) VOCMS Group3- Case Narrative	4
3) Qualifier Page	6
4) QA Checklist	7
5) VOCMS Group3 Data	8
6) Shipping Document	16
6.1) CHAIN OF CUSTODY	17
6.2) ROC	22
6.3) Lab Certificate	23
6.4) Internal COC	24

1
2
3
4
5
6

Cover Page

Order ID : Q1533

Project ID : 3555 Victory Blvd, Staten Island NY

Client : GFE LLC

Lab Sample Number

Q1533-01
Q1533-02
Q1533-03
Q1533-04
Q1533-05

Client Sample Number

SV1
IA1
SV2
IA2
OA1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 3/12/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

GFE LLC

Project Name: 3555 Victory Blvd, Staten Island NY

Project # N/A

Chemtech Project # Q1533

Test Name: VOCMS Group3

A. Number of Samples and Date of Receipt:

5 Air samples were received on 03/07/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOCMS Group3. This data package contains results for VOCMS Group3.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax , Entech 7100A Preconcentrator. The analysis of VOCMS Group3 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

Due to potential high concentration of target analytes, Samples SV1 and SV2 were initially diluted.

E. Additional Comments:

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1533

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 03/12/2025

Hit Summary Sheet
SW-846

SDG No.: Q1533
Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	SV1							
Q1533-01	SV1	Air	Trichloroethene	1.83		0.38	0.64	ug/m3
Q1533-01	SV1	Air	Tetrachloroethene	136		0.41	0.81	ug/m3
			Total Voc :	137				
			Total Concentration:	137				
Client ID:	IA1							
Q1533-02	IA1	Air	Trichloroethene	0.64		0.11	0.16	ug/m3
Q1533-02	IA1	Air	Tetrachloroethene	38.0		0.14	0.20	ug/m3
			Total Voc :	38.6				
			Total Concentration:	38.6				
Client ID:	SV2							
Q1533-03	SV2	Air	Trichloroethene	1.72		0.38	0.64	ug/m3
Q1533-03	SV2	Air	Tetrachloroethene	121		0.41	0.81	ug/m3
			Total Voc :	122				
			Total Concentration:	122				
Client ID:	IA2							
Q1533-04	IA2	Air	Tetrachloroethene	7.46		0.14	0.20	ug/m3
			Total Voc :	7.46				
			Total Concentration:	7.46				
Client ID:	OA1							
Q1533-05	OA1	Air	Tetrachloroethene	0.20		0.14	0.20	ug/m3
			Total Voc :	0.20				
			Total Concentration:	0.20				



SAMPLE DATA

Report of Analysis

Client:	GFE LLC	Date Collected:	03/07/25
Project:	3555 Victory Blvd, Staten Island NY	Date Received:	03/07/25
Client Sample ID:	SV1	SDG No.:	Q1533
Lab Sample ID:	Q1533-01	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042135.D	4		03/10/25 20:35	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	0.060	0.15	U	0.15	0.31	ug/m3
75-35-4	1,1-Dichloroethene	0.56	2.22	U	2.22	7.93	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.36	1.43	U	1.43	7.93	ug/m3
71-55-6	1,1,1-Trichloroethane	0.040	0.22	U	0.22	0.65	ug/m3
79-01-6	Trichloroethene	0.34	1.83		0.38	0.64	ug/m3
127-18-4	Tetrachloroethene	20.0	136		0.41	0.81	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.5			65 - 135	105%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	156000			2.797		
540-36-3	1,4-Difluorobenzene	363000			3.975		
3114-55-4	Chlorobenzene-d5	361000			8.898		

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	03/07/25
Project:	3555 Victory Blvd, Staten Island NY	Date Received:	03/07/25
Client Sample ID:	IA1	SDG No.:	Q1533
Lab Sample ID:	Q1533-02	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042132.D	1		03/10/25 18:57	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
79-01-6	Trichloroethene	0.12	0.64		0.11	0.16	ug/m3
127-18-4	Tetrachloroethene	5.60	38.0		0.14	0.20	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.0			65 - 135	100%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	160000			2.8		
540-36-3	1,4-Difluorobenzene	376000			3.981		
3114-55-4	Chlorobenzene-d5	321000			8.901		

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	03/07/25
Project:	3555 Victory Blvd, Staten Island NY	Date Received:	03/07/25
Client Sample ID:	SV2	SDG No.:	Q1533
Lab Sample ID:	Q1533-03	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042137.D	4		03/10/25 21:38	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	0.060	0.15	U	0.15	0.31	ug/m3
75-35-4	1,1-Dichloroethene	0.56	2.22	U	2.22	7.93	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.36	1.43	U	1.43	7.93	ug/m3
71-55-6	1,1,1-Trichloroethane	0.040	0.22	U	0.22	0.65	ug/m3
79-01-6	Trichloroethene	0.32	1.72		0.38	0.64	ug/m3
127-18-4	Tetrachloroethene	17.8	121		0.41	0.81	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	11.1			65 - 135	111%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	158000			2.8		
540-36-3	1,4-Difluorobenzene	370000			3.978		
3114-55-4	Chlorobenzene-d5	340000			8.898		

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	03/07/25
Project:	3555 Victory Blvd, Staten Island NY	Date Received:	03/07/25
Client Sample ID:	IA2	SDG No.:	Q1533
Lab Sample ID:	Q1533-04	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042133.D	1		03/10/25 19:30	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
127-18-4	Tetrachloroethene	1.10	7.46		0.14	0.20	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.0			65 - 135	100%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	163000			2.8		
540-36-3	1,4-Difluorobenzene	382000			3.978		
3114-55-4	Chlorobenzene-d5	316000			8.901		

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	03/07/25
Project:	3555 Victory Blvd, Staten Island NY	Date Received:	03/07/25
Client Sample ID:	OA1	SDG No.:	Q1533
Lab Sample ID:	Q1533-05	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042134.D	1		03/10/25 20:03	VL031025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
127-18-4	Tetrachloroethene	0.030	0.20		0.14	0.20	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.1			65 - 135	101%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	160000			2.8		
540-36-3	1,4-Difluorobenzene	372000			3.978		
3114-55-4	Chlorobenzene-d5	309000			8.901		

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

LAB CHRONICLE

OrderID: Q1533	OrderDate: 3/10/2025 9:20:00 AM
Client: GFE LLC	Project: 3555 Victory Blvd, Staten Island NY
Contact: Frank Galdun	Location: F11,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1533-01	SV1	Air	VOCMS Group3	TO-15	03/07/25		03/10/25	03/07/25
Q1533-02	IA1	Air	VOCMS Group3	TO-15	03/07/25		03/10/25	03/07/25
Q1533-03	SV2	Air	VOCMS Group3	TO-15	03/07/25		03/10/25	03/07/25
Q1533-04	IA2	Air	VOCMS Group3	TO-15	03/07/25		03/10/25	03/07/25
Q1533-05	OA1	Air	VOCMS Group3	TO-15	03/07/25		03/10/25	03/07/25



SHIPPING DOCUMENTS

Client Contact Information				Bottle Order ID : B2503004				Courier : FGALDUN				1 of 3 COCs									
Client ID : GFEL01				Project ID : 97-34 63RD RD, Rego Park NY				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix							
Customer Name : GFE LLC Address : 58 Nokomis Ave				Project Manager : FRANK GALDUN				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified				Indoor/Ambient Air Soil Gas									
				Phone Number : 646-542-3465																	
				Fax Number : 973-334-1692																	
City : Lake Hiawatha				Site Details : 3555 VICTORY BLD.																	
State : NJ				Analysis Turnaround Time : 5 DAY																	
Zip Code : 07034				Standard : 10 Business days OR				Data Package Type : RESULTS ONLY													
Country :				Rush (Specify) : 5 Days				EDD Type : PDF													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID	Flow Controller Readout (ml/min)	Can Cert ID								
SU1	3/7/25	8:50	10:50	6	6	68	68	-30	-6.3	10707	10296	6 L	50	VL041966.D							
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) [Signature]											
		Ambient		Maximum		Minimum															
Start																					
Stop																					
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. REPORT ONLY: PCE, TCE, cis-1,2, DCE, 1,1-DCE, 1,1-TCM VINYL CHLORIDE Please follow the instructions on the back of this COC.											
		Ambient		Maximum		Minimum															
Start																					
Stop																					
Special Instructions/QC Requirements & Comments :																					
Suspected Contamination: High Medium Low PID Readings: 0.0																					
Sampling site (State):																					
Quick Connector required : NO																					
Canisters Shipped by: [Signature]				Date/Time: 3/5/25				Canisters Received by: [Signature]				Date/Time: 3.7.25 14:00				B2503004 - 4					
Samples Relinquished by: [Signature]				Date/Time: 3/11/25				Received by:				Date/Time:									
Relinquished by:				Date/Time:				Received by:				Date/Time:									

Client Contact Information		Bottle Order ID : B2503004	Courier : F GALDUN	2 of 3 COCs 6.1	
Client ID : GFEL01	Project ID : 97-24-63RD RD, Rego Park NY	Sampler Name(s) : FRANK GALDUN		Analysis	Matrix
Customer Name : GFE LLC	Project Manager : FRANK GALDUN	AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified			
Address : 58 Nokomis Ave	Phone Number : 646-542-3465				
City : Lake Hiawatha	Fax Number : 973-334-1692				
State : NJ	Site Details: 3555 VICTORY BLD. STATEN ISLAND, NY				
Zip Code : 07034	Standard : 10 business days OR	Analysis Turnaround Time : 5 DAY	Data Package Type : RESULTS ONLY		
Country :	Rush (Specify): 5 Days	EDD Type : PDF			

Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID	Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor Ambient Air	Soil Gas
IA1	3/7/25	8:56	10:46	30	6.5	68	60	-30	-7.1	10579	10321	6 L 50	VL041966.D	1	1	

Temperature (Fahrenheit)			
	Ambient	Maximum	Minimum
Start			
Stop			

GC/MS Analyst Signature (TO-15) Semy

Pressure (Inches of Hg)			
	Ambient	Maximum	Minimum
Start			
Stop			

** Submittal of this COC indicates approval of the analysis based on existing conditions.
REPORT ONLY: PCE, TCE, cis-1,2-DCE, 1,1-DCE, 1,1,1-TCA, VINYL CHLORIDE
 Please follow the instructions on the back of this COC.

Special Instructions/QC Requirements & Comments :

Suspected Contamination: High Medium **Low** PID Readings: **0.0**

Sampling site (State):

Quick Connector required : No		
Canisters Shipped by: Semy	Date/Time: 3/5/25	Canisters Received by: dr
Samples Relinquished by: FRANK GALDUN	Date/Time: 3/7/25	Received by:
Relinquished by:	Date/Time:	Received by:

B2503004 - 5

Client Contact Information				Bottle Order ID : B2503004				Courier : Results only				3 of 5 COCs 6.1				
Client ID : GFEL01 Project ID : 97-34-03RD RD, STATEN ISLAND, NY				Project Manager : FRANK GALDUN				Sampler Name(s) : PDF				Analysis		Matrix		
Customer Name : GFE LLC				Phone Number : 646-542-3465				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified								
Address : 58 Nokomis Ave				Fax Number : 973-334-1692												
City : Lake Hiawatha				Site Details: 3555 VICTORY BLVD STATEN ISLAND, NY												
State : NJ				Analysis Turnaround Time 5 DAY												
Zip Code : 07034				Standard : 10 business days OR				Data Package Type : Results only								
Country :				Rush (Specify): 5 Days				EDD Type : PDF								
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID	Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
SV2	3/7/25	8:51	10:51	30	5.5	68	68	-30	-5.1	10648	10410	6 L 50	VL041954.D	J		/
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) 						
		Ambient		Maximum		Minimum										
Start																
Stop																
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. REPORT ONLY: PCE, TCE, cis-1,2-DCE, 1,1-DCE, Please follow the instructions on the back of this COC.						
		Ambient		Maximum		Minimum										
Start																
Stop																
Special Instructions/QC Requirements & Comments :																
Suspected Contamination: High Medium Low PID Readings: 0.0																
Sampling site (State):																
Quick Connector required : NJO																
Canisters Shipped by: Sam				Date/Time: 03/11/25				Canisters Received by: OP				Date/Time: 3-7-25 14:00				B2503004 - 6
Samples Relinquished by: Frank				Date/Time: 3/7/25				Received by:				Date/Time:				
Relinquished by:				Date/Time:				Received by:				Date/Time:				

Client Contact Information		Bottle Order ID : B2503004				Courier: F GALDUN				4 of 5 COCs ⁶¹																															
Client ID : GFEL01		Project ID : 97-34-63RD RD, Kings Park NY				Sampler Name(s): FRANK GALDUN				Analysis		Matrix																													
Customer Name : GFE LLC		Project Manager : FRANK GALDUN				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified																																			
Address : 58 Nokomis Ave		Phone Number : 646-542-3465																																							
City : Lake Hiawatha		Fax Number : 973-334-1692																																							
State : NJ		Site Details: 3955 VICTORY BLVD. STATEN ISLAND NY																																							
Zip Code : 07034		Analysis Turnaround Time 5 DAY				Data Package Type : RESULTS ONLY																																			
Country :		Standard : 10 business days OR				EDD Type : PDF																																			
Rush (Specify): 5 Days		Can Vacuum in Field ("Hg) (Start)		Can Vacuum in Field ("Hg) (Stop)**		Interior Temp. (F) (Start)		Interior Temp. (F) (Stop)		Out going Can Pressure ("Hg)(Lab)		In coming Can Pressure ("Hg)(Lab)		Flow Reg. ID		Can ID		Flow Controller Readout (ml/min)		Can Cert ID		TO-15		Indoor Ambient Air		Soil Gas															
Sample Identification		Sample Date(s)		Time Start (24 hr Clock)		Time Stop (24 hr Clock)		30		4.5		68		68		-30		-4.7		10613		10266		6 L		50		VL041954.D		1		1									
Temperature (Fahrenheit)												GC/MS Analyst Signature (TO-15) 																													
		Ambient		Maximum		Minimum																																			
Start																																									
Stop																																									
Pressure (Inches of Hg)												** Submission of this COC indicates approval of the analysis based on existing conditions. REPORT ONLY: PCE, TCE, cis-1,2-DCE, 1,1-DCE, 1,1,1-TCA, VINYL CHLORIDE Please follow the instructions on the back of this COC.																													
		Ambient		Maximum		Minimum																																			
Start																																									
Stop																																									
Special Instructions/QC Requirements & Comments :																																									
Suspected Contamination: High Medium <u>Low</u> PID Readings: <u>0.0</u>																																									
Sampling site (State):																																									
Quick Connector required : <u>N/D</u>																																									
Canisters Shipped by: <u>Serrin</u>						Date/Time: <u>03/15/25</u>						Canisters Received by: <u>AR</u>						Date/Time: <u>3-7-25 11:00</u>																							
Samples Relinquished by: <u>FRANK</u>						Date/Time: <u>3/17/25</u>						Received by:						Date/Time:																							
Relinquished by:						Date/Time:						Received by:						Date/Time:																							

B2503004 - 3

Client Contact Information				Bottle Order ID : B2503004				Courier : F GALDUN				5 of 5 COCs 6.1				
Client ID : GFEL01 Project ID : 9X34 63RD RD, Rego Park NY				Project Manager : FRANK GALDUN				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix		
Customer Name : GFE LLC				Phone Number : 646-542-3465				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified								
Address : 58 Nokomis Ave				Fax Number : 973-334-1692												
City : Lake Hiawatha				Site Details: 3999 VICTORY BLD STATEN ISLAND NY												
State : NJ				Analysis Turnaround Time 5 DAY				Data Package Type : RESULTS ONLY				Indoor/Ambient Air Soil Gas				
Zip Code : 07034				Standard : 10 business days OR				EDD Type : PDF								
Country :				Rush (Specify): 5 Days												
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID	Flow Controller Readout (ml/min)	Can Cert ID	(TO-15)	Indoor/Ambient Air	Soil Gas
OAI	7/15/25	9:22	11:22	30	2	39	43	-30	-3.9	10617	10153	6 L	50	VL041954.D		
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) [Signature]						
		Ambient	Maximum	Minimum												
Start																
Stop																
Pressure (Inches of Hg)										* Submittal of this COC indicates approval of the analysis based on existing conditions. REPORT ONLY: PCE, TCE, cis-1,2-DCE, 1,1-DCE, 1,1,1-TCA, VINYL CHLORIDE Please follow the instructions on the back of this COC.						
		Ambient	Maximum	Minimum												
Start																
Stop																
Special Instructions/QC Requirements & Comments :																
Suspected Contamination: High Medium Low PID Readings: 0.0																
Sampling site (State):																
Quick Connector required : NO																
Canisters Shipped by: san				Date/Time: 03/5/25				Canisters Received by: ar				Date/Time: 3-7-25 14:00				
Samples Relinquished by: FUN				Date/Time: 3/7/25				Received by:				Date/Time:				
Relinquished by:				Date/Time:				Received by:				Date/Time:				

B2503004 - 1

From: frankg4 <frankg4@optonline.net>
Sent: Thursday, March 13, 2025 11:33 AM
Subject: RE: Project 3555 Victory Blvd - Q1533

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Sent via the Samsung Galaxy S9, an AT&T 5G Evolution capable smartphone
Yes mistake analyze for 6

----- Original message -----

From: Kiran Saleem <Kiran.Saleem@alliancetg.com>
Date: 3/13/25 10:21 AM (GMT-05:00)
To: frankg4 <frankg4@optonline.net>
Subject: Project 3555 Victory Blvd - Q1533


Good Morning Frank,

I am reaching out to check with you the Parameter list for Q1533 - All samples parameter list are same, but as per coc page 3 ,sample#03, SV2, parameter list is 4 compounds only, is it a mistake? Please see the COC attached.

Thanks.

NOTE: Chemtech is now an Alliance Technical Group company. Please add AllianceTG.com to your safe senders list to ensure receipt of important emails.

Regards,



Kiran Saleem
Project Manager
Alliance Technical Group
Main: 908-789-8900
Direct: 908-728-3148
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488


Internal Chain of Custody

Instructions: Use 1 form for each 20 samples of aliquot

Laboratory Person Breaking Field Seal on Sample Shuttle & Accepting Responsibility for Sample

Laboratory: <u>Chemtech</u>	Location: <u>284 Sheffield Street, Mountainside, NJ 7092</u>
Field Sample Seal No. <u>Q1533</u>	Title: <u>Sample Custodian</u>
Case No.: <u>3555 Victory Blvd, Staten</u>	Date Broken <u>3/7/2025</u> Military Time Seal Broken: <u>14:00:00</u>
	Analytical Parameter/Fraction: <u>MOCMS Group3</u>

Sample No.	Aliquot/Extract No.	Sample No.	Aliquot/Extract No.
Q1533-01	SV1		
Q1533-02	IA1		
Q1533-03	SV2		
Q1533-04	IA2		
Q1533-05	OA1		

Date	Time	Relinquished By	Received By	Purpose of Change of Custody
3/10/25	9:51	Signature 	Signature 	
		Printed Name <u>Casmax Perri</u>	Printed Name <u>Samantha Yelton</u>	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	

Distribution: White - Original (Sent With Report) Yellow - Contractor Archive Pink - Sample Custodian - Interim Copy