

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GFEL01
 Lab Code: CHEM Case No.: Q1533 SAS No.: Q1533 SDG No.: Q1533
 Instrument ID: MSVOA_L Calibration Date(s): 02/25/2025 02/25/2025
 Heated Purge: (Y/N) N Calibration Time(s): 08:56 12:07
 GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042041.D		RRF002 = VL042042.D		RRF001 = VL042043.D		
		RRF0.5 = VL042044.D		RRF0.1 = VL042045.D		RRF.03 = VL042046.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
Vinyl Chloride	0.337	0.361	0.346	0.376	0.336	0.447	0.364	10.9
1,1-Dichloroethene	0.320	0.386	0.437	0.398			0.376	12.6
cis-1,2-Dichloroethene	1.237	1.450	1.401	1.421			1.353	7.3
1,1,1-Trichloroethane	1.771	1.944	1.916	1.883	1.875	1.921	1.866	4.1
Trichloroethene	0.378	0.447	0.447	0.433	0.462	0.447	0.427	8.1
Tetrachloroethene	0.327	0.385	0.383	0.380	0.374	0.322	0.358	7.9
1-Bromo-4-Fluorobenzene	0.764	0.765	0.777	0.768	0.760	0.773	0.769	0.9

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.