

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL022525AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Wed Feb 26 01:08:36 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL042046.D 0.1 =VL042045.D 0.5 =VL042044.D 1 =VL042043.D 2 =VL042042.D 10 =VL042041.D 15 =VL042047.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.109	1.084	1.091	0.943	0.973	1.040	7.31
3) Chlorodifluoro...			1.493	1.566	1.535	1.311	1.362	1.453	7.67
4) Chloromethane			0.438	0.398	0.399	0.344	0.344	0.385	10.53
5) T Vinyl Chloride	0.447	0.336	0.376	0.346	0.361	0.337	0.343	0.364	10.89
6) T Bromomethane			0.170	0.184	0.186	0.160	0.165	0.173	6.63
7) Chloroethane			0.151	0.158	0.156	0.135	0.140	0.148	6.83
8) T Dichlorotetra...			0.892	0.910	0.901	0.779	0.791	0.855	7.47
9) T Propene			0.614	0.630	0.636	0.522	0.558	0.592	8.47
10) T Heptane			1.971	1.946	1.960	1.662	1.679	1.844	8.59
11) T Trichlorofluor...			1.056	1.122	1.119	0.944	0.968	1.042	7.98
12) T 1,1,2-Trichlor...			0.919	0.902	0.865	0.697	0.773	0.831	11.28
13) Ethanol			0.053	0.057	0.040	0.020	0.018	0.038#	48.50
14) T Bromoethene			0.325	0.304	0.299	0.264	0.275	0.293	8.41
15) T Acetone			1.084	0.960	0.933	0.714	0.723	0.883	18.17
16) T 1,3-Butadiene			0.476	0.435	0.418	0.358	0.367	0.411	11.96
17) tert-Butyl alc...			1.184	1.185	1.201	0.950	1.004	1.105	10.71
18) T 1,1-Dichloroet...			0.398	0.437	0.386	0.320	0.338	0.376	12.57
19) T Isopropyl Alcohol			0.561	0.571	0.557	0.419	0.429	0.507	15.06
20) T Methylene Chlo...			0.429	0.374	0.329	0.261	0.278	0.334	20.72
21) T Allyl Chloride			0.707	0.717	0.639	0.538	0.590	0.638	11.94
22) T trans-1,2-Dich...			0.481	0.506	0.450	0.358	0.395	0.438	13.90
23) T Vinyl Acetate			0.689	0.614	0.695	0.520	0.586	0.621	11.82
24) T 1,1-Dichloroet...			0.927	0.856	0.863	0.669	0.758	0.815	12.42
25) T Ethyl Acetate			3.833	3.674	3.776	3.174	3.217	3.535	8.91
26) T Hexane			1.596	1.462	1.519	1.282	1.287	1.429	9.83
27) T Carbon Disulfide			1.042	0.972	1.032	0.856	0.942	0.969	7.78
28) T Methyl tert-Bu...			0.650	0.844	0.838	0.522	0.705	0.711	19.02
29) T Chloroform			2.165	2.044	1.971	1.722	1.716	1.923	10.36
30) T Cyclohexane			1.154	1.239	1.255	1.076	1.076	1.160	7.39
31) T cis-1,2-Dichlo...			1.421	1.401	1.450	1.237	1.255	1.353	7.33
32) T 1,1,1-Trichlor...	1.921	1.875	1.883	1.916	1.944	1.771	1.751	1.866	4.05

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.833	0.799	0.812	0.692	0.759	0.779	7.15
35) T Carbon Tetrach...	0.664	0.705	0.675	0.723	0.725	0.655	0.681	0.690	4.09
36) T Benzene			1.079	1.084	1.104	0.969	0.999	1.047	5.66
37) T 1,2-Dichloroet...			0.624	0.604	0.598	0.520	0.556	0.580	7.20
38) T Trichloroethene	0.447	0.462	0.433	0.447	0.447	0.378	0.379	0.427	8.06
39) T 1,2-Dichloropr...			0.421	0.415	0.414	0.357	0.379	0.397	7.02

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\ Method File : VL022525AIR.M											
40)	T	1,4-Dioxane			0.192	0.196	0.181	0.164	0.173	0.181	7.31
41)	T	Tetrahydrofuran			0.469	0.492	0.473	0.418	0.450	0.460	6.04
42)	T	Bromodichlorom...			0.806	0.791	0.810	0.739	0.765	0.782	3.81
43)		Methyl Methacr...			0.400	0.418	0.426	0.392	0.409	0.409	3.35
44)	T	2,2,4-Trimethy...			1.928	1.935	1.955	1.670	1.750	1.848	6.99
45)	T	t-1,3-Dichloro...			0.414	0.392	0.418	0.393	0.414	0.406	3.12
46)	T	cis-1,3-Dichlo...			0.514	0.523	0.548	0.516	0.526	0.526	2.59
47)	T	1,1,2-Trichlor...			0.426	0.429	0.426	0.361	0.375	0.403	8.12
48)	T	Dibromochlorom...			0.618	0.667	0.667	0.594	0.613	0.632	5.25
49)	T	Bromoform			0.553	0.551	0.588	0.525	0.526	0.549	4.68
50)	T	4-Methyl-2-Pen...			0.996	0.974	0.975	1.014	1.079	1.008	4.27
51)	T	2-Hexanone			0.577	0.570	0.608	0.834	0.879	0.693	21.69
52)	T	Tetrachloroethene	0.322	0.374	0.380	0.383	0.385	0.327	0.337	0.358	7.88
53)	T	Toluene			1.150	1.210	1.284	1.088	1.118	1.170	6.67
54)	T	1,2-Dibromoethane		0.563	0.556	0.553	0.586	0.500	0.514	0.545	5.87
-----ISTD-----											
55)	I	Chlorobenzene-d5									
56)		1,1,1,2-Tetrac...			0.607	0.596	0.595	0.505	0.524	0.565	8.37
57)	T	Chlorobenzene			1.102	1.113	1.121	0.910	0.938	1.037	10.02
58)	T	Ethyl Benzene			1.886	1.906	1.974	1.694	1.749	1.842	6.31
59)	T	m/p-Xylene			1.488	1.550	1.580	1.330	1.380	1.466	7.35
60)	T	o-Xylene			1.572	1.550	1.577	1.317	1.351	1.473	8.68
61)	T	Styrene			0.659	0.719	0.738	0.658	0.690	0.693	5.18
62)		Isopropylbenzene			2.299	2.339	2.399	1.969	2.026	2.207	8.84
63)	T	1,1,2,2-Tetrac...	0.999	0.975	1.026	1.041	1.031	0.848	0.872	0.970	8.13
64)		n-propylbenzene			0.596	0.607	0.632	0.524	0.535	0.579	8.11
65)		tert-Butylbenzene			2.080	2.251	2.222	1.733	1.753	2.008	12.47
66)	T	Benzyl Chloride			0.141	0.159	0.163	0.153	0.150	0.153	5.40
67)		sec-Butylbenzene			2.973	3.229	3.182	2.482	2.494	2.872	12.65
68)	S	1-Bromo-4-Fluo...	0.773	0.760	0.768	0.777	0.765	0.764	0.779	0.769	0.92
69)		p-Isopropyltol...			2.240	2.533	2.575	2.021	2.056	2.285	11.38
70)		n-Butylbenzene			2.332	2.567	2.572	2.070	2.120	2.332	10.20
71)		2-Chlorotoluene			1.758	1.829	1.853	1.514	1.551	1.701	9.30
72)	T	4-Ethyltoluene			1.775	1.944	1.985	1.621	1.670	1.799	8.99
73)	T	1,3,5-Trimethy...			1.592	1.732	1.696	1.381	1.400	1.560	10.47
74)	T	1,2,4-Trimethy...			1.855	1.982	1.990	1.471	1.488	1.757	14.76
75)	T	1,3-Dichlorobe...			1.112	1.110	1.083	0.869	0.882	1.011	12.30
76)	T	1,4-Dichlorobe...			1.120	1.061	1.092	0.861	0.886	1.004	12.07
77)	T	1,2-Dichlorobe...			1.059	1.097	1.063	0.836	0.852	0.981	12.88
78)	T	Hexachloro-1,3...			0.840	0.832	0.844	0.637	0.666	0.764	13.53
79)	T	Naphthalene		0.841	0.978	1.237	1.352	1.371	1.450	1.205	20.15
80)	T	Naphthalene,2-...			0.090	0.142	0.240	0.562	0.615	0.330	73.74
81)	T	1,2,4-Trichlor...			0.589	0.687	0.765	0.644	0.673	0.671	9.58

(#) = Out of Range