

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GFEL01
 Lab Code: CHEM Case No.: Q1533 SAS No.: Q1533 SDG NO.: Q1533
 Lab File ID: VL042115.D Date Analyzed: 03/10/2025
 Instrument ID: MSVOA_L Time Analyzed: 09:08
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
1DUP	0	0.00				
SV1	0	0.00				
IA1	0	0.00				
SV2	0	0.00				
IA2	0	0.00				
OA1	0	0.00				
VL0310ABL01	0	0.00				
VL0310ABS01	0	0.00				

IS4 =

AREA UPPER LIMIT = +40% of internal standard area

AREA LOWER LIMIT = -40% of internal standard area

RT UPPER LIMIT = +0.33 minutes of internal standard RT

RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.