

Preparation Blank Summary

Client: Weston Solutions

SDG No.: Q1539

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: lb134983BL							
Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.003	0.01	03/11/2025
Sample ID: LB134995BLW							
TOC	mg/L	0.27	0.5000	J	0.19	1	03/12/2025
Sample ID: LB135005BLW							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	03/11/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	03/11/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	03/11/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	03/11/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	03/11/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	03/11/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	03/11/2025
Sample ID: LB135027BL							
Oil and Grease	mg/L	< 2.5000	2.5000	U	0.4	5.0	03/14/2025
Sample ID: PB167063BL							
Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	03/12/2025