

Laboratory Control Sample Summary

Client: Weston Solutions

SDG No.: Q1539

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Run No.: LB134983

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	Ib134983BS							
Hexavalent Chromium	mg/L	0.5	0.51		102	1	90-111	03/11/2025

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	Q1539
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Run No.:	LB134995

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134995BSW							
TOC	mg/L	10	9.40		94	1	90-110	03/12/2025

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	Q1539
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Run No.:	LB135005

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135005BSW							
Bromide	mg/L	10	10.5		105	1	90-110	03/11/2025
Chloride	mg/L	3	3.10		103	1	90-110	03/11/2025
Fluoride	mg/L	2	2.10		105	1	90-110	03/11/2025
Nitrite	mg/L	3	3.10		103	1	90-110	03/11/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	03/11/2025
Sulfate	mg/L	15	15.5		103	1	90-110	03/11/2025
Orthophosphate as P	mg/L	5	4.70		94	1	90-110	03/11/2025

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	Q1539
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Run No.:	LB135027

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135027BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	03/14/2025

Laboratory Control Sample Summary

Client: Weston Solutions

SDG No.: Q1539

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Run No.: LB135012

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167063BS							
Ammonia as N	mg/L	1	0.98		98	1	90-110	03/12/2025