



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Weston Solutions SDG No.: Q1539
Contract: WEST04 Lab Code: CHEM Case No.: Q1539 SAS No.: Q1539
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.22	0.2	109	40 - 160	CV	03/12/2025	11:15	LB134996
CRI	Aluminum	17.7	20.0	88	70 - 130	P	03/23/2025	16:11	LB135143
	Antimony	1.92	2.0	96	70 - 130	P	03/23/2025	16:11	LB135143
	Arsenic	1.17	1.0	117	70 - 130	P	03/23/2025	16:11	LB135143
	Barium	8.95	10.0	90	70 - 130	P	03/23/2025	16:11	LB135143
	Beryllium	1.00	1.0	100	70 - 130	P	03/23/2025	16:11	LB135143
	Cadmium	0.91	1.0	91	70 - 130	P	03/23/2025	16:11	LB135143
	Calcium	458	500	92	70 - 130	P	03/23/2025	16:11	LB135143
	Chromium	1.82	2.0	91	70 - 130	P	03/23/2025	16:11	LB135143
	Cobalt	0.92	1.0	92	50 - 150	P	03/23/2025	16:11	LB135143
	Copper	2.06	2.0	103	70 - 130	P	03/23/2025	16:11	LB135143
	Iron	52.1	50.0	104	70 - 130	P	03/23/2025	16:11	LB135143
	Lead	0.79	1.0	79	70 - 130	P	03/23/2025	16:11	LB135143
	Magnesium	438	500	88	70 - 130	P	03/23/2025	16:11	LB135143
	Manganese	0.84	1.0	84	50 - 150	P	03/23/2025	16:11	LB135143
	Nickel	0.86	1.0	86	70 - 130	P	03/23/2025	16:11	LB135143
	Potassium	456	500	91	70 - 130	P	03/23/2025	16:11	LB135143
	Selenium	5.46	5.0	109	70 - 130	P	03/23/2025	16:11	LB135143
	Silver	0.98	1.0	98	70 - 130	P	03/23/2025	16:11	LB135143
	Sodium	443	500	88	70 - 130	P	03/23/2025	16:11	LB135143
	Thallium	0.91	1.0	91	70 - 130	P	03/23/2025	16:11	LB135143
	Vanadium	4.48	5.0	90	70 - 130	P	03/23/2025	16:11	LB135143
	Zinc	4.52	5.0	90	50 - 150	P	03/23/2025	16:11	LB135143