

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Weston Solutions **SDG No.:** Q1539
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1539 **SAS No.:** Q1539
Matrix: Water **Level:** LOW **Client ID:** TAPFTA-MW01I-031025-00-T2A
Sample ID: Q1539-02 **Spiked ID:** Q1539-02A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	84 - 116	16.4		16.9		500	0		P
Barium	ug/L	86 - 114	2140		14.7		2500	85		P
Iron	ug/L	87 - 118	45100		24300		25000	83		P
Manganese	ug/L	87 - 115	4400		70.2		5000	87		P
Potassium	ug/L	87 - 115	22000		820		25000	85		P
Silver	ug/L	85 - 116	66.0		1.00	U	500	13		P