

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1539
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1539 **SAS No.:** Q1539
Matrix: Water **Sample ID:** Q1539-02 **Client ID:** TAPFTA-MW01I-031025-00-T2DUP
Percent Solids for Sample: NA **Duplicate ID** Q1539-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	67.3		85.8		24	*	P
Antimony	ug/L	20	2.00	U	2.00	U			P
Arsenic	ug/L	20	16.9		16.7		1		P
Barium	ug/L	20	14.7		15.1		3		P
Beryllium	ug/L	20	1.00	U	1.00	U			P
Cadmium	ug/L	20	1.00	U	1.00	U			P
Calcium	ug/L	20	1950		1930		1		P
Chromium	ug/L	20	0.65	J	0.70	J	7		P
Cobalt	ug/L	20	1.00	U	1.00	U			P
Copper	ug/L	20	0.75	J	0.62	J	19		P
Iron	ug/L	20	24300		24100		1		P
Lead	ug/L	20	0.40	J	0.41	J	2		P
Magnesium	ug/L	20	480	J	479	J	0		P
Manganese	ug/L	20	70.2		70.1		0		P
Nickel	ug/L	20	0.35	J	1.00	U	200.0		P
Potassium	ug/L	20	820		835		2		P
Selenium	ug/L	20	5.00	U	5.00	U			P
Silver	ug/L	20	1.00	U	1.00	U			P
Sodium	ug/L	20	3910		3940		1		P
Thallium	ug/L	20	1.00	U	1.00	U			P
Vanadium	ug/L	20	0.53	J	0.54	J	2		P
Zinc	ug/L	20	1.87	J	1.38	J	30		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1539
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1539 **SAS No.:** Q1539
Matrix: Water **Sample ID:** Q1539-02MS **Client ID:** TAPFTA-MW01I-031025-00-T2MSD
Percent Solids for Sample: NA **Duplicate ID** Q1539-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	8610		9040		5		P
Antimony	ug/L	20	494		479		3		P
Arsenic	ug/L	20	16.2		16.0		1		P
Barium	ug/L	20	2190		2110		4		P
Beryllium	ug/L	20	430		421		2		P
Cadmium	ug/L	20	452		436		4		P
Calcium	ug/L	20	47500		47500		0		P
Chromium	ug/L	20	459		453		1		P
Cobalt	ug/L	20	473		461		3		P
Copper	ug/L	20	4270		4290		0		P
Iron	ug/L	20	44700		44400		1		P
Lead	ug/L	20	2360		2340		1		P
Magnesium	ug/L	20	46600		46300		1		P
Manganese	ug/L	20	4360		4350		0		P
Nickel	ug/L	20	465		460		1		P
Potassium	ug/L	20	21900		22000		0		P
Selenium	ug/L	20	490		475		3		P
Silver	ug/L	20	67.4		64.1		5		P
Sodium	ug/L	20	49600		49600		0		P
Thallium	ug/L	20	469		460		2		P
Vanadium	ug/L	20	478		471		1		P
Zinc	ug/L	20	4360		4330		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1539
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1539 **SAS No.:** Q1539
Matrix: Water **Sample ID:** Q1541-01 **Client ID:** 205703DUP
Percent Solids for Sample: NA **Duplicate ID** Q1541-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1539
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1539 **SAS No.:** Q1539
Matrix: Water **Sample ID:** Q1541-01MS **Client ID:** 205703MSD
Percent Solids for Sample: NA **Duplicate ID** Q1541-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.94		3.82		3		CV