

Analytical Sequence

Client: Weston Solutions

SDG No.: Q1539

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Instrument ID: FID_G

GC Column: RXI-1MS **ID:** 0.18 (mm)

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

MEAN SUROGATE RT FROM INITIAL CALIBRATION 15.0338					
EPA SAMPLE NO.	LAB SAMPLE ID	DATE AND TIME ANALYZED	DATAFILE	RT	#
PIBLK01	LBLK01	12 Mar 2025 10:29	FG015471.D	15.029	
50 PPM TRPH STD	50 PPM TRPH STD	12 Mar 2025 10:59	FG015472.D	15.031	
PB167101BL	PB167101BL	12 Mar 2025 13:57	FG015474.D	15.027	
PB167101BS	PB167101BS	12 Mar 2025 14:26	FG015475.D	15.027	
PB167101BSD	PB167101BSD	12 Mar 2025 14:55	FG015476.D	15.026	
TAPIAL3-MW03D-031025-00-T1	Q1539-01	12 Mar 2025 16:23	FG015479.D	14.970	
TAPFTA-MW01I-031025-00-T2	Q1539-02	12 Mar 2025 16:53	FG015480.D	15.027	
PIBLK02	LBLK02	12 Mar 2025 17:22	FG015481.D	15.027	
50 PPM TRPH STD	50 PPM TRPH STD	12 Mar 2025 18:21	FG015482.D	15.030	

Column used to flag RT values with an * values outside of QC limits

QC Limits
(± 0.10 minutes)

Lower Limit
14.9338

Upper Limits
15.1338