

**WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE**

**Lab Name:** Chemtech **Client:** Weston Solutions  
**Lab Code:** CHEM **Cas No:** Q1539 **SAS No :** Q1539 **SDG No:** Q1539  
**Matrix Spike - EPA Sample No :** PB167101BS **Datafile:** FG015475.D

| COMPOUND | SPIKE<br>ADDED<br>ug/L | CONCENTRATION<br>ug/L | LCS/LCSD<br>CONCENTRATION<br>ug/L | % REC | QC LIMITS |
|----------|------------------------|-----------------------|-----------------------------------|-------|-----------|
| DRO      | 200                    | 0                     | 206                               | 103   | 78-117    |